



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of Govt of Jammu & Kashmir)

Corporate Head Office: 1st Floor, Drug Store Building, Govt. Medical College, Bakshi Nagar, Jammu

Corporate Office: Opp. UT of Motor Garages Near Haj House, Bemina, Srinagar

Telephone: 0191-2580842, Fax: 0191-2581845 (Jammu); Telefax: 0194-2432008 (Srinagar)

NOTICE

Subject: Evaluation of Technical bids for finalization of Rate Contract for procurement of "Instrument Sets" for Bone & Joint Hospital with reference No. NIT/JKMSCL/M&E/2023/599 dated: 17-08-2023 with reference to clause 2.1.9 Chapter II of the Standard Procurement Procedure approved during 2nd Board Meeting of JKMSCL.

With reference to the report of the Sub-committee, the evaluation of bid documents for the finalization of the rate contract for the procurement of the "Instruments" with reference to the NIT No. JKMSCL/M&E/2023/599 dated 17.08.2023, the preliminary evaluation of the bid documents have been done on the basis of uploaded documents and following infirmities have been found:


S No.	Name of instrument	Name of bidders	Items quoted (As per System Generated)	Shortfall documents
1.	General Shoulder Instruments	ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080 Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	1. General Shoulder Instruments 2. Delta Retractor 3. Delta Retractor, pointed 4. Bone Retractor, Humeral Head 5. Ring Retractor 6. Ring Narrow 7. Ring Retractor, Drop-Shaped 8. Ring Retractor, Drop-Shaped, narrow 9. Ring Retractor, delta-Shaped 10. Shoulder Spreader, left bend 11. Shoulder Spreader, right bend	To submit:- 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd.). 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		SHARMA ENTERPRISES Importer:- M/S Ahuja sons Surgicals Manufacture:- INNOMED, INC - USA	1. General Shoulder Instruments 2. Delta Retractor 3. Delta Retractor, pointed 4. Bone Retractor, Humeral Head 5. Ring Retractor 6. Ring Narrow 7. Ring Retractor, Drop-Shaped 8. Ring Retractor, Drop-Shaped, narrow 9. Ring Retractor, delta-Shaped 10. Shoulder Spreader, left bend 11. Shoulder Spreader, right bend	To submit:- 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable.

				12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
2.	COLLINEAR REDUCTION CLAMP	ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080 Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	1. COLLINEAR REDUCTION CLAMP 2. Collinear Reduction Clamp Sliding Mechanism 3. Combination Wrench dia 8 4. Pelvic Arm f/ Collinear Reduction Clamp 5. Percut-Arm f/ Collinear Reduction Clamp 6. BoneHook-shape Arm f/ Collinear Reduction 7. LCP Reduct-Attachm f/ Collinear Reduction 8. Reduction-Attachm f/ Collinear Reduction 9. Combination Wrench dia 13 10. K-Wire o2.6 w/ spade point tip L500 11. Disc spiked f/ Reduct-Forceps 12. Sterilization Case	To submit: - 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd). 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		SHARMA ENTERPRISES Importer: - M/S Ahuja sons Surgicals Manufacture: - INNOMED, INC - USA	1. COLLINEAR REDUCTION CLAMP 2. Collinear Reduction Clamp Sliding Mechanism 3. Combination Wrench dia 8 4. Pelvic Arm f/ Collinear Reduction Clamp 5. Percut-Arm f/ Collinear Reduction Clamp 6. BoneHook-shape Arm f/ Collinear Reduction 7. LCP Reduct-Attachm f/ Collinear Reduction 8. Reduction-Attachm f/ Collinear Reduction 9. Combination Wrench dia 13 10. K-Wire o2.6 w/ spade point tip L500 11. Disc spiked f/ Reduct-Forceps Sterilization Case	To submit: - 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of importer product). 11. Quality Certifications on the products viz ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
3.	Specification for Broken Screw and Nail Removal	ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080	1. The set should have screwdrivers & screwdriver shafts, forceps for screw removal, conical extraction screws with T handle, hollow reamers & extraction bolts with T handle. 2. The instrument set should contain a Hollow Reamer for 3.5/4.0 mm 3. The instrument set should contain a Spare Reamer Tube for no. 309.035 4. The instrument set should contain an Extraction Bolt for 3.5/4.0mm	To submit: - 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).

		Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	Screws 5. The instrument set should contain an Extraction Screw, conical, for 2.7mm, 3.5mm and 4.0mm Screws 6. The instrument set should contain a Hollow Reamer for 4.5mm Screws 7. The instrument set should contain a Spare Reamer Tube 8. The instrument set should contain an Extraction Bolt for 4.5mm Screws 9. The instrument set should contain an Extraction Screw, conical, for 4.5/6.5mm Screws 10. The instrument set should contain a Hollow Reamer for 5.0/6.0/6.5/7.0mm Screws 11. The instrument set should contain a Spare Reamer Tube for no. 309.065 12. The instrument set should contain an Extraction Bolt, for 5.0/6.0/6.5/7.0mm Screws 13. The instrument set should contain an anodized Aluminum Plate 14. The instrument set should contain a Sharp Hook of length 155mm 15. The instrument set should contain a Forceps for Screw Removal, L 205 mm 16. The instrument set should contain a Gouge, 10mm wide and length 205mm 17. The instrument set should contain a T-Handle with quick coupling of length 80mm Aluminum Case 18. Aluminum Case	4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd.) 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		SHARMA ENTERPRISES Manufacturer: Nebula Surgical Pvt Ltd-Gujrat - 360005	1. The set should have screwdrivers & screwdriver shafts, forceps for screw removal, conical extraction screws with T handle, hollow reamers & extraction bolts with T handle. 2. The instrument set should contain a Hollow Reamer for 3.5/4.0 mm Screws 3. The instrument set should contain a Spare Reamer Tube for no. 309.035 4. The instrument set should contain an Extraction Bolt for 3.5/4.0mm Screws 5. The instrument set should contain an Extraction Screw, conical, for 2.7mm, 3.5mm and 4.0mm Screws 6. The instrument set should contain a Hollow Reamer for 4.5mm Screws 7. The instrument set should contain a Spare Reamer Tube 8. The instrument set should contain an Extraction Bolt for 4.5mm Screws 9. The instrument set should contain an Extraction Screw, conical, for 4.5/6.5mm Screws 10. The instrument set should contain a Hollow Reamer for 5.0/6.0/6.5/7.0mm Screws 11. The instrument set should contain a Spare Reamer Tube for no. 309.065 12. The instrument set should contain an Extraction Bolt, for 5.0/6.0/6.5/7.0mm Screws 13. The instrument set should contain an anodized Aluminum Plate 14. The instrument set should contain a Sharp Hook of length 155mm 15. The instrument set should contain a Forceps for Screw Removal, L 205 mm 16. The instrument set should contain a Gouge, 10mm wide and length 205mm 17. The instrument set should contain a T-Handle with quick coupling of length 80mm Aluminum Case 18. Aluminum Case	To submit:- 1. Compliance Sheet for each equipment. 2. Technical bid submission sheet duly filled 3. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Principal manufacturer as per proforma duly notarized. 4. Model of each quoted instruments. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 7. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 8. Latest GST Returns of the Bidder 9. Declaration of bidder regarding acceptance Bid for terms & conditions 10. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
4.	BROKEN NAIL REMOVAL	Zero bid		
5.	Pelvic Set	ORTHOFIX (Bidder)	1. The Instrument should have Schanz Screw dia 6.0 mm length 190/50 mm, Stainless Steel 2. The Instrument should have	To submit:- 1. Copy of catalogue of each quoted item (self-attested).

	<p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacturer:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<p>Universal Chuck with T-Handle, L 150+20mm</p> <ol style="list-style-type: none"> 3. The Instrument should have Bone Hook, sharp, medium, L 225+20mm 4. The Instrument should have Bone Hook, sharp, large, L 200+20mm 5. The Instrument should have Forceps for Screw Removal, L 200+20mm 6. The Instrument should have Pelvic Reduction Forceps, small, length, 200+20mm, for use with Cortex Screws dia 3.5 and 4.5mm 7. The Instrument should have Pelvic Reduction Forceps, medium, length, 250+20mm, for use with Cortex Screws dia 3.5 and 4.5mm 8. The instrument should have Pelvic Reduction Forceps, large, adjustable, speed lock, length 325+20mm 9. The Instrument should have Reduction Forceps, large, with points, L200+20mm 10. The Instrument should have Ball Spike, straight, with pointed ball tips dia 6.5mm, length 300+20mm 11. The Instrument should have Ball Spike, straight, long, with pointed ball tips dia 6.5mm, length 400+20mm 12. The Instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 250+20mm 13. The instrument should have Pelvic Reduction Forceps, oblique, with pointed ball tips dia 6.5mm, length 200+20mm 14. The instrument should have Pelvic Reduction Forceps, oblique, with pointed ball tips dia 6.5mm, length 250+20mm 15. The instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 400+20mm 16. The Instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 250+20mm 17. The Instrument should have Pelvic Reduction Forceps, Small, for Screws dia 3.5mm, length 250+20mm 18. The Instrument should have Disc, hole dia 6.5mm 19. The Instrument should have Disc, rectangular, hole dia 6.5mm 20. The Instrument should have Bone Hook, medium, long, length 325+20mm 21. The Instrument should have Bone Hook, large, long, length 325+20mm 22. The Instrument should have Pelvic Retractor, blunt, L 275+20mm 23. The Instrument should have Radiolucent Pelvic Retractor, blunt, length 275+20mm 24. The Instrument should have retractor, 18mm wide, long narrow tip, L 225+20mm 25. The Instrument should have Radiolucent Hohmann Retractor, width 18mm, length 250+20mm 26. The Instrument should have Pelvic Retractor, medium, length 275+20mm 27. The Instrument should have Radiolucent Pelvic Retractor, medium length 275+20mm 28. The Instrument should have Pelvic Retractor, large, length 325+20mm 29. The Instrument should have Radiolucent Pelvic Retractor, large, length 325+20mm 30. The Instrument should have Retractor, can be contoured, width 	<ol style="list-style-type: none"> 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd). 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
--	--	---	---

			<p>20mm</p> <p>31. The Instrument should have Retractor, can be contoured, width 30mm</p> <p>32. The instrument should have Retractor, can be contoured, width 40mm</p> <p>33. The instrument should have Radiolucent Hohmann Retractor, width 24 mm lengths 275±20mm</p> <p>34. The Instrument should have In-situ Bending and Twisting Handle, straight</p> <p>35. The Instrument should have In-situ Bending and Twisting Handle, 90°</p> <p>36. The Instrument should have In-situ Bending and Twisting Handle, 120°</p> <p>37. The Instrument should have In-situ Bender</p> <p>38. The Instrument should have Drill Bit, 3.5mm dia., L 195/170mm, for quick coupling</p> <p>39. The Instrument should have Tap for Cortex Screws dia 3.5 mm, calibrated, length 175 ±10mm</p> <p>40. The Instrument should have Holding Sleeve.</p> <p>41. The Instrument should have Screwdriver, hexagonal, small, dia 2.5 mm. length 270 +10 mm</p> <p>42. The Instrument should have Drill Bit, 2.5mm dia., L 230/205. +10 mm, 3-flute, with depth marking</p> <p>43. The Instrument should have Drill Bit dia 2.5 mm, calibrated, length 300 mm ±20mm, with Quick Coupling, for Percutaneous Insertion</p> <p>44. The Instrument should have Depth Gauge for Cortex Screws dia 3.5 mm, measuring range up to 150 mm +20mm</p> <p>45. The Instrument should have Bending Iron for Reconstruction Plates 3.5 and 4.5, length 190 mm, ±10mm</p> <p>46. The Instrument should have Bending Pliers for Reconstruction Plates 3.5</p> <p>47. The Instrument should have Handle with Quick Coupling</p> <p>48. The Instrument should have Screwdriver Shaft, hexagonal for Screws dia 3.5 mm, length 250 mm, +10mm</p> <p>49. The Instrument should have Screwdriver Shaft Star drive, length 250mm, ± 10mm</p> <p>50. The instrument set should contain blunt aluminium Radiolucent Pelvic Retractor with length 274 mm+10mm</p> <p>51. The instrument set should contain the Radiolucent hohmann retractors with the following Sizes.</p> <p>52. Width 35 mm, Length 275 mm</p> <p>53. Width 18 mm, Length 240mm</p> <p>54. Width 24 mm, Length 267 mm</p> <p>55. The set should contain 3.5mm Polyaxial angle Drill Guide Its grips should be made of vulcanized ethylene propylene diene monomer</p> <p>56. Instrument should Have Graft cases and lids As per instruments</p> <p>57. T&C- The company should be able to provide complete set as per given item list</p> <p>58. The above said Instrument should be USFDA certified Warranty 2 years</p> <p>59. For instruments, steel grade used should be AISI 316 & 410</p> <p>60. Instrument should be satin finished for glare free under OT Light Company should be able to do demo if required.</p>	
		SHARMA ENTERPRISES	<p>1. The Instrument should have Schanz Screw dia 6.0 mm length</p>	<p>To submit: -</p> <p>1. Copy of Catalogue of the Quoted product</p>




5

		<p>Importer: - M/S Ahuja sons Surgicals</p> <p>Manufacture: - INNOMED, INC - USA</p>	<ol style="list-style-type: none"> 2. 190/50 mm, Stainless Steel 3. The Instrument should have Universal Chuck with T-Handle, L 150+20mm 4. The Instrument should have Bone Hook, sharp, medium, L 225+20mm 5. The Instrument should have Bone Hook, sharp, large, L 200+20mm 6. The Instrument should have Forceps for Screw Removal, L 200+20mm 7. The Instrument should have Pelvic Reduction Forceps, small, length, 200+20mm, for use with Cortex Screws dia 3.5 and 4.5mm 8. The Instrument should have Pelvic Reduction Forceps, medium, length, 250+20mm, for use with Cortex Screws dia 3.5 and 4.5mm 9. The instrument should have Pelvic Reduction Forceps, large, adjustable, speed lock, length 325+20mm 10. The Instrument should have Reduction Forceps, large, with points, L200+20mm 11. The Instrument should have Ball Spike, straight, with pointed ball tips dia 6.5mm, length 300+20mm 12. The Instrument should have Ball Spike, straight, long, with pointed ball tips dia 6.5mm, length 400+20mm 13. The Instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 250+20mm 14. The instrument should have Pelvic Reduction Forceps, oblique, with pointed ball tips dia 6.5mm, length 200+20mm 15. The instrument should have Pelvic Reduction Forceps, oblique, with pointed ball tips dia 6.5mm, length 250+20mm 16. The instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 400+20mm 17. The instrument should have Pelvic Reduction Forceps, with pointed ball tips dia 6.5mm, length 250+20mm 18. The instrument should have Pelvic Reduction Forceps, Small, for Screws dia 3.5mm, length 250+20mm 19. The Instrument should have Disc, hole dia 6.5mm 20. The Instrument should have Disc, rectangular, hole dia 6.5mm 21. The Instrument should have Bone Hook, medium, long, length 325+20mm 22. The Instrument should have Bone Hook, large, long, length 325+20mm 23. The Instrument should have Pelvic Retractor, blunt, L 275+20mm 24. The Instrument should have Radiolucent Pelvic Retractor, blunt, length 275+20mm 25. The Instrument should have retractor, 18mm wide, long narrow tip, L 225+20mm 26. The Instrument should have Radiolucent Hohmann Retractor, width 18mm. length 250+20mm 27. The Instrument should have Pelvic Retractor, medium, length 275+20mm 28. The Instrument should have Radiolucent Pelvic Retractor, medium length 275+20mm 29. The Instrument should have Retractor, large, length 325+20mm 30. The Instrument should have Radiolucent Pelvic Retractor, large, length 325+20mm 	<ol style="list-style-type: none"> 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
--	--	--	---	--



6

			<p>30. The Instrument should have Retractor, can be contoured, width 20mm</p> <p>31. The Instrument should have Retractor, can be contoured, width 30mm</p> <p>32. The instrument should have Retractor, can be contoured, width 40mm</p> <p>33. The instrument should have Radiolucent Hohmann Retractor, width 24 mm lengths 275±20mm</p> <p>34. The Instrument should have In-situ Bending and Twisting Handle, straight</p> <p>35. The Instrument should have In-situ Bending and Twisting Handle, 90°</p> <p>36. The Instrument should have In-situ Bending and Twisting Handle, 120°</p> <p>37. The Instrument should have In-situ Bender</p> <p>38. The Instrument should have Drill Bit, 3.5mm dia., L 195/170mm, for quick coupling</p> <p>39. The Instrument should have Tap for Cortex Screws dia 3.5 mm, calibrated, length 175 ±10mm</p> <p>40. The Instrument should have Holding Sleeve.</p> <p>41. The Instrument should have Screwdriver, hexagonal, small, dia 2.5 mm, length 270 +10 mm</p> <p>42. The Instrument should have Drill Bit, 2.5mm dia., L 230/205, +10 mm, 3-flute, with depth marking</p> <p>43. The Instrument should have Drill Bit dia 2.5 mm, calibrated, length 300 mm ±20mm, with Quick Coupling, for Percutaneous Insertion</p> <p>44. The Instrument should have Depth Gauge for Cortex Screws dia 3.5 mm, measuring range up to 150 mm +20mm</p> <p>45. The Instrument should have Bending Iron for Reconstruction Plates 35 and 4.5, length 190 mm, ±10mm</p> <p>46. The Instrument should have Bending Pliers for Reconstruction Plates 3.5</p> <p>47. The Instrument should have Handle with Quick Coupling</p> <p>48. The Instrument should have Screwdriver Shaft, hexagonal for Screws dia 3.5 mm, length 250 mm, +10mm</p> <p>49. The Instrument should have Screwdriver Shaft Star drive, length 250mm, t 10mm</p> <p>50. The instrument set should contain blunt aluminium Radiolucent Pelvic Retractor with length 274 mm+10mm</p> <p>51. The instrument set should contain the Radiolucent hohmann retractors with the following Sizes.</p> <p>52. Width 35 mm, Length 275 mm</p> <p>53. Width 18 mm, Length 240mm</p> <p>54. Width 24 mm, Length 267 mm</p> <p>55. The set should contain 3.5mm Polyaxial angle Drill Guide Its grips should be made of vulcanized ethylene propylene diene monomer</p> <p>56. Instrument should Have Graft cases and lids As per instruments</p> <p>57. T&C- The company should be able to provide complete set as per given item list</p> <p>58. The above said Instrument should be USFDA certified Warranty 2 years</p> <p>59. For instruments, steel grade used should be AISI 316 & 410</p> <p>60. Instrument should be satin finished for glare free under OT Light Company should be able to do demo if required.</p>	
--	--	--	---	--



6.	Power Flexible Reaming Set	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<ol style="list-style-type: none"> 1. T-Handle with quick coupling. L 85mm 2. Holding Forceps for Reaming Rod 2.5mm 3. Removing Tool for Reaming set 4. Hand Reamer, 8.0mm dia., for predrilling in pseudarthroses 5. Reaming Rod, 2.5mm dia, L.950mm oliva 6. Flexible Shaft, 7.0mm dia., reaming depth to 470mm 7. Cleaning Brush for 3.6mm flexible Shaft . L. 600mm 8. Reduction Head straight 9. Reduction dia, Displacement 2.5mm 10. Reamer Head , 8.5mm dia 11. Reamer Head. 9.0mm dia 12. Reamer Head. 9.5mm dia 13. Reamer Head.10.0mm dia 14. Reamer Head.10.5mm dia 15. Reamer Head.11.0mm dia 16. Reamer Head. 11.5mm dia. 17. Reamer Head.12.0m 18. Reamer Head.12.5mm dia 19. Reamer Head...13.0mm dia 20. Box for above items 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested) . 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		<p>SHARMA ENTERPRISES</p> <p>Importer: - M/S Ahuja sons Surgicals</p> <p>Manufacture: - INNOMED, INC - USA</p>	<ol style="list-style-type: none"> 1. T-Handle with quick coupling. L 85mm 2. Holding Forceps for Reaming Rod 2.5mm 3. Removing Tool for Reaming set 4. Hand Reamer, 8.0mm dia., for predrilling in pseudarthroses 5. Reaming Rod, 2.5mm dia, L.950mm oliva 6. Flexible Shaft, 7.0mm dia., reaming depth to 470mm 7. Cleaning Brush for 3.6mm flexible Shaft . L. 600mm 8. Reduction Head straight 9. Reduction dia, Displacement 2.5mm 10. Reamer Head , 8.5mm dia 11. Reamer Head. 9.0mm dia 12. Reamer Head. 9.5mm dia 13. Reamer Head.10.0mm dia 14. Reamer Head.10.5mm dia 15. Reamer Head.11.0mm dia 16. Reamer Head. 11.5mm dia. 17. Reamer Head.12.0m 18. Reamer Head.12.5mm dia 19. Reamer Head...13.0mm dia 20. Box for above items 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
7.	SPECIFICATION FOR UNIVERSAL SMALL FRAGMENT	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<ol style="list-style-type: none"> 1. The instrument should be made of stainless steel and medical grade polymers - polyaryletherketone, vulcanized ethylene propylene diene monomer and polypropylene. 2. All instrument grips should be made of vulcanized ethylene propylene diene monomer 3. The instruments should be able to maintain stiffness while withstanding high temperatures during sterilization 4. The set should contain 3.5 mm Neutral Sleeve Adapter for 3.5 Non-Locking Drill Guide 5. The set should contain 3.5mm Non-Locking Drill Guide 6. The set should contain 3.5 mm Polyaxial angle Drill Guide 7. The set should contain 2.8mm Threaded Guide for 3.5mm Screw 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested) . 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered

			<p>for both Polyaxial angle locking and standard locking</p> <p>8. The set should 2.7mm Neutral Sleeve Adapter for 2.7 Non-locking Drill Guide</p> <p>9. The set should contain 2.7mm Polyaxialangle Drill Guide</p> <p>10. The set should contain 2.0mm Threaded guide for 2.7mm screw for polyaxial angle and standard locking</p> <p>11. The set should contain centering sleeve for k-wire with diameter 1.6mm with length 70mm</p> <p>12. The set should contain 2.7/3.5mm Depth Gauge 0-60mm</p> <p>13. The set should contain quick coupling 2.0mm Drill Bit with length 30mm</p> <p>14. The set should contain quick coupling 2.0mm Drill Bit with length 60mm</p> <p>15. The set should contain quick coupling 2.5mm Drill Bit with length 45mm</p> <p>16. The set should contain quick coupling 2.5mm Drill Bit with length 80mm</p> <p>17. The set should contain quick coupling 2.7mm Drill Bit with length 45mm</p> <p>18. The set should contain quick coupling 2.8mm Drill Bit with length 45mm</p> <p>19. The set should contain quick coupling 2.8mm Drill Bit with length 80mm</p> <p>20. The set should contain quick coupling 3.5mm Drill Bit with length 60mm</p> <p>21. The set should contain counter sink for 3.5mm cortex and 4.0mm cancellous bone crews</p> <p>22. The set should contain Universal Screwdriver Handle</p> <p>23. The set should contain 2.5mm Hexagonal Driver shaft</p> <p>24. The set should contain stardrive screwdriver shaft with coupling</p> <p>25. The set should contain stardrive screwdriver shaft length 105mm</p> <p>26. The set should contain holding sleeve</p> <p>27. The set should contain 2.5mm torque limiting handle with quick coupling</p> <p>28. The set should contain torque limiting attachment 1.5mm with quick coupling</p> <p>29. The set should contain torque limiting attachment 1.2mm</p> <p>30. The set should contain torque limiting attachment 0.8mm with quick coupling</p> <p>31. The set should contain plate Bender Irons closed for 2.7/3.5mm plates</p> <p>32. The set should contain 1.25mm kirschner wire with trocar point 150mm-Pack of 10's</p> <p>33. The set should contain 2.0mm kirschner wire with trocar point 150mm Pack of 10's</p> <p>34. The set should contain 1.6mm kirschner wire with trocar point 150mm - Pack of 10's</p> <p>35. The set should contain Periosteal Elevator 6mm Curved</p> <p>36. The set should contain sharp hook-small taper</p> <p>37. The set should contain small hohmann retractor 8mm short narrow tip 160mm</p> <p>38. The set should contain hohmann retractor 15mm 160mm</p> <p>39. The set should contain 2.7/3.5mm Depth Gauge length 40-100</p> <p>40. The set should contain 2.5mm Drill Bit with quick coupling and length 150mm</p>	<p>7. Accountant of the Importer with UDIN. Self-attested photocopy of IEC (Import Export Certificate).</p> <p>8. Quality Certification of the quoted equipment as per NIT</p> <p>9. Copy of CDSCO Registration Certification.</p>
--	--	--	---	---

			41. The set should contain 2.8mm Drill Bit with quick coupling and length 110mm 42. The set should contain 3.5mm Drill Bit with quick coupling and length 105mm 43. The set should contain 2.7/3.5mm Depth Gauge with length 40-100mm 44. The set should contain 2.5mm Drill Bit with quick coupling and length 150mm 45. The set should contain 2.8mm Drill Bit with quick coupling and length 110mm 46. The set should contain 3.5mm Drill Bit with quick coupling and length 1.5mm 47. Aluminium case	
		SHARMA ENTERPRISES Importer: - M/S Ahuja sons Surgicals Manufacture: - INNOMED, INC - USA	1. The instrument should be made of stainless steel and medical grade polymers - polyaryletherketone, vulcanized ethylene propylene diene monomer and polypropylene. 2. All instrument grips should be made of vulcanized ethylene propylene diene monomer 3. The instruments should be able to maintain stiffness while withstanding high temperatures during sterilization 4. The set should contain 3.5 mm Neutral Sleeve Adapter for 3.5 Non-Locking Drill Guide 5. The set should contain 3.5mm Non-Locking Drill Guide 6. The set should contain 3.5 mm Polyaxial angle Drill Guide 7. The set should contain 2.8mm Threaded Guide for 3.5mm Screw for both Polyaxial angle locking and standard locking 8. The set should contain 2.7mm Neutral Sleeve Adapter for 2.7 Non-locking Drill Guide 9. The set should contain 2.7mm Polyaxial angle Drill Guide 10. The set should contain 2.0mm Threaded guide for 2.7mm screw for polyaxial angle and standard locking 11. The set should contain centering sleeve for k-wire with diameter 1.6mm with length 70mm 12. The set should contain 2.7/3.5mm Depth Gauge 0-60mm 13. The set should contain quick coupling 2.0mm Drill Bit with length 30mm 14. The set should contain quick coupling 2.0mm Drill Bit with length 60mm 15. The set should contain quick coupling 2.5mm Drill Bit with length 45mm 16. The set should contain quick coupling 2.5mm Drill Bit with length 80mm 17. The set should contain quick coupling 2.7mm Drill Bit with length 45mm 18. The set should contain quick coupling 2.8mm Drill Bit with length 45mm 19. The set should contain quick coupling 2.8mm Drill Bit with length 80mm 20. The set should contain quick coupling 3.5mm Drill Bit with length 60mm 21. The set should contain counter sink for 3.5mm cortex and 4.0mm cancellous bone crews 22. The set should contain Universal Screwdriver Handle 23. The set should contain 2.5mm Hexagonal Driver shaft 24. The set should contain stardrive	To submit: - 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.

			<p>screwdriver shaft with coupling</p> <p>25. The set should contain stardrive screwdriver shaft length 105mm</p> <p>26. The set should contain holding sleeve</p> <p>27. The set should contain 2.5mm torque limiting handle with quick coupling</p> <p>28. The set should contain torque limiting attachment 1.5mm with quick coupling</p> <p>29. The set should contain torque limiting attachment 1.2mm</p> <p>30. The set should contain torque limiting attachment 0.8mm with quick coupling</p> <p>31. The set should contain plate Bender Irons closed for 2.7/3.5mm plates</p> <p>32. The set should contain 1.25mm kirschner wire with trocar point 150mm-Pack of 10's</p> <p>33. The set should contain 2.0mm kirschner wire with trocar point 150mm Pack of 10's</p> <p>34. The set should contain 1.6mm kirschner wire with trocar point 150mm - Pack of 10's</p> <p>35. The set should contain Periosteal Elevator 6mm Curved</p> <p>36. The set should contain sharp hook-small taper</p> <p>37. The set should contain small hohmann retractor 8mm short narrow tip 160mm</p> <p>38. The set should contain hohmann retractor 15mm 160mm</p> <p>39. The set should contain 2.7/3.5mm Depth Gauge length 40-100</p> <p>40. The set should contain 2.5mm Drill Bit with quick coupling and length 150mm</p> <p>41. The set should contain 2.8mm Drill Bit with quick coupling and length 110mm</p> <p>42. The set should contain 3.5mm Drill Bit with quick coupling and length 105mm</p> <p>43. The set should contain 2.7/3.5mm Depth Gauge with length 40-100mm</p> <p>44. The set should contain 2.5mm Drill Bit with quick coupling and length 150mm</p> <p>45. The set should contain 2.8mm Drill Bit with quick coupling and length 110mm</p> <p>46. The set should contain 3.5mm Drill Bit with quick coupling and length 1.5mm</p> <p>47. Aluminium case</p>	
8.	Percutaneous Cerclage wire passer	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<p>1. The Instruments should be able to be used in Periprosthetic features of the femur, Subtrochanteric fractures, Hip and knee prostheses.</p> <p>2. The instruments should be able to be used in temporary reduction of fractures</p> <p>3. The instruments should enable Minimally invasive surgery</p> <p>4. The Cerclage wire passer should be available in size of 45mm & 60mm (5% deviation)</p> <p>5. The Cerclage passer should be designed as two separate halves to facilitates equential insertion through on small incision</p> <p>6. The instruments should have a cerclage twister which should be compatible to be used with wirers with diameter ranging from 0.8 to 1.5mm</p> <p>7. The cerclage twister should adapt force by scrolling the adjusting sleeve to prevent wire breakage</p> <p>8. The instrument set should contain tunnelling deives as per the size of the wire passer for 46mm & 60mm</p> <p>9. The instrument set contain bending</p>	<p>To submit: -</p> <p>1. Copy of catalogue of each quoted item (self-attested).</p> <p>2. Technical bid submission sheet duly filled (Annexures II).</p> <p>3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).</p> <p>4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd).</p> <p>5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21).</p> <p>6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN.</p> <p>7. Self-attested photocopy of IEC (Import Export Certificate).</p> <p>8. Quality Certification of the quoted equipment as per NIT</p> <p>9. Copy of CDSCO Registration Certification.</p>

			<p>pliers for bending the wires</p> <p>10. The set should incorporate two beaded trocars for Cerclage Passer</p> <p>11. The set should have a front cutter for cutting wires</p>	
		<p>SHARMA ENTERPRISES</p> <p>Importer:- M/S Ahuja sons Surgicals</p> <p>Manufacture:- INNOMED, INC - USA</p>	<p>1. The Instruments should be able to be used in Periprosthetic features of the femur, Subtrochanteric fractures, Hip and knee prostheses.</p> <p>2. The instruments should be able to be used in temporary reduction of fractures</p> <p>3. The instruments should enable Minimally invasive surgery</p> <p>4. The Cerclage wire passer should be available in size of 45mm & 60mm (5% deviation)</p> <p>5. The Cerclage passer should be designed as two separate halves to facilitates equential insertion through on small incision</p> <p>6. The instruments should have a cerclage twister which should be compatible to be used with wirers with diameter ranging from 0.8 to 1.5mm</p> <p>7. The cerclage twister should adapt force by scrolling the adjusting sleeve to prevent wire breakage</p> <p>8. The instrument set should contain tunnelling deives as per the size of the wire passer for 46mm & 60mm</p> <p>9. The instrument set contain bending pliers for bending the wires</p> <p>10. The set should incorporate two beaded trocars for Cerclage Passer</p> <p>11. The set should have a front cutter for cutting wires</p>	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
9.	Large Fragment	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<p>1. The instrument set should contain Drill Bit with diameter 3.2mm, Length 145/120mm for quick coupling</p> <p>2. The instrument set should contain Drill Bit with diameter 4.5mm, length 147/120mm for quick coupling</p> <p>3. The instrument set should contain large countersink with length 180mm</p> <p>4. The instrument set should contain T- Handle with quick coupling of length 80mm</p> <p>5. The instrument set should contain Tap for 4.5mm Cortex Screws with length 70/125mm</p> <p>6. The instrument set should contain Tap for 6.5mm Cancellous Bone Screws Length 195mm</p> <p>7. The instrument set should contain Double Drill sleeve 4.5/3.2</p> <p>8. The instrument set should contain Double Drill sleeve 4.5/3.2 length 80mm</p> <p>9. The instrument set should contain Double Drill sleeve 6.5/3.2</p> <p>10. The instrument set should contain large Screwdriver Shaft with hexagonal recess & length 100mm</p> <p>11. The instrument set should contain large screwdriver with hexagonal recess & recess & groove with length 240mm</p> <p>12. The instrument set should contain large Holding sleeve of length 120mm</p> <p>13. The instrument set should contain Depth Guage for 4.5 to 6.5mm screws</p> <p>14. The instrument set should contain sharp Hook, length 155mm</p>	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.

			15. The instrument set should contain Articulated Tension Device 16. The instrument set should contain combination wrench, 11mm, length 140mm 17. The instrument set should contain DPC Drill Sleeve 4.5 18. The instrument set should contain LC-DCP Drill Sleeve 4.5 19. The instrument set should contain Universal Drill Sleeve 4.5 20. The instrument set should contain Bending Template for DCP 4.5 and LC-DCP 4.5, Length 210mm 21. The instrument set should contain Bending Template for DCP 4.5 and LC-DCP 4.5, Length 155mm 22. The instrument set should contain Ratchet Wrench, 11mm, Length 140mm 23. The instrument set should contain Torque-Limiting Screwdriver, Length 25.5mm for 3.5mm hexagon 24. The instrument set should contain Drill bit with diameter 3.5mm for metal 25. The instrument set should contain Conical extraction screw for 4.5/6.5mm screws 26. The instrument set should contain Drill Bit with diameter 4.3mm, Length 221mm 27. Screwdriver Shaft 3.5, self-retaining, L 110mm 28. The instrument set should contain Threaded LCP Drill Guide for 4.3mm Drill Bit 29. The instrument set should contain screw Holding sleeve for LCP 4.5/5.0 30. The instrument set should contain Handle with Quick coupling 31. Radiolucent Pelvic Retractor, blunt, length 274 mm, Aluminium 32. Radiolucent Hohmann Retractor, width 18 mm, length 240 mm Aluminium 33. Radiolucent Pelvic Retractor, medium, length 268 mm, Aluminium 34. Radiolucent Pelvic Retractor, large, length 323 mm, Aluminium 35. Radiolucent Hohmann Retractor, width 35 mm, length 275 mm, Aluminium 36. Radiolucent Hohmann Retractor, width 267 mm, Aluminium 37. Reduction Forceps with points Narrow Soft Lock, length 146 mm 38. Bone Holding Forcep , Self Centering , Soft Lock 191 mm 39. Reduction Forceps with Narrow Soft Lock Length 127 mm 40. Reduction Forceps with Point Soft Lock Length 222 mm 41. Reduction Forceps Toothed, Soft Lock, Length 250mm 42. Bone Holding Forceps, Self Centering , Soft Lock 239mm 43. Aluminium Case	
		SHARMA ENTERPRISES Manufacturer: Nebula Surgical Pvt Ltd-Gujrat - 360005	1. The instrument set should contain Drill Bit with diameter 3.2mm, Length 145/120mm for quick coupling 2. The instrument set should contain Drill Bit with diameter 4.5mm, length 147/120mm for quick coupling 3. The instrument set should contain large countersink with length 180mm 4. The instrument set should contain T- Handle with quick coupling of length 80mm 5. The instrument set should contain Tap for 4.5mm Cortex Screws with length 70/125mm 6. The instrument set should contain Tap for 6.5mm Cancellous Bone	To submit: - 1. Compliance Sheet for each equipment. 2. Technical bid submission sheet duly filled 3. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Principal manufacturer as per proforma duly notarized. 4. Model of each quoted instruments. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with

			<p>Screws Length 195mm</p> <p>7. The instrument set should contain Double Drill sleeve 4.5/3.2</p> <p>8. The instrument set should contain Double Drill sleeve 4.5/3.2 length 80mm</p> <p>9. The instrument set should contain Double Drill sleeve 6.5/3.2</p> <p>10. The instrument set should contain large Screwdriver Shaft with hexagonal recess & length 100mm</p> <p>11. The instrument set should contain large screwdriver with hexagonal recess & recess & groove with length 240mm</p> <p>12. The instrument set should contain large Holding sleeve of length 120mm</p> <p>13. The instrument set should contain Depth Gauge for 4.5 to 6.5mm screws</p> <p>14. The instrument set should contain sharp Hook, length 155mm</p> <p>15. The instrument set should contain Articulated Tension Device</p> <p>16. The instrument set should contain combination wrench, 11mm, length 140mm</p> <p>17. The instrument set should contain DPC Drill Sleeve 4.5</p> <p>18. The instrument set should contain LC-DCP Drill Sleeve 4.5</p> <p>19. The instrument set should contain Universal Drill Sleeve 4.5</p> <p>20. The instrument set should contain Bending Template for DCP 4.5 and LC-DCP 4.5, Length 210mm</p> <p>21. The instrument set should contain Bending Template for DCP 4.5 and LC-DCP 4.5, Length 155mm</p> <p>22. The instrument set should contain Ratchet Wrench, 11mm, Length 140mm</p> <p>23. The instrument set should contain Torque-Limiting Screwdriver, Length 25.5mm for 3.5mm hexagon</p> <p>24. The instrument set should contain Drill bit with diameter 3.5mm for metal</p> <p>25. The instrument set should contain Conical extraction screw for 4.5/6.5mm screws</p> <p>26. The instrument set should contain Drill Bit with diameter 4.3mm, Length 221mm</p> <p>27. Screwdriver Shaft 3.5, self-retaining, L 110mm</p> <p>28. The instrument set should contain Threaded LCP Drill Guide for 4.3mm Drill Bit</p> <p>29. The instrument set should contain screw Holding sleeve for LCP 4.5/5.0</p> <p>30. The instrument set should contain Handle with Quick coupling</p> <p>31. Radiolucent Pelvic Retractor, blunt, length 274 mm, Aluminium</p> <p>32. Radiolucent Hohmann Retractor, width 18 mm, length 240 mm Aluminium</p> <p>33. Radiolucent Pelvic Retractor, medium, length 268 mm, Aluminium</p> <p>34. Radiolucent Pelvic Retractor, large, length 323 mm, Aluminium</p> <p>35. Radiolucent Hohmann Retractor, width 35 mm, length 275 mm, Aluminium</p> <p>36. Radiolucent Hohmann Retractor, width 267 mm, Aluminium</p> <p>37. Reduction Forceps with points Narrow Soft Lock, length 146 mm</p> <p>38. Bone Holding Forcep, Self Centering, Soft Lock 191 mm</p> <p>39. Reduction Forceps with Narrow Soft Lock Length 127 mm</p> <p>40. Reduction Forceps with Point Soft Lock Length 222 mm</p> <p>41. Reduction Forceps Toothed, Soft Lock, Length 250mm</p>	<p>UDIN (2018-19, 2019-20 and 2020-21).</p> <p>7. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN.</p> <p>8. Latest GST Returns of the Bidder</p> <p>9. Declaration of bidder regarding acceptance Bid for terms & conditions</p> <p>10. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.</p>
--	--	--	--	---

			42. Bone Holding Forceps, Self Centering, Soft Lock 239mm 43. Aluminium Case	
10.	Specification for Hand Surgery System	Zero Bidder		
11.	Wire instruments Set	<p>SHARMA ENTERPRISES Importer:- M/S Ahuja sons Surgicals</p> <p>Manufacture:- INNOMED, INC -USA</p>	<ol style="list-style-type: none"> 1. Drill Bit, 2.0 mm dia, L 120/75mm for quick coupling 2. Triple Drill Guide 2.0 with 3 holes, opposite side 1 hole 3. Wire passer, 45mm bending diameter 4. Wire passer, 70mm bending diameter 5. Wire tightener with handle and two pegs, L 240mm 6. Holding Forceps for Cerclage Wires, L 170mm 7. Wire Bending Pliers, L 155mm 8. Parallel Pliers, flat nosed, L 160mm 9. Wire Cutter, large, L 220mm 10. Wire Cutter, short, L 175mm 11. Bending iron, for kirschner Wires 1.25 to 2.5mm dia, L 120mm 12. Wire mount 13. Cerclage wire, 1.0mm dia, with eye L 280mm 14. Cerclage wire, 1.25mm dia, with eye L 280mm 15. Wire Coil, 1.0mm dia, L 10m 16. Wire Coil, 1.25mm dia L 10m 17. Kirschner Wire, 1.0m dia with trocar tip, L 150mm(packet of 10) 18. Kirschner Wire, 1.25m dia with trocar tip, L 150mm (packet of 10) 19. Kirschner Wire, 1.6m dia with trocar tip, L 150mm (packet of 10) 20. Kirschner Wire, 2.0m dia with trocar tip, L 150mm (packet of 10) 21. Kirschner Wire, 2.5m dia with trocar tip, L 150mm (packet of 10) 22. Sterilization Case 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<ol style="list-style-type: none"> 1. Drill Bit, 2.0 mm dia, L 120/75mm for quick coupling 2. Triple Drill Guide 2.0 with 3 holes, opposite side 1 hole 3. Wire passer, 45mm bending diameter 4. Wire passer, 70mm bending diameter 5. Wire tightener with handle and two pegs, L 240mm 6. Holding Forceps for Cerclage Wires, L 170mm 7. Wire Bending Pliers, L 155mm 8. Parallel Pliers, flat nosed, L 160mm 9. Wire Cutter, large, L 220mm 10. Wire Cutter, short, L 175mm 11. Bending iron, for kirschner Wires 1.25 to 2.5mm dia, L 120mm 12. Wire mount 13. Cerclage wire, 1.0mm dia, with eye L 280mm 14. Cerclage wire, 1.25mm dia, with eye L 280mm 15. Wire Coil, 1.0mm dia, L 10m 16. Wire Coil, 1.25mm dia L 10m 17. Kirschner Wire, 1.0m dia with trocar tip, L 150mm(packet of 10) 18. Kirschner Wire, 1.25m dia with trocar tip, L 150mm (packet of 10) 19. Kirschner Wire, 1.6m dia with trocar tip, L 150mm (packet of 10) 20. Kirschner Wire, 2.0m dia with trocar tip, L 150mm (packet of 10) 21. Kirschner Wire, 2.5m dia with trocar tip, L 150mm (packet of 10) 22. Sterilization Case 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
12.	HIP PRESERVATION INSTRUMENTS SET	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson &</p>	<ol style="list-style-type: none"> 1. The Hip preservation surgery Set contain semi-radiolucent retractor, osteostomes, chisels and femoral head templates to assist the surgeon with periacetabular osteotomy and hip 	<p>To submit:-</p> <p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled

9

15

		<p>Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<p>impingement procedures.</p> <ol style="list-style-type: none"> The instrument set should contains a blunt pelvic Retractor The instrument set should contain Bone Lever with wide tip and width 22 mm with length 250mm The Instrument set should contain long narrow tip Bone lever with 18mm and length 235mm Long narrow tip wide tip Bone lever with width 24 mm and length 270mm The instruments set should contains a chisel Handle with the following chisel blades Chisel Blade , width 10mm Chisel Blade, width 16mm Chisel Blade, width 5mm Long Chisel Blade, width 10mm Long chisel Blade, width 16mm Long chisel Blade, width 25 mm Long chisel Blade, width 5.0mm The instrument set should contain Pelvic Osteotomy chisel with width 15 mm and length 304/134mm The instrument set should contain Pelvic Osteotomy Chisel with width 20 mm and length 304/134 mm The instrument set should contain wide tip Bone lever with width 35 mm and length 284mm The instrument set should contain curved Pelvic Osteotomy chisel with widths 15 mm and 20mm and length 309 mm The instrument set should contain blunt aluminium Radiolucent Pelvic Retractor with length 274mm The instrument set should contain the Radiolucent hohmann retractors with the following Sizes: Width 35 mm, length 275 mm Width 18 mm, length 240 mm Width 24 mm, Length 267 mm The instrument set should contains aluminium Radiolucent Pelvic Retractor in medium and large size wit lengths 268 mm and 323mm The instrument set should contain aluminium Radiolucent Hohann Retractor with width 22 mm and length 250mm Pelvic Osteotome, straight 15 mm Pelvic Osteotome, straight 20 mm Pelvic Osteotome, angled 15 mm Pelvic Osteotome, angledt 20 mm Pelvic Osteotome, short angled 15 mm Pelvic Osteotome, bayonet - shaped , 15 mm Pelvic Osteotome, bayonet - shaped , 20 mm The instrument set should contain the following templates: Template for femoral Head dia 42 mm Template for femoral Head dia 44 mm Template for femoral Head dia 46 mm Template for femoral Head dia 48 mm Template for femoral Head dia 50 mm Template for femoral Head dia 52 mm Template for femoral Head dia 54 mm Template for femoral Head dia 56 mm Template for femoral Head dia 58 mm Graphic Case for Hip Preservation set 	<p>(Annexures II).</p> <ol style="list-style-type: none"> Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. Self-attested photocopy of IEC (Import Export Certificate). Quality Certification of the quoted equipment as per NIT Copy of CDSCO Registration Certification.
		<p>SHARMA ENTERPRISES Importer:- M/S Ahuja sons Surgicals</p> <p>Manufacture:- INNOMED, INC -USA</p>	<ol style="list-style-type: none"> The Hip preservation surgery Set contain semi-radiolucent retractor, osterostomes, chisels and femoral head templates to assits the surgeon with periacetabular osteotomy and hip impingement procedures. The instrument set should contains a blunt pelvic Retractor The instrument set should contain Bone Lever with wide tip and width 22 mm with length 250mm The Instrument set should contain long narrow tip Bone lever with 18mm and length 235mm 	<p>To submit: -</p> <ol style="list-style-type: none"> Copy of Catalogue of the Quoted product Compliance Sheet for each equipment. Technical bid submission sheet duly filled Model of each quoted instruments. Copy of Catalogue of the Quoted product Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).

			<ol style="list-style-type: none"> 5. Long narrow tip wide tip Bone lever with width 24 mm and length 270mm 6. The instruments set should contains a chisel Handle with the following chisel blades 7. Chisel Blade , width 10mm 8. Chisel Blade, width 16mm 9. Chisel Blade, width 5mm 10. Long Chisel Blade, width 10mm 11. Long chisel Blade, width 16mm 12. Long chisel Blade, width 25 mm 13. Long chisel Blade, width 5.0mm 14. The instrument set should contain Pelvic Osteotomy chisel with width 15 mm and length 304/134mm 15. The instrument set should contain Pelvic Osteotomy Chisel with width 20 mm and length 304/134 mm 16. The instrument set should contain wide tip Bone lever with width 35 mm and length 284mm 17. The instrument set should contain curved Pelvic Osteotomy chisel with widths 15 mm and 20mm and length 309 mm 18. The instrument set should contain blunt aluminium Radiolucent Pelvic Retractor with length 274mm 19. The instrument set should contain the Radiolucent hohmann retractors with the following Sizes: 20. Width 35 mm, length 275 mm 21. Width 18 mm, length 240 mm 22. Width 24 mm, Length 267 mm 23. The instrument set should contains aluminium Radiolucent Pelvic Retractor in medium and large size wit lengths 268 mm and 323mm 24. The instrument set should contain aluminium Radiolucent Hohann Retractor with width 22 mm and length 250mm 25. Pelvic Osteotome, straight 15 mm 26. Pelvic Osteotome, straight 20 mm 27. Pelvic Osteotome, angled 15 mm 28. Pelvic Osteotome, angledt 20 mm 29. Pelvic Osteotome, short angled 15 mm 30. Pelvic Osteotome, bayonet - shaped , 15 mm 31. Pelvic Osteotome, bayonet - shaped , 20 mm 32. The instrument set should contain the following templates: 33. Template for femoral Head dia 42 mm 34. Template for femoral Head dia 44 mm 35. Template for femoral Head dia 46 mm 36. Template for femoral Head dia 48 mm 37. Template for femoral Head dia 50 mm 38. Template for femoral Head dia 52 mm 39. Template for femoral Head dia 54 mm 40. Template for femoral Head dia 56 mm 41. Template for femoral Head dia 58 mm 42. Graphic Case for Hip Preservation 	<ol style="list-style-type: none"> 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
--	--	--	---	--

13.	DHS/DCS Instrument Set	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<p>set</p> <ol style="list-style-type: none"> 1. Material -Stainless Steel (instrument) 2. Grade-ISO 5832-1 3. The instrument should have DHS/DCS Threaded Guide Wire , 2.5mm dia.. L 230/5mm 4. The instrument should have Angled Guide 135 150° 5. The instrument should have DHS/DCS Direct Measuring Device 6. The instrument should Wrench for one step insertion Length 230mm 7. The instrument should have T- Handle with quick coupling, Length 80mm 8. The instrument should have DHS Triple Reamer 9. The instrument should have Impactor for one-step insertion, L 260mm 10. The instrument should have DHS/ DCS Tap Length 220Mm 11. The instrument should have Locking Centering Sleeve 12. The instrument should have coupling Screw, cannulated 13. The instrument should have Triple Reamer 14. The instrument should have tapered ends to allow for submuscular insertion of the plate 15. The instrument should be compatible with Trochanter Stabilizing plates 16. The triple reamer should be detachable in 3 parts - Reamer for screw. Reamer for DHS.DCS plate barrel and Nut 17. The should use one step surgical technique for the insertion of lag screw coupling for DHS/DCS 18. Aluminum Case. 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested) . 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		<p>SHARMA ENTERPRISES</p> <p>Importer:- . M/S Ahuja sons Surgicals</p> <p>Manufacture:- INNOMED, INC -USA</p>	<ol style="list-style-type: none"> 1. Material -Stainless Steel (instrument) 2. Grade-ISO 5832-1 3. The instrument should have DHS/DCS Threaded Guide Wire , 2.5mm dia.. L 230/5mm 4. The instrument should have Angled Guide 135 150° 5. The instrument should have DHS/DCS Direct Measuring Device 6. The instrument should Wrench for one step insertion Length 230mm 7. The instrument should have T- Handle with quick coupling, Length 80mm 8. The instrument should have DHS Triple Reamer 9. The instrument should have Impactor for one-step insertion, L 260mm 10. The instrument should have DHS/ DCS Tap Length 220Mm 11. The instrument should have Locking Centering Sleeve 12. The instrument should have coupling Screw, cannulated 13. The instrument should have Triple Reamer 14. The instrument should have tapered ends to allow for submuscular insertion of the plate 15. The instrument should be compatible with Trochanter Stabilizing plates 16. The triple reamer should be detachable in 3 parts - Reamer for screw. Reamer for DHS.DCS plate barrel and Nut 17. The should use one step surgical technique for the insertion of lag screw coupling for DHS/DCS 18. Aluminum Case. 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		<p>DKSORTHO AND SHARMA PHARMACEUTICAL PRIVATE LIMITED (Manufacturer / Bidder)</p>	<ol style="list-style-type: none"> 1. The instrument should have DHS/DCS Threaded Guide Wire , 2.5mm dia.. L 230/5mm 2. The instrument should have Angled Guide 135 150° 3. The instrument should have DHS/DCS Direct Measuring Device 4. The instrument should have T- Handle with quick coupling, Length 80mm 5. The instrument should have DHS 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item(self attested) . 3. The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi) . Please Clarify along with relevant documents 4. Compliance Sheet for each equipment 5. Client base with reference of the supply order on letter

			6. Triple Reamer The instrument should have DHS/ DCS Tap Length 220Mm 7. The instrument should have Locking Centering Sleeve 8. The instrument should have coupling Screw, cannulated 9. The instrument should have Triple Reamer 10. The instrument should be compatible with Trochanter Stabilizing plates 11. The triple reamer should be detachable in 3 parts – Reamer for screw. Reamer for DHS.DCS plate barrel and Nut 12. The should use one step surgical technique for the insertion of lag screw coupling for DHS/DCS	head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Latest GST Returns of the Bidder 7. Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled 8. Copy of CDSCO Registration Certification.
14	General Instrument and Chisel, Impactor Set	ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080 Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	1. Bone Hook sharp, small, L 230mm 2. Bone Hook, sharp, medium, L 230mm 3. Retractor, 8mm wide, short narrow tip, L 220mm 4. Retractor, 18mm wide, short narrow tip, L 235 5. Retractor, 24mm wide, long and wide tip, L 270mm 6. Periosteal Elevator, curved shaft, 14mm wide, L 200mm 7. Periosteal Elevator, round edge, 6mm wide, L 200mm 8. Periosteal Elevator straight shaft, 14mm wide, L 200mm 9. Hammer 500g, L 185mm 10. Chisel Handle, L 185mm 11. Chisel Blade, 10mm wide thickness 0.9mm, L 81 mm 12. Chisel Blade 16mm wide, thickness 0.9mm, L 81mm 13. Chisel Blade 25mm wide, thickness 0.9mm, L 81mm 14. Gouge, curved, for cancellous bone graft harvest, 10mm wide, L 250mm 15. Handle with quick coupling, L 150mm 16. Cancellous Bone impactor, 6.0mm dia, round L 140mm 17. Cancellous Bone impactor, 6.0mm dia, flattened, L 140mm 18. Cancellous Bone impactor, 8.0mm dia, round L 140mm 19. Cancellous Bone impactor, rectangular 6x16mm, L 140mm 20. Cancellous Bone impactor, rectangular 10x20mm, L 140mm 21. Cancellous Bone impactor, rectangular 10x30mm, L 140mm 22. Gouge, curved, 10mm wide, L 140mm 23. Gouge, curved, 15mm wide, L 140mm 24. Chisel, flat, straight, 16mm wide, L 140mm	To submit: - 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		SHARMA ENTERPRISES Importer: - M/S Ahuja sons Surgicals Manufacture: - INNOMED, INC -USA	1. Bone Hook sharp, small, L 230mm 2. Bone Hook, sharp, medium, L 230mm 3. Retractor, 8mm wide, short narrow tip, L 220mm 4. Retractor, 18mm wide, short narrow tip, L 235 5. Retractor, 24mm wide, long and wide tip, L 270mm 6. Periosteal Elevator, curved shaft, 14mm wide, L 200mm 7. Periosteal Elevator, round edge, 6mm wide, L 200mm 8. Periosteal Elevator straight shaft, 14mm wide, L 200mm 9. Hammer 500g, L 185mm 10. Chisel Handle, L 185mm 11. Chisel Blade, 10mm wide thickness 0.9mm, L 81 mm 12. Chisel Blade 16mm wide, thickness 0.9mm, L 81mm 13. Chisel Blade 25mm wide, thickness 0.9mm, L 81mm 14. Gouge, curved, for cancellous bone graft harvest, 10mm wide, L 250mm 15. Handle with quick coupling, L 150mm 16. Cancellous Bone impactor, 6.0mm dia, round L 140mm 17. Cancellous Bone impactor, 6.0mm dia, flattened, L 140mm 18. Cancellous Bone impactor, 8.0mm dia, round L 140mm	To submit: - 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz.

			19. Cancellous Bone impactor, rectangular 6x16mm, L 140mm 20. Cancellous Bone impactor, rectangular 10x20mm, L 140mm 21. Cancellous Bone impactor, rectangular 10x30mm, L 140mm 22. Gouge, curved, 10mm wide, L 140mm 23. Gouge, curved, 15mm wide, L 140mm 24. Chisel, flat, straight, 16mm wide, L 140mm	12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		DKSORTHO AND SHARMA PHARMACEUTICAL PRIVATE LIMITED (Manufacturer / Bidder)	1. Bone Hook sharp, small, L 230mm 2. Bone Hook, sharp, medium, L 230mm 3. Retractor, 8mm wide, short narrow tip, L 220mm 4. Retractor, 18mm wide, short narrow tip, L 235 5. Retractor, 24mm wide, long and wide tip, L 270mm 6. Periosteal Elevator, curved shaft, 14mm wide, L 200mm 7. Periosteal Elevator, round edge, 6mm wide, L 200mm 8. Periosteal Elevator straight shaft, 14mm wide, L 200mm 9. Hammer 500g, L 185mm 10. Chisel Handle, L 185mm 11. Gouge, curved, for cancellous bone graft harvest, 10mm wide, L 250mm 12. Handle with quick coupling, L 150mm 13. Cancellous Bone impactor, 6.0mm dia, round L 140mm 14. Cancellous Bone impactor, 6.0mm dia, flattened, L 140mm 15. Cancellous Bone impactor, 8.0mm dia, round L 140mm 16. Cancellous Bone impactor, rectangular 6x16mm, L 140mm 17. Cancellous Bone impactor, rectangular 10x20mm, L 140mm 18. Cancellous Bone impactor, rectangular 10x30mm, L 140mm 19. Gouge, curved, 10mm wide, L 140mm 20. Gouge, curved, 15mm wide, L 140mm 21. Chisel, flat, straight, 16mm wide, L 140mm	To submit: - 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item(self attested) . 3. The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi). Please Clarify along with relevant documents 4. Compliance Sheet for each equipment 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Latest GST Returns of the Bidder 7. Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled 8. Copy of CDSCO Registration Certification.
15	CSS 4.5mm instrument Set	DKSORTHO AND SHARMA PHARMACEUTICAL PRIVATE LIMITED (Manufacturer / Bidder)	1. Drill Bit Dia 1.5L110/85 2flute 2. Drill Bit Dia 3.2/1.75 cann L170/140 4flute 3. T-Handle w/Quick-Coupl 4. Double Drill Guide-Coupl 5. Parallel- Guide f/ Guide-Wires Dia 1.6 adjust 6. Trocar Dia 1.6 Dia 4.5 7. Drill Sleeve 3.2/1.6 Dia 4.5 8. Protect - sleeve 9.5/7 Dia 4.5 9. Scr Driver - hex-cannDia 4.5 10. Direct Measur-Device Dia 4.5 11. Hold Clip f/ Washers 12. Scr Forceps self hold L 85 13. Guide wire Dia 1.6w/Thread-tip w/trocar L15	To submit: - 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item(self attested) . 3. The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi) . Please Clarify along with relevant documents 4. Compliance Sheet for each equipment 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Latest GST Returns of the Bidder 7. Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled 8. Copy of CDSCO Registration Certification.
16.	Css 6.5mm Instrument Set	DKSORTHO AND SHARMA PHARMACEUTICAL PRIVATE LIMITED (Manufacturer / Bidder)	1. Drill Bit Dia 5 cann L 300/250 3flute 2. Parallel Guide f/ Guide -wires Dia 2.8 adjust 3. Trocar Dia 2.8 4. Protect- Sleeve 12/8.5 5. Drill Sleeve 8.5/2.8 6. Protect- Sleeve 15.5/13 7. Scr Driver- hex-cannDia 6.5+7.3 8. Hold- Sleeve large 9. Clean-Style Dia 2.8 10. Direct Measure- Device f/ Guide Wires Dia 2+2.8 11. Scr Forceps self-hold L85 12. Guide Wire Dia 2.8w/thread-tip w/trocar L 3	To submit: - 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item(self attested) . 3. The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi) . Please Clarify along with relevant documents 4. Compliance Sheet for each equipment 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Latest GST Returns of the Bidder 7. Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled 8. Copy of CDSCO Registration Certification.

17	Bone Forceps Range	<p>ORTHOFIX (Bidder)</p> <p>Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080</p> <p>Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380</p>	<ol style="list-style-type: none"> 1. Reduc-Forceps w/Points narrow ratchet-lock 2. Reduc-Forceps w/Points wide ratchet-lock 3. Reduc-Forceps w/Points ratchet-lock L180 4. Reduc-Forceps toothed speed-lock L170 5. Reduc-Forceps w/Points speed-lock L130 6. Reduc-Forceps large w/Points speed-lock 7. Reduc-Forceps toothed speed lock L140 8. Reduc-Forceps w/Points ratchet-lock L160 9. Reduc-Forceps w/Points ratchet-lock L130 10. Reduc-Forceps toothed ratchet-lock L140 11. Bone Holding Forceps self -center speed -1 12. Bone Holding-Forceps self -center speed -1 13. Bone Holding -Forceps self-center speed-1 14. Hold-Forceps f/Tib-Edge Fragn L210 15. Hold-Forcepsw/Ball ratchet-lock L180 16. Bone-spreader Speed-lock L140 17. Stagbeetle-Forceps ratchet-Lock L120 18. Patella Forceps speed Lock L175 19. Malleolar Forceps-lock L210 20. Reduc-Forceps toothed speed-lock L170 21. Plate Holding Forceps, size 2 22. Plate Holding Forceps, size 0 23. Reduction Forceps with Points Narrow Soft Lock, Length 146 MM 24. Reduction Forceps with Points Narrow Soft Lock Length 127MM 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of catalogue of each quoted item (self-attested). 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd). 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		<p>SHARMA ENTERPRISES</p> <p>Importer: - M/S Ahuja sons Surgicals</p> <p>Manufacture: - INNOMED, INC -USA</p>	<ol style="list-style-type: none"> 1. Reduc-Forceps w/Points narrow ratchet-lock 2. Reduc-Forceps w/Points wide ratchet-lock 3. Reduc-Forceps w/Points ratchet-lock L180 4. Reduc-Forceps toothed speed-lock L170 5. Reduc-Forceps w/Points speed-lock L130 6. Reduc-Forceps large w/Points speed-lock 7. Reduc-Forceps toothed speed lock L140 8. Reduc-Forceps w/Points ratchet-lock L160 9. Reduc-Forceps w/Points ratchet-lock L130 10. Reduc-Forceps toothed ratchet-lock L140 11. Bone Holding Forceps self -center speed -1 12. Bone Holding-Forceps self -center speed -1 13. Bone Holding -Forceps self-center speed-1 14. Hold-Forceps f/Tib-Edge Fragn L210 15. Hold-Forcepsw/Ball ratchet-lock L180 16. Bone-spreader Speed-lock L140 17. Stagbeetle-Forceps ratchet-Lock L120 18. Patella Forceps speed Lock L175 19. Malleolar Forceps-lock L210 20. Reduc-Forceps toothed speed-lock L170 21. Plate Holding Forceps, size 2 22. Plate Holding Forceps, size 0 23. Reduction Forceps with Points Narrow Soft Lock, Length 146 MM 24. Reduction Forceps with Points Narrow Soft Lock Length 127MM 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		<p>DKSORTHO AND SHARMA PHARMACEUTICAL PRIVATE LIMITED</p>	<ol style="list-style-type: none"> 1. Reduc-Forceps w/Points narrow ratchet-lock 2. Reduc-Forceps w/Points wide ratchet-lock 3. Reduc-Forceps w/Points ratchet-lock L180 4. Reduc-Forceps toothed speed-lock L170 5. Reduc-Forceps w/Points speed-lock L130 6. Reduc-Forceps large w/Points speed-lock 7. Reduc-Forceps toothed speed lock L140 8. Reduc-Forceps w/Points ratchet-lock 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item (self attested). 3. The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi). Please Clarify along with relevant documents 4. Compliance Sheet for each equipment 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of

21

			L160 9. Reduc-Forceps w/Points ratchet-lock L130 10. ReducForceps toothed ratchet-lock L140 11. Bone Holding Forceps self-center speed -1 12. Bone Holding-Forceps self-center speed -1 13. Bone Holding -Forceps self-center speed-1 14. Hold-Forceps f/Tib-Edge Fragm L210 15. Hold-Forcepsw/Ball ratchet-lock L180 16. Bone-spreader Speed-lock L140 17. Stagbeetle-Forceps ratchet-Lock L120 18. Patella Forceps speed Lock L175 19. Malleolar Forceps-lock L210 20. Reduc-Forceps toothed speed-lockL170 21. Plate Holding Forceps, size 2 22. Plate Holding Forceps, size 0 23. Reduction Forceps with Points Narrow Soft Lock, Length 146 MM 24. Reduction Forceps with Points Narrow Soft Lock Length 127MM	reference supply order and satisfactory performance needs to be attached). 6. Latest GST Returns of the Bidder 7. Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled 8. Copy of CDSCO Registration Certification.
18	Specification - Foot Distractor	ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080 Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	1. Tray for compression/Distractor Device, for Orthopaedic Foot Instruments, without Lid, without contents 2. LID F/Modular Insert Size 1/4 3. Compression/Distractor Device for Foot and Ankle 4. Holding Sleeve Ø2.5 mm for Compression/Distractor Device for Foot and Ankle 5. Holding Sleeve Ø 2.5 mm for Compression/Distractor Device for Foot and Ankle 6. Holding Sleeve Ø 3.0 mm for compression/Distractor Device for Foot and Ankle 7. Holding Sleeve Tightening Device for Compression/Distractor Device 8. K-Wire Ø2.5 L150 SST 10U 9. K-Wire Ø2.5 Thread- Tip L 150/15 SST10U 10. K-Wire Ø2 Thread-Tip L 150/15 SST 10U	To submit: - 1. Copy of catalogue of each quoted item (self-attested) . 2. Technical bid submission sheet duly filled (Annexures II). 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. 5. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 6. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 7. Self-attested photocopy of IEC (Import Export Certificate). 8. Quality Certification of the quoted equipment as per NIT 9. Copy of CDSCO Registration Certification.
		SHARMA ENTERPRISES Importer: - M/S Ahuja sons Surgicals Manufacture: - INNOMED, INC -USA	1. Tray for compression/Distractor Device, for Orthopaedic Foot Instruments, without Lid, without contents 2. LID F/Modular Insert Size 1/4 3. Compression/Distractor Device for Foot and Ankle 4. Holding Sleeve Ø2.5 mm for Compression/Distractor Device for Foot and Ankle 5. Holding Sleeve Ø 2.5 mm for Compression/Distractor Device for Foot and Ankle 6. Holding Sleeve Ø 3.0 mm for compression/Distractor Device for Foot and Ankle 7. Holding Sleeve Tightening Device for Compression/Distractor Device 8. K-Wire Ø2.5 L150 SST 10U 9. K-Wire Ø2.5 Thread- Tip L 150/15 SST10U 10. K-Wire Ø2 Thread-Tip L 150/15 SST 10U	To submit: - 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the

				said firm along with complete documentary evidence.
19	Specification for Moreland type cement Revision System	Zero bidder		
20	Specification of MIS Joint Replacement Instrument System	Zero bidder		
21	Specification for MIS Acetabular Reamer System	Zero bidder		
22	Specification fore Control cable instruments	Zero bidder		
23	Technical Specification of HP Moreland	Zero bidder		
24	ACL-PCL RECONSTR UCTION COMPREH ENSIVE INSTRUME NTS SET	SHARMA ENTERPRISES Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	<ol style="list-style-type: none"> 1. Drill Guide 2. Angled Bullet, 2.4mm ID 3. ACL foot print aimer 4. PCL Tibial Aimer 5. ACL/PCL Femoral Aimer 6. Angled Bullet, 3.5mm ID 7. Drill Bit Passing Pin, Dia, 2.4mm: Len, 250mm 8. Drill Tip Graft passing Pin, (Beadth Pin) Dia, 2.4mm: Len, 355mm 9. Tibial Reamer, Dia,6.0mm/7.0mm/8.0mm/9.0m m Cannulated 10. Tibial Reamer, Dia,6.5mm/7.5mm/8.5mm/9.5m m Cannulated 11. Femoral Reamer, Dia,6.0mm/7.0mm/8.0mm/9.0m m Cannulated 12. Femoral Reamer, Dia,6.5mm/7.5mm/8.5mm/9.5m m Cannulated 13. Bone Tunnel Pug 9-10mm 14. Tunnel Dilator 6mm /7.7mm/8mm/9mm/10mm 15. T-Handle for Tunnel Dialators 16. Endoscopic Reamer, Dia, 4.7mm, Cannulated 17. Screwdriver for Bio Interference Screws Cannulated 18. Screwdriver for PEEK Interference Screws Cannulated 19. Guide Wire for interference Screws Dia, 1.2mm, Len 300 (NitinoL) 20. Depth Gauge for ACL/PCL reconstruction 21. Transportal Femoral Aimer 7mm 22. Arthroscopic Probe 23. Mirofracture Awl 30° 24. Graft Prep Station includes Graft Board, Sliding Base, Button Holder with scale, Graft Grasper Clamp-Left & Right and Suture Holder with tensioner 25. Graft Sizing Block for Dia, 6-7-8-9-10-11-12mm 26. Graft Sizing Block for Dia, 6.5-7-5-8.5-10.5-11.5mm 27. Tendon stripper, Closed end 7mmk 28. Tendon Stripper, Open and 6mm 29. Notchplasty Curette 30. PCL Rasp 31. PCL Hook 32. Suture Retriever 33. Tissue Grasper 34. Guide Sleeve for al Reamer 35. Retrograde reamer for all inside ACL (7mm/8mm/9mm/10mm) 36. Graphic Case for ACL-PCL Reconstruction 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Model of each quoted instruments. 5. Copy of Catalogue of the Quoted product 6. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 7. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 8. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 9. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 10. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). 11. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 12. Latest GST Returns of the Bidder 13. Copy of CDSCO Registration Certification. 14. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		CHSPL COMMONHE ALTH	<ol style="list-style-type: none"> 1. Drill Guide 2. Angled Bullet, 2.4mm ID 3. ACL foot print aimer 4. PCL Tibial Aimer 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Latest Authorization from principal manufacturer to

		<p>SYSTEMS PVT LTD</p> <p>Importer M/s Avana Medical Devices Pvt Ltd</p> <p>Manufacturer: M/s ARTHREX, NC-?</p>	<ol style="list-style-type: none"> 5. ACL/PCL Femoral Aimer 6. Angled Bullet, 3.5mm ID 7. Drill Bit Passing Pin, Dia, 2.4mm: Len, 250mm 8. Drill Tip Graft passing Pin, (Beadth Pin) Dia, 2.4mm: Len, 355mm 9. Tibial Reamer, Dia, 6.0mm/7.0mm/8.0mm/9.0mm Cannulated 10. Tibial Reamer, Dia, 6.5mm/7.5mm/8.5mm/9.5mm Cannulated 11. Femoral Reamer, Dia, 6.0mm/7.0mm/8.0mm/9.0mm Cannulated 12. Femoral Reamer, Dia, 6.5mm/7.5mm/8.5mm/9.5mm Cannulated 13. Bone Tunnel Pug 9-10mm 14. Tunnel Dilator 6mm /7mm/8mm/9mm/10mm 15. T-Handle for Tunnel Dialators 16. Endoscopic Reamer, Dia, 4.7mm, Cannulated 17. Screwdriver for Bio Interference Screws Cannulated 18. Screwdriver for PEEK Interference Screws Cannulated 19. Guide Wire for interference Screws Dia, 1.2mm, Len 300 (NitinoL) 20. Depth Gauge for ACL/PCL reconstruction 21. Transportal Femoral Aimer 7mm 22. Arthroscopic Probe 23. Mirofracture Awl 30° 24. Graft Prep Station includes Graft Board, Sliding Base, Button Holder with scale, Graft Grasper Clamp- Left & Right and Suture Holder with tensioner 25. Graft Sizing Block for Dia, 6-7-8-9-10-11-12mm 26. Graft Sizing Block for Dia, 6.5-7-8-8.5-10.5-11.5mm 27. Tendon Stripper, Open and 6mm 28. Notchplasty Curette 29. PCL Rasp 30. PCL Hook 31. Suture Retriever 32. Tissue Grasper 33. Guide Sleeve for al Reamer 34. Retrograde reamer for all inside ACL (7mm/8mm/9mm/10mm) 35. Graphic Case for ACL-PCL Reconstruction 	<p>Importer</p> <ol style="list-style-type: none"> 2. List of each quoted items as per the NIT(Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		<p>JK MEDITECH CO-OPERATIVE LTD</p> <p>Manufacturer M/S Chetan Meditech Pvt Ltd - Gujrat:</p>	<ol style="list-style-type: none"> 1. Drill Guide 2. Angled Bullet, 2.4mm ID 3. ACL foot print aimer 4. PCL Tibial Aimer 5. ACL/PCL Femoral Aimer 6. Angled Bullet, 3.5mm ID 7. Drill Bit Passing Pin, Dia, 2.4mm: Len, 250mm 8. Drill Tip Graft passing Pin, (Beadth Pin) Dia, 2.4mm: Len, 355mm 9. Tibial Reamer, Dia, 6.0mm/7.0mm/8.0mm/9.0mm Cannulated 10. Tibial Reamer, Dia, 6.5mm/7.5mm/8.5mm/9.5mm Cannulated 11. Femoral Reamer, Dia, 6.0mm/7.0mm/8.0mm/9.0mm Cannulated 12. Femoral Reamer, Dia, 6.5mm/7.5mm/8.5mm/9.5mm Cannulated 13. Bone Tunnel Pug 9-10mm 14. Tunnel Dilator 6mm /7mm/8mm/9mm/10mm 15. T-Handle for Tunnel Dialators 16. Endoscopic Reamer, Dia, 4.7mm, Cannulated 17. Screwdriver for Bio Interference Screws Cannulated 18. Screwdriver for PEEK Interference 	<ol style="list-style-type: none"> 1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.

			<div>19. Screws Cannulated</div> <div>20. Guide Wire for interference Screws Dia, 1.2mm, Len 300 (NitinoL)</div> <div>21. Depth Gauge for ACL/PCL reconstruction</div> <div>22. Transportal Femoral Aimer 7mm</div> <div>23. Arthroscopic Probe</div> <div>24. Mirofracture Awl 30°</div> <div>25. Graft Prep Station includes Graft Board, Sliding Base, Button Holder with scale, Graft Grasper Clamp- Left & Right and Suture Holder with tensioner</div> <div>26. Graft Sizing Block for Dia, 6-7-8-9- 10-11-12mm</div> <div>27. Graft Sizing Block for Dia, 6.5-7-5- 8.5-10.5-11.5mm</div> <div>28. Tendon stripper, Closed end 7mmk</div> <div>29. Tendon Stripper, Open and 6mm</div> <div>30. Notchplasty Curette</div> <div>31. PCL Rasp</div> <div>32. PCL Hook</div> <div>33. Suture Retriever</div> <div>34. Tissue Grasper</div> <div>35. Guide Sleeve for al Reamer</div> <div>36. Retrograde reamer for all inside ACL (7mm/8mm/9mm/10mm)</div> <div>37. Graphic Case for ACL-PCL Reconstruction</div>	
	<div>C.P. TRADERS & GENERAL ORDER SUPPLIERS</div> <div>Manufacturer M/S Relief ortho -Gujrat</div>	<div>1. Drill Guide</div> <div>2. Angled Bullet, 2.4mm ID</div> <div>3. ACL foot print aimer</div> <div>4. PCL Tibial Aimer</div> <div>5. ACL/PCL Femoral Aimer</div> <div>6. Angled Bullet, 3.5mm ID</div> <div>7. Drill Bit Passing Pin, Dia, 2.4mm: Len, 250mm</div> <div>8. Drill Tip Graft passing Pin, (Beadth Pin) Dia, 2.4mm: Len, 355mm</div> <div>9. Tibial Reamer, Dia,6.0mm/7.0mm/8.0mm/9.0m m Cannulated</div> <div>10. Tibial Reamer, Dia,6.5mm/7.5mm/8.5mm/9.5m m Cannulated</div> <div>11. Femoral Reamer, Dia,6.0mm/7.0mm/8.0mm/9.0m m Cannulated</div> <div>12. Femoral Reamer, Dia,6.5mm/7.5mm/8.5mm/9.5m m Cannulated</div> <div>13. Bone Tunnel Pug 9-10mm</div> <div>14. Tunnel Dilator 6mm /7.mm/8mm/9mm/10mm</div> <div>15. T-Handle for Tunnel Dialators</div> <div>16. Endoscopic Reamer, Dia, 4.7mm, Cannulated</div> <div>17. Screwdriver for Bio Interference Screws Cannulated</div> <div>18. Screwdriver for PEEK Interference Screws Cannulated</div> <div>19. Guide Wire for interference Screws Dia, 1.2mm, Len 300 (NitinoL)</div> <div>20. Depth Gauge for ACL/PCL reconstruction</div> <div>21. Transportal Femoral Aimer 7mm</div> <div>22. Arthroscopic Probe</div> <div>23. Mirofracture Awl 30°</div> <div>24. Graft Prep Station includes Graft Board, Sliding Base, Button Holder with scale, Graft Grasper Clamp- Left & Right and Suture Holder with tensioner</div> <div>25. Graft Sizing Block for Dia, 6-7-8-9- 10-11-12mm</div> <div>26. Graft Sizing Block for Dia, 6.5-7-5- 8.5-10.5-11.5mm</div> <div>27. Tendon stripper, Closed end 7mmk</div> <div>28. Tendon Stripper, Open and 6mm</div> <div>29. Notchplasty Curette</div> <div>30. PCL Rasp</div> <div>31. PCL Hook</div> <div>32. Suture Retriever</div> <div>33. Tissue Grasper</div> <div>34. Guide Sleeve for al Reamer</div> <div>35. Retrograde reamer for all inside ACL (7mm/8mm/9mm/10mm)</div> <div>36. Graphic Case for ACL-PCL</div>	<div>1. Copy of Catalogue of the each Quoted product (self attested)</div> <div>2. Compliance Sheet for each equipment (self attested)</div> <div>3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).</div> <div>4. Quality certification of the quoted equipment</div> <div>5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any..</div> <div>6. Specify point of supply with full Address.</div> <div>7. Declaration of bidder regarding acceptance Bid for terms & conditions</div> <div>8. Copy of CDSO Registration Certification.</div>	

			Reconstruction	
25	OPEN LATARJET / CORACOID TRANSFER PROCESS INSTRUMENT SET	SHARMA ENTERPRISES Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	<ol style="list-style-type: none"> Shoulder Retractor with 6 Blades Subscapularis Retractor-Blunt Tip Gelpi Retractor-Blunt Tip Self Retaining Retractor Glenoid Retractor 15mm wide blade with double prong Glenoid Retractor 18mm wide blade with double prong Glenoid Retractor 22mm wide blade with double prong Fukuda Humeral Head Retractor Coracoid Holding Forceps Glenoid Offset Parallel Drill Guide Drill Bit, Dia: 2.7mm, Cannulated for glenoid Cannulated Screw Driver SW 2.5 for LCS screw Guidewire Sleeve, Dia. 2.7mm for Coracoid Threaded Guidewire Dia. 1.2; Len. 180mm, SS Curved Osteotome for coracoid Drill Bit, Dia. 2.7mm, Len. 100.0mm for coracoid Depth measuring probe, Straight Graphic Case for BioLatarjet® Open Shoulder Instruments 	<p>To submit -</p> <ol style="list-style-type: none"> Copy of Catalogue of the Quoted product Compliance Sheet for each equipment. Technical bid submission sheet duly filled Model of each quoted instruments. Copy of Catalogue of the Quoted product Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product). Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. Latest GST Returns of the Bidder Copy of CDSCO Registration Certification. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		JK MEDITECH CO- OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujarat:	<ol style="list-style-type: none"> Shoulder Retractor with 6 Blades Subscapularis Retractor-Blunt Tip Gelpi Retractor-Blunt Tip Self Retaining Retractor Glenoid Retractor 15mm wide blade with double prong Glenoid Retractor 18mm wide blade with double prong Glenoid Retractor 22mm wide blade with double prong Fukuda Humeral Head Retractor Coracoid Holding Forceps Glenoid Offset Parallel Drill Guide Drill Bit, Dia: 2.7mm, Cannulated for glenoid Cannulated Screw Driver SW 2.5 for LCS screw Guidewire Sleeve, Dia. 2.7mm for Coracoid Threaded Guidewire Dia. 1.2; Len. 180mm, SS Curved Osteotome for coracoid Drill Bit, Dia. 2.7mm, Len. 100.0mm for coracoid Depth measuring probe, Straight Graphic Case for BioLatarjet® Open Shoulder Instruments 	<ol style="list-style-type: none"> Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Quality certification of the quoted equipment Copy of CDSCO Registration Certification.
		C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat	<ol style="list-style-type: none"> Shoulder Retractor with 6 Blades Subscapularis Retractor-Blunt Tip Gelpi Retractor-Blunt Tip Self Retaining Retractor Glenoid Retractor 15mm wide blade with double prong Glenoid Retractor 18mm wide blade with double prong Glenoid Retractor 22mm wide blade with double prong Fukuda Humeral Head Retractor Coracoid Holding Forceps Glenoid Offset Parallel Drill Guide Drill Bit, Dia: 2.7mm, Cannulated for glenoid Cannulated Screw Driver SW 2.5 for LCS screw Guidewire Sleeve, Dia. 2.7mm for Coracoid Threaded Guidewire Dia. 1.2; Len. 180mm, SS Curved Osteotome for coracoid Drill Bit, Dia. 2.7mm, Len. 100.0mm for coracoid Depth measuring probe, Straight 	<ol style="list-style-type: none"> Copy of Catalogue of the each Quoted product (self attested) Compliance Sheet for each equipment (self attested) Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Quality certification of the quoted equipment Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. Specify point of supply with full Address. Declaration of bidder regarding acceptance Bid for terms & conditions Copy of CDSCO Registration Certification.

28

			18. Graphic Case for BioLatarjet® Open Shoulder Instruments	
26	SHOULDER TRACTION KIT	SHARMA ENTERPRISES Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	1. Assembly for shoulder positioning (Lateral Decubitus Position) with OT Table Clamp	To submit: - 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Authorization from principal manufacturer in favor of Importer 7. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 8. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 9. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product) 10. Model of each quoted instruments. 11. Latest GST Returns of the Bidder 12. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 13. Declaration of bidder regarding acceptance Bid for terms & conditions 14. Copy of CDSCO Registration Certification. 15. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		CHSPL COMMONHEALTH SYSTEMS PVT LTD Importer M/s Avana Medical Devices Pvt Ltd Manufacturer: M/s ARTHREX, NC-?	1. Assembly for shoulder positioning (Lateral Decubitus Position) with OT Table Clamp	To submit: - 1. Latest Authorization from principal manufacturer to Importer 2. List of each quoted items as per the NIT(Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		JK MEDITECH CO-OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujarat	1. Assembly for shoulder positioning (Lateral Decubitus Position) with OT Table Clamp	1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.
		C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat	1. Assembly for shoulder positioning (Lateral Decubitus Position) with OT Table Clamp	1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions

27

27	Meniscus Root Repair Zig	SHARMA ENTERPRISES Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	<ol style="list-style-type: none"> 1. Meniscus Root Repair Zig consists of: 2. Drill Guide 3. Angled Bullet, 2.4mm ID 4. Aimer for Meniscus Root Repair 5. SHUTTLE FERRY® Suture Passer Mini for Knee 	<p>8. Copy of CDSCO Registration Certification.</p> <p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product 2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Authorization from principal manufacturer in favor of Importer 7. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 8. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 9. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product) 10. Model of each quoted instruments. 11. Latest GST Returns of the Bidder 12. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 13. Declaration of bidder regarding acceptance Bid for terms & conditions 14. Copy of CDSCO Registration Certification. 15. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		CHSPL COMMONHEALTH SYSTEMS PVT LTD Importer M/s Avana Medical Devices Pvt Ltd Manufacturer: M/s ARTHREX, NC-?	<ol style="list-style-type: none"> 1. Meniscus Root Repair Zig consists of: 2. Drill Guide 3. Angled Bullet, 2.4mm ID 4. Aimer for Meniscus Root Repair 5. SHUTTLE FERRY® Suture Passer Mini for Knee 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Latest Authorization from principal manufacturer to Importer 2. List of each quoted items as per the NIT(Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		JK MEDITECH CO-OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujarat:	<ol style="list-style-type: none"> 1. Drill Guide 2. Angled Bullet, 2.4mm ID 3. Aimer for Meniscus Root Repair 4. SHUTTLE FERRY® Suture Passer Mini for Knee 	<ol style="list-style-type: none"> 1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.
		C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat	<ol style="list-style-type: none"> 1. Drill Guide 2. Angled Bullet, 2.4mm ID 3. Aimer for Meniscus Root Repair 4. SHUTTLE FERRY® Suture Passer Mini for Knee 	<ol style="list-style-type: none"> 1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions 8. Copy of CDSCO Registration Certification.
28	MENISCUS REPAIR	SHARMA ENTERPRISES	<ol style="list-style-type: none"> 1. Curved Probe for Meniscus Repair 	<p>To submit: -</p> <ol style="list-style-type: none"> 1. Copy of Catalogue of the Quoted product

9

28

4	INSTRUMENT SET	Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	2. Meniscus Rasp 3. Meniscus Curette 4. Guide Sleeve (Double Lumen) for Meniscus Repair, Straight 5. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Left 6. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Right 7. Guide Sleeve (Single Lumen) for Meniscus Repair, Straight 8. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 10° 9. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 20° 10. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 30° 11. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 40° 12. Needle Pusher 13. Sleeve Bender 14. Knot Pusher for Meniscus Repair 15. Slotted Sleeve For all inside Meniscus repair 16. Graphic Case for Meniscus Repair Instruments	2. Compliance Sheet for each equipment. 3. Technical bid submission sheet duly filled 4. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Authorization from principal manufacturer in favor of Importer 7. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019- 20 and 2020-21). 8. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 9. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product) 10. Model of each quoted instruments. 11. Latest GST Returns of the Bidder 12. Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. 13. Declaration of bidder regarding acceptance Bid for terms & conditions 14. Copy of CDSCO Registration Certification. 15. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX,-Inc USA(importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		CHSPL COMMONHE ALTH SYSTEMS PVT LTD Importer M/s Avana Medical Devices Pvt Ltd Manufacturer: M/s ARTHREX, NC-?	1. Curved Probe for Meniscus Repair 2. Meniscus Rasp 3. Meniscus Curette 4. Knot Pusher for Meniscus Repair 5. Slotted Sleeve For all inside Meniscus repair 6. Graphic Case for Meniscus Repair Instruments	To submit: - 1. Latest Authorization from principal manufacturer to Importer 2. List of each quoted items as per the NIT(Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		JK MEDITECH CO- OPERATIVE LTD Manufacturer: M/S Chetan Meditech Pvt Ltd - Gujrat:	1. Curved Probe for Meniscus Repair 2. Meniscus Rasp 3. Meniscus Curette 4. Guide Sleeve (Double Lumen) for Meniscus Repair, Straight 5. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Left 6. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Right 7. Guide Sleeve (Single Lumen) for Meniscus Repair, Straight 8. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 10° 9. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 20° 10. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 30° 11. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 40° 12. Needle Pusher	1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.

			13. Sleeve Bender 14. Knot Pusher for Meniscus Repair 15. Slotted Sleeve For all inside Meniscus repair 16. Graphic Case for Meniscus Repair Instruments	
		C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat	1. Curved Probe for Meniscus Repair 2. Meniscus Rasp 3. Meniscus Curette 4. Guide Sleeve (Double Lumen) for Meniscus Repair, Straight 5. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Left 6. Guide Sleeve (Double Lumen) for Meniscus Repair, Curved, Right 7. Guide Sleeve (Single Lumen) for Meniscus Repair, Straight 8. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 10° 9. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 20° 10. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 30° 11. Guide Sleeve (Single Lumen) for Meniscus Repair, Curved, 40° 12. Needle Pusher 13. Sleeve Bender 14. Knot Pusher for Meniscus Repair 15. Slotted Sleeve For all inside Meniscus repair 16. Graphic Case for Meniscus Repair Instruments	1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions 8. Copy of CDSCO Registration Certification.
29	Foot AND Ankle Instrument Set	J K MEDITECH CO-OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujarat	1. Fluted Reamer Cannulated, Dia. 4.0mm 2. Fluted Reamer Cannulated, Dia. 5.0mm 3. Fluted Reamer Cannulated, Dia. 6.0mm 4. Fluted Reamer Cannulated, Dia. 7.0mm 5. Fluted Reamer Cannulated, Dia. 8.0mm 6. Screwdriver for Peek Interference Screw, Cannulated, Dia. 4.0mm 7. Screwdriver for Peek Interference Screws, Cannulated, Dia: 5.0mm 8. Screwdriver for Peek Interference Screw Len. 23.0mm, Cannulated 9. Guide Wire, Blunt Tip for Dia. 4.0mm Peek Interference Screws Dia. 0.8mm; Len. 200mm (SS) 10. Guide Wire, Blunt Tip for Dia. 5.0mm-8.0mm Peek Interference Screws Dia. 1.2mm; Len. 200mm (NITINOL) 11. Threaded Guidewire Dia. 1.2; Len. 180mm, SS 12. Drill Bit Dia 1.2mm , Len.- 80mm for Soft Ligament Anchor Dia 1.2mm with Needle 13. Drill Bit Dia 1.6mm , Len.- 100 mm for Soft Ligament Anchor Dia 1.5mm and 1.8mm with Needles 14. Drill bit, Dia. 1.3mm for 1.8mm Titanium Ligament Anchor 15. Drill bit, Dia. 1.5mm for 2.0mm Titanium Ligament Anchor 16. Drill bit, Dia. 1.5mm for 1.8mm Peek Ligament Anchor 17. Drill bit, Dia. 1.7mm for 2.0mm Peek Ligament Anchor 18. Drill bit Dia. 2.6mm, Len. 100mm for Peek Double loaded 3.0mm Ligament Anchor 19. Drill bit, Cannulated Dia. 2.8mm, Len. 100mm for Peek 3.5mm Ligament Anchor, knotless 20. Drill Bit Dia 3.2mm, Len. 150 mm for Syndesmosis Fixation Button with Loop 21. Drill bit, Dia. 3.5mm, Len. 100mm for 4.75mm Lateral Row Peek Ligament Anchor, knotless 22. Suture Passing Pin, Spade- Tip (Beath Pin) Dia. 3.2mm; Len. 250mm 23. Punch for Soft Ligament Anchor,	1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.

			<div>24. Dia 2.9mm Tap for Peek, Dia 4.5mm & 5.5mm Peek Ligament Anchor and Dia 4.75mm & 5.5mm Peek Lateral Row Anchor</div> <div>25. Slotted Sleeve for Dia. 1.8mm & 2.0mm Titanium & Peek Ligament Anchor</div> <div>26. Slotted Sleeve for Foot & Ankle Drill bits-10004 ,10005 , 10067</div> <div>27. Graft Sizer for Foot & Ankle, Dia. 4-5-6-7-8 mm</div> <div>28. Suture passer, Banana Curved</div> <div>29. Suture Passing wire, Blunt Tip with Loop Len. 200mm</div> <div>30. Graphic Case for Foot & Ankle Instruments</div>	
	<div>C.P. TRADERS & GENERAL ORDER SUPPLIERS</div> <div>Manufacturer M/S Relief ortho -Gujrat</div>	<div>1. Fluted Reamer Cannulated, Dia. 4.0mm</div> <div>2. Fluted Reamer Cannulated, Dia. 5.0mm</div> <div>3. Fluted Reamer Cannulated, Dia. 6.0mm</div> <div>4. Fluted Reamer Cannulated, Dia. 7.0mm</div> <div>5. Fluted Reamer Cannulated, Dia. 8.0mm</div> <div>6. Screwdriver for Peek Interference Screw, Cannulated, Dia. 4.0mm</div> <div>7. Screwdriver for Peek Interference Screws, Cannulated, Dia: 5.0mm</div> <div>8. Screwdriver for Peek Interference Screw Len. 23.0mm, Cannulated</div> <div>9. Guide Wire, Blunt Tip for Dia. 4.0mm Peek Interference Screws Dia. 0.8mm; Len. 200mm (SS)</div> <div>10. Guide Wire, Blunt Tip for Dia. 5.0mm-8.0mm Peek Interference Screws Dia. 1.2mm; Len. 200mm (NITINOL)</div> <div>11. Threaded Guidewire Dia. 1.2; Len. 180mm, SS</div> <div>12. Drill Bit Dia 1.2mm , Len.- 80mm for Soft Ligament Anchor Dia 1.2mm with Needle</div> <div>13. Drill Bit Dia 1.6mm , Len.- 100 mm for Soft Ligament Anchor Dia 1.5mm and 1.8mm with Needles</div> <div>14. Drill bit, Dia. 1.3mm for 1.8mm Titanium Ligament Anchor</div> <div>15. Drill bit, Dia. 1.5mm for 2.0mm Titanium Ligament Anchor</div> <div>16. Drill bit, Dia. 1.5mm for 1.8mm Peek Ligament Anchor</div> <div>17. Drill bit, Dia. 1.7mm for 2.0mm Peek Ligament Anchor</div> <div>18. Drill bit Dia. 2.6mm, Len. 100mm for Peek Double loaded 3.0mm Ligament Anchor</div> <div>19. Drill bit, Cannulated Dia. 2.8mm, Len. 100mm for Peek 3.5mm Ligament Anchor, knotless</div> <div>20. Drill Bit Dia 3.2mm, Len. 150 mm for Syndesmosis Fixation Button with Loop</div> <div>21. Drill bit, Dia. 3.5mm, Len. 100mm for 4.75mm Lateral Row Peek Ligament Anchor, knotless</div> <div>22. Suture Passing Pin, Spade- Tip (Beath Pin) Dia. 3.2mm; Len. 250mm</div> <div>23. Punch for Soft Ligament Anchor, Dia 2.9mm</div> <div>24. Tap for Peek, Dia 4.5mm & 5.5mm Peek Ligament Anchor and Dia 4.75mm & 5.5mm Peek Lateral Row Anchor</div> <div>25. Slotted Sleeve for Dia. 1.8mm & 2.0mm Titanium & Peek Ligament Anchor</div> <div>26. Slotted Sleeve for Foot & Ankle Drill bits-10004 ,10005 , 10067</div> <div>27. Graft Sizer for Foot & Ankle, Dia. 4-5-6-7-8 mm</div> <div>28. Suture passer, Banana Curved</div> <div>29. Suture Passing wire, Blunt Tip with Loop Len. 200mm</div> <div>30. Graphic Case for Foot & Ankle Instruments</div>	<div>1. Copy of Catalogue of the each Quoted product (self attested)</div> <div>2. Compliance Sheet for each equipment (self attested)</div> <div>3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).</div> <div>4. Quality certification of the quoted equipment</div> <div>5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any..</div> <div>6. Specify point of supply with full Address.</div> <div>7. Declaration of bidder regarding acceptance Bid for terms & conditions</div> <div>8. Copy of CDSCO Registration Certification.</div>	

37

30	Specifications of Small joint scope AND shaver and instruments	SHARMA ENTERPRISES Importer : Cure Surgical- new Delhi Manufacturer: M/s ARTHREX, NC-USA	<ol style="list-style-type: none"> 2.7 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 2.7mm Wide Angle, Direct View High Definition autoclavable Arthroscope Light Guide insertion on opposite side of the direction of view with a J-lock fixation for cannula. Working Length of Not more than 120mm Angle of view: 30 degree Diameter 2.7 mm Fiber optic light transmission incorporated Standard ocular window for coupling the camera head Scratch resistance sapphire quoted tip lens Advanced Rod lens system for optimum brightness, contrast and definition Arthroscopes should be supplied with compatible cannulas high flow, double valve, fully rotatable with fenestrated tip & conical and blunt tip obturator. Sheath- 3mm to 4mm, high flow diagnostic cannula, double valve, fully rotatable cannula with fenestrated tip. 1 qty Trocar-3mm to 4mm conical obturator to fit with cannula. 1qty 1.9 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 1.9mm, 30 degree, length 65-70 direct view HD small joint arthroscope 2.2 mm Short Cannula, with flow port, obturator 2.2mm Compatible light guide cable 2mm TMJ/Wrist Probe Small joint Shaver Small Joint Shaver hand piece (for 2mm to 3mm blades and burrs) with hand control, suction lever and blade multi-positioning, 10 ft. (3m) cord. Shaver should be supplied with 6 small joint blades Small Joint Handheld instruments Micro Pitbull Grasper, straight, pinless hinge design Raptor Punch, straight, pinless hinge design BLUNT NOSE JR. PUNCH pinless hinge design Alligator grasper 	<p>To submit:-</p> <ol style="list-style-type: none"> Copy of Catalogue of the Quoted product Compliance Sheet for each equipment. Technical bid submission sheet duly filled Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Authorization from principal manufacturer in favor of Importer Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. Self attested photocopy of IEC certificate and Permission/ Authorization for sale for sale from the foreign principle manufacturer (in case of imported product) Model of each quoted instruments. Latest GST Returns of the Bidder Quality Certifications on the products viz. ISI/CE/USFDA etc. whichever applicable. Declaration of bidder regarding acceptance Bid for terms & conditions Copy of CDSCO Registration Certification. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		CHSPL COMMONHEALTH SYSTEMS PVT LTD	<ol style="list-style-type: none"> 2.7 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 2.7mm Wide Angle, Direct View High Definition autoclavable Arthroscope Light Guide insertion on opposite side of the direction of view with a J-lock fixation for cannula. Working Length of Not more than 120mm Angle of view: 30 degree Diameter 2.7 mm Fiber optic light transmission incorporated Standard ocular window for coupling the camera head Scratch resistance sapphire quoted tip lens Advanced Rod lens system for optimum brightness, contrast and definition Arthroscopes should be supplied with compatible cannulas high flow, double valve, fully rotatable with fenestrated tip & conical and blunt tip obturator. Sheath- 3mm to 4mm, high flow diagnostic cannula, double valve 	<p>To submit:-</p> <ol style="list-style-type: none"> Latest Authorization from principal manufacturer to Importer List of each quoted items as per the NIT (Submitted is Arthroscopy instrumentation only) Copy of Catalogue of the Quoted product Compliance Sheet for each equipment. Model of each quoted instruments. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Copy of CDSCO Registration Certification. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.

			13. fully rotatable cannula with fenestrated tip. 1 qty 14. Trocar-3mm to 4mm conical obturator to fit with cannula. 1qty 15. 1.9 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 16. 1.9mm, 30 degree, length 65-70 direct view HD small joint arthroscope 17. 2.2 mm Short Cannula, with flow port, obturator 2.2mm 18. Compatible light guide cable 2mm 19. TMJ/Wrist Probe 20. Small joint Shaver 21. Small Joint Shaver hand piece (for 2mm to 3mm blades and burrs) with hand control, suction lever and blade multi-positioning, 10 ft. (3m) cord. Shaver should be supplied with 6 small joint blades	
		JAI DURGA MEDICAL AND SURGICAL AGENCIES Importer: Smith and Nephew Health Care Pvt Ltd Manufacturer: Smith and Nephew-Inc	1. 2.7 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 2. 2.7mm Wide Angle, Direct View High Definition autoclavable Arthroscope 3. Sheath- 3mm to 4mm, high flow diagnostic cannula, double valve, fully rotatable cannula with fenestrated tip. 1 qty 4. Trocar-3mm to 4mm conical obturator to fit with cannula. 1qty 5. 1.9 mm Small joint Arthroscope complete Set (Arthroscope, Sheath and Obturator) 6. 1.9mm, 30 degree, length 65-70 direct view HD small joint arthroscope 7. 2.2 mm Short Cannula, with flow port, obturator 2.2mm 8. Compatible light guide cable 2mm 9. TMJ/Wrist Probe 10. Small Joint Shaver hand piece (for 2mm to 3mm blades and burrs) with hand control, suction lever and blade multi-positioning, 10 ft. (3m) cord. Shaver should be supplied with 6 small joint blades 11. Micro Pitbull Grasper, straight, pinless hinge design 12. Raptor Punch, straight, pinless hinge design 13. BLUNT NOSE JR. PUNCH pinless hinge design 14. Alligator grasper	1. Authorization from principal manufacturer (Smith and Nephew-Inc) in favour of importer (Smith and Nephew Health Care Pvt Ltd-mumbai). 2. Client base with reference of er for any of the three years in last 5 years along with the supply order on letter head of the bidder / import satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 3. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Sole Importer as per proforma duly notarized. 4. Readable copy Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 5. Copy of CDSO Registration Certification.
31	HIGH TIBIAL OSTEOTOMY (HTO) INSTRUMENTS ET	SHARMA ENTERPRISES Manufacturer: Nebula Surgical Pvt Ltd-Gujrat - 360005	1. Bone Spreader 2. Osteotomy Chisel, 10mm 3. Osteotomy Chisel, 15mm 4. Osteotomy Chisel, 20mm 5. Osteotomy Chisel, 25mm 6. Drillbit Dia. 3.2mm; Len. 150mm for 4.5 cortiscrews 7. Drill Bit, Dia. 4.3mm, Length 220mm for 5mm Locking Head Screw 8. Depth Gauge, Large 9. Alignment Rod, Male/Female 10. Threaded Drill Sleeve Dia. 4.3mm 11. Drill Sleeve, Threaded, Dia. 3.2mm 12. K-Wire Sleeve, dia. 1.8mm 13. Osteotomy Jack with Allen Key SW 3.5 14. Gap Measuring Device 15. Screwdriver, Star Driver For 4.5/5.0 mm Screw 16. Screwdriver Shaft, Star Driver For 4.5/5.0 mm Screw 17. Torque Screwdriver, Star Driver For 4.5/5.0 mm Screw 18. Tibial Retractor, Radioluscent 19. K-Wire, Dia. 1.8mm, Len. 230mm (Pack of 4 in 1) 20. Threaded Guidewire Dia. 1.2mm; Len. 180mm (Pack of 2 in 1) 21. Graphic Case for HTO Instrumentation Set	To submit:- 1. Compliance Sheet for each equipment. 2. Technical bid submission sheet duly filled 3. Declaration for Latest Non-Conviction, non blacklisting on non-judicial Stamp paper of Rs 100 furnished by the Principal manufacturer as per proforma duly notarized. 4. Model of each quoted instruments. 5. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 6. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). 7. Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. 8. Latest GST Returns of the Bidder 9. Declaration of bidder regarding acceptance Bid for terms & conditions 10. The irrelevant documents of Johnson & Johnson Pvt and ARTHREX, Inc USA (importer Cure Surgical) have been uploaded by you. clarify the relation of the said firm along with complete documentary evidence.
		DKS ORTHO AND	1. Bone Spreader 2. Osteotomy Chisel, 10mm	To submit:- 1. List of quoted items as per NIT

		SHARMA PHARMACE UTICAL PRIVATE LIMITED (Manufacturer / Bidder)	<ol style="list-style-type: none"> OsteotomyChisel,15mm OsteotomyChisel,20mm OsteotomyChisel,25mm DrillbitDia.3.2mm;Len.150mmfor4.5cortexscrews DrillBit,Dia.4.3mm,Length220mmf or5mmLockingHeadScrew Depth Gauge, Large Alignment Rod, Male/Female ThreadedDrillSleeveDia.4.3mm DrillSleeve,Threaded,Dia.3.2mm K-WireSleeve,dia.1.8mm OsteotomyJackIIwithAllenKeySW 3.5 Gap Measuring Device Screwdriver,StarDriverFor4.5/5.0 mmScrew ScrewdriverShaft,StarDriverFor4.5 /5.0mmScrew TorqueScrewdriver,StarDriverFor4 .5/5.0mmScrew Tibial Retractor, Radioluscent K- Wire,Dia.1.8mm,Len.230mm(Pack of4in1) ThreadedGuidewireDia.1.2mm;Le n.180mm(Packof2in1) Graphic Case for HTO Instrumentation Set 	<ol style="list-style-type: none"> Copy of catalogue of each quoted item(self attested) The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi) . Please Clarify along with relevant documents Compliance Sheet for each equipment Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Latest GST Returns of the Bidder Quality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filled Copy of CDSCO Registration Certification.
		ORTHOFIX (Bidder) Importer:- Johnson & Johnson Pvt. Ltd. LBS Marg Mulund (W) Mumbai 400080 Manufacture:- Synthesis USA Products, LLC 1302 rights land east west chester, PA 19380	<ol style="list-style-type: none"> Bone Spreader OsteotomyChisel,10mm OsteotomyChisel,15mm OsteotomyChisel,20mm OsteotomyChisel,25mm DrillbitDia.3.2mm;Len.150mmfor4.5cortexscrews DrillBit,Dia.4.3mm,Length220mmf or5mmLockingHeadScrew Depth Gauge, Large Alignment Rod, Male/Female ThreadedDrillSleeveDia.4.3mm DrillSleeve,Threaded,Dia.3.2mm K-WireSleeve,dia.1.8mm OsteotomyJackIIwithAllenKeySW 3.5 Gap Measuring Device Screwdriver,StarDriverFor4.5/5.0 mmScrew ScrewdriverShaft,StarDriverFor4.5 /5.0mmScrew TorqueScrewdriver,StarDriverFor4 .5/5.0mmScrew Tibial Retractor, Radioluscent K- Wire,Dia.1.8mm,Len.230mm(Pack of4in1) ThreadedGuidewireDia.1.2mm;Le n.180mm(Packof2in1) Graphic Case for HTO Instrumentation Set 	To submit:- <ol style="list-style-type: none"> Copy of catalogue of each quoted item (self-attested) . Technical bid submission sheet duly filled (Annexures II). Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Authorization from principal manufacturer (Synthesis GmbH) in favour of importer (Johnson & Johnson Pvt. Ltd. Average Annual Turnover Statement for Last 3 financial Years of the Importer issued by Chartered Accountant/competent authority with UDIN (2018-19, 2019-20 and 2020-21). Copies of Audited Balance sheet & profit loss account for last three financial years certified by Chartered Accountant of the Importer with UDIN. Self-attested photocopy of IEC (Import Export Certificate). Quality Certification of the quoted equipment as per NIT Copy of CDSCO Registration Certification.
		CHSPL COMMONHE ALTH SYSTEMS PVT LTD Importer M/s Avana Medical Devices Pvt Ltd Manufacturer: M/s ARTHREX, NC-7	<ol style="list-style-type: none"> Bone Spreader OsteotomyChisel,15mm OsteotomyChisel,20mm OsteotomyChisel,25mm DrillbitDia.3.2mm;Len.150mmfor4.5cortexscrews DrillBit,Dia.4.3mm,Length220mmf or5mmLockingHeadScrew Depth Gauge, Large Alignment Rod, Male/Female ThreadedDrillSleeveDia.4.3mm DrillSleeve,Threaded,Dia.3.2mm Screwdriver,StarDriverFor4.5/5.0 mmScrew TorqueScrewdriver,StarDriverFor4 .5/5.0mmScrew Tibial Retractor, Radioluscent Graphic Case for HTO Instrumentation Set 	To submit:- <ol style="list-style-type: none"> Latest Authorization from principal manufacturer to Importer List of each quoted items as per the NIT(Submitted is Arthroscopy instrumentation only) Copy of Catalogue of the Quoted product Compliance Sheet for each equipment. Model of each quoted instruments. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Copy of CDSCO Registration Certification. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		JK MEDITECH CO- OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujrat:	<ol style="list-style-type: none"> Bone Spreader OsteotomyChisel,10mm OsteotomyChisel,15mm OsteotomyChisel,20mm OsteotomyChisel,25mm DrillbitDia.3.2mm;Len.150mmfor4.5cortexscrews DrillBit,Dia.4.3mm,Length220mmf or5mmLockingHeadScrew Depth Gauge, Large 	<ol style="list-style-type: none"> Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). Quality certification of the quoted equipment Copy of CDSCO Registration Certification.

34

			9. Alignment Rod, Male/Female 10. Threaded Drill Sleeve Dia. 4.3mm 11. Drill Sleeve, Threaded, Dia. 3.2mm 12. K-Wire Sleeve, dia. 1.8mm 13. Osteotomy Jack II with Allen Key SW 3.5 14. Gap Measuring Device 15. Screwdriver, Star Driver For 4.5/5.0 mm Screw 16. Screwdriver Shaft, Star Driver For 4.5/5.0 mm Screw 17. Torque Screwdriver, Star Driver For 4.5/5.0 mm Screw 18. Tibial Retractor, Radioluscent 19. K-Wire, Dia. 1.8mm, Len. 230mm (Pack of 4 in 1) 20. Threaded Guidewire Dia. 1.2mm; Len. 180mm (Pack of 2 in 1) 21. Graphic Case for HTO Instrumentation Set	
		C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat	1. Bone Spreader 2. Osteotomy Chisel, 10mm 3. Osteotomy Chisel, 15mm 4. Osteotomy Chisel, 20mm 5. Osteotomy Chisel, 25mm 6. Drill bit Dia. 3.2mm; Len. 150mm for 4.5 cortex screws 7. Drill Bit, Dia. 4.3mm, Length 220mm for 5mm Locking Head Screw 8. Depth Gauge, Large 9. Alignment Rod, Male/Female 10. Threaded Drill Sleeve Dia. 4.3mm 11. Drill Sleeve, Threaded, Dia. 3.2mm 12. K-Wire Sleeve, dia. 1.8mm 13. Osteotomy Jack II with Allen Key SW 3.5 14. Gap Measuring Device 15. Screwdriver, Star Driver For 4.5/5.0 mm Screw 16. Screwdriver Shaft, Star Driver For 4.5/5.0 mm Screw 17. Torque Screwdriver, Star Driver For 4.5/5.0 mm Screw 18. Tibial Retractor, Radioluscent 19. K-Wire, Dia. 1.8mm, Len. 230mm (Pack of 4 in 1) 20. Threaded Guidewire Dia. 1.2mm; Len. 180mm (Pack of 2 in 1) 21. Graphic Case for HTO Instrumentation Set	1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions 8. Copy of CDSCO Registration Certification.
32	BBQR OF ALEX-III SHOULDER MODEL	CHSPL COMMONHEALTH SYSTEMS PVT LTD	1. Wide-angle modular portal covers and a quick-release base. In addition to capsular plication, SLAP (superior labrum, anterior to posterior) lesion repair, rotator cuff repair (side-to-side and suture anchor to bone), and knot tying techniques, this model now also allows for biceps tenodesis (arthroscopic), PASTA (partial articular supraspinatus avulsion), and subscapularis tendon repair : 1 Pc	To submit :- 1. Latest Authorization from principal manufacturer to Importer 2. List of each quoted items as per the NIT (Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		J K MEDITECH CO-OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd -Gujrat:	1. Wide-angle modular portal covers and a quick-release base. In addition to capsular plication, SLAP (superior labrum, anterior to posterior) lesion repair, rotator cuff repair (side-to-side and suture anchor to bone), and knot tying techniques, this model now also allows for biceps tenodesis (arthroscopic), PASTA (partial articular supraspinatus avulsion), and subscapularis tendon repair : 1 Pc	1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.
		C.P. TRADERS & GENERAL ORDER	1. Wide-angle modular portal covers and a quick-release base. In addition to capsular plication, SLAP (superior lab	1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head

35

		<p>SUPPLIERS</p> <p>Manufacturer M/S Relief ortho -Gujrat</p>	<p>rum, anterior to posterior) lesion repair, rotator cuff repair (side-to-side and suture anchor to bone), and knot tying techniques, this model now also allows for biceps tenodesis (arthroscopic), PASTA (partial articular supraspinatus avulsion), and subscapularis tendon repair : 1 Pc</p>	<p>of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).</p> <ol style="list-style-type: none"> 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions 8. Copy of CDSCO Registration Certification.
33	BBQR OF KNEE ALEX BONE MODEL	<p>CHSPL COMMONHE ALTH SYSTEMS PVT LTD Importer M/s Avana Medical Devices Pvt Ltd Manufacturer: M/s ARTHREX, NC-?</p>	<ol style="list-style-type: none"> 1. Arthroscopy Workstation with Knot Tester - 1 Set Complete Kit Includes: A - Clear Shell B - Base C - Opaque White Shell D - Knot Tester E - Resection-Suture Passage Module F - Simple Probing and Targeting Module G - 1 Inch Spacer H - Maze Module I - Horizontal Ring Transfer Module J - Smart Phone Module K - Vertical Ring Transfer Module L - Shoulder Module M - Knot Tying Module 2. ACL Knee Trainer, Medium. Includes soft tissue sleeve, clamps and replaceable - 2 Sets 3. Knot Pusher - 2 pc (Blunt end and does not cut suture) 4. Probe - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 5. Grasper - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. Latest Authorization from principal manufacturer to Importer 2. List of each quoted items as per the NIT (Submitted is Arthroscopy instrumentation only) 3. Copy of Catalogue of the Quoted product 4. Compliance Sheet for each equipment. 5. Model of each quoted instruments. 6. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 7. Copy of CDSCO Registration Certification. 8. Submitted letter of authorization is of M/ S Arthrex GmbH but quality certification submitted is M/ S Arthrex USA Clarify the issue with document proof.
		<p>JK MEDITECH CO- OPERATIVE LTD Manufacturer M/S Chetan Meditech Pvt Ltd - Gujarat:</p>	<ol style="list-style-type: none"> 1. Arthroscopy Workstation with Knot Tester - 1 Set Complete Kit Includes: A - Clear Shell B - Base C - Opaque White Shell D - Knot Tester E - Resection-Suture Passage Module F - Simple Probing and Targeting Module G - 1 Inch Spacer H - Maze Module I - Horizontal Ring Transfer Module J - Smart Phone Module K - Vertical Ring Transfer Module L - Shoulder Module M - Knot Tying Module 2. ACL Knee Trainer, Medium. Includes soft tissue sleeve, clamps and replaceable - 2 Sets 3. Knot Pusher - 2 pc (Blunt end and does not cut suture) 4. Probe - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 5. Grasper - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 	<ol style="list-style-type: none"> 1. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 2. Quality certification of the quoted equipment 3. Copy of CDSCO Registration Certification.
		<p>C.P. TRADERS & GENERAL ORDER SUPPLIERS Manufacturer M/S Relief ortho -Gujrat</p>	<ol style="list-style-type: none"> 1. Arthroscopy Workstation with Knot Tester - 1 Set Complete Kit Includes: A - Clear Shell B - Base C - Opaque White Shell D - Knot Tester E - Resection-Suture Passage Module F - Simple Probing and Targeting Module G - 1 Inch Spacer H - Maze Module I - Horizontal Ring Transfer Module J - Smart Phone Module K - Vertical Ring Transfer Module L - Shoulder Module M - Knot Tying Module 2. ACL Knee Trainer, Medium. Includes soft tissue sleeve, clamps and replaceable - 2 Sets 3. Knot Pusher - 2 pc (Blunt end and does not cut suture) 4. Probe - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 5. Grasper - Surgical training tool can be used in conjunction with Knot Pusher - 2 Pc 	<ol style="list-style-type: none"> 1. Copy of Catalogue of the each Quoted product (self attested) 2. Compliance Sheet for each equipment (self attested) 3. Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached). 4. Quality certification of the quoted equipment 5. Name, photograph & specimen signature of the designated officer/ representative of the Bidder who is authorized to make correspondence with the JKMSCL, if any.. 6. Specify point of supply with full Address. 7. Declaration of bidder regarding acceptance Bid for terms & conditions 8. Copy of CDSCO Registration Certification.
34	FIXATOR SYSTEM	<p>DKSORTHO AND SHARMA</p>	<ol style="list-style-type: none"> 1. Rail Hinge-Female 2. Rail End Cap-Female 3. Rail End Cap-Male 	<p>To submit:-</p> <ol style="list-style-type: none"> 1. List of quoted items as per NIT 2. Copy of catalogue of each quoted item (self attested)

		PHARMACE UTICAL PRIVATE LIMITED (Manufacturer / Bidder)	<ol style="list-style-type: none">Rail Hinge-MaleArc Small 90*Arc Medium 120*Arc Large 180*Rail 150mmRail 200mmRail 250mmRail 300mmRail 350mmRail Bolt M6 Cannulated 28mmRail Bolt M6 Cannulated 38mmDrill Guides 6x5x40mmDrill Guides 6x5x80mmDrill Guides 6x3.2x40mmDrill Guides 6x3.2x80mmRail Bolt M6 27mmRail Bolt M6 30mmAllen Key 6mmLocking Nut M6X6mmRail Bolt M6X7mm90*Screw Driver3mm Straight Screw DriverDrill bitcannulated 4.8X1.8X200mmDrillbitcannulated4.3.2X1.8X200mmDrill bit 3.2X200mmDrill bit 4.8X200mmGuide wire sleeve 5X1.8X40mmGuide wire sleeve 5X1.8X80mm	<ol style="list-style-type: none">The firm has participated as MSME (Micro) but has also submitted authorization in favour of M/s Lexical Health Care Pvt Ltd (New Delhi) . Please Clarify along with relevant documentsCompliance Sheet for each equipmentClient base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).Latest GST Returns of the BidderQuality Certification of the quoted equipment as per NIT Technical bid submission sheet duly filledCopy of CDSCO Registration Certification.
		JAI DURGA MEDICAL AND SURGICAL AGENCIES Manufacturer (M/S HPitkar Orth tools PVT Ltd	<ol style="list-style-type: none">Rotation End ConnectorRail Hinge-FemaleStandard 3 Hole ClampStacked 3 Hole ClampStandard Clamp-Hinge ConnectionRail End Cap-FemaleStacked 5 Hole ClampRail Arc ClampRail End Cap-MaleRail Arc Rotation endRail Hinge-MaleArc Small 90*Arc Medium 120*Arc Large 180*Stacked 5 Hole OverhangRail 150mmRail 200mmRail 250mmRail 300mmRail 350mmStandard 1 Hole ClampA-P Angle Attachment clampRail Bolt M6 Cannulated 28mmRail Bolt M6 Cannulated 38mmDrill Guides 6x5x40mmDrill Guides 6x5x80mmDrill Guides 6x3.2x40mmDrill Guides 6x3.2x80mmRail Bolt M6 27mmRail Bolt M6 30mmAllen Key 6mmRail Set Screw M6X7.5mmLocking Nut M6X6mmHalf Pin Introducer D TypeRail Bolt M6X7mm90*Screw DriverCompression And Distraction Device3mm Straight Screw DriverDrill bitcannulated 4.8X1.8X200mmDrillbitcannulated4.3.2X1.8X200mmDrill bit 3.2X200mmDrill bit 4.8X200mmGuide wire sleeve 5X1.8X40mmGuide wire sleeve 5X1.8X80mmContainer for fixture instrument Set	<ol style="list-style-type: none">Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).Copy of CDSCO Registration Certification.
PEDIATRIC SYSTEM	JAI DURGA MEDICAL AND SURGICAL AGENCIES Manufacturer(M/S HPitkar Orth tools	<ol style="list-style-type: none">Pediatric Standard 1 Hole ClampPediatric Standard Clamp -Hinge ConnectionPediatric A-Angle Attachment ClampPediatric Stacked 3 Hole Overhang ClampPediatric Rail End Cap-FemalePediatric Rail Arc Clamp	<ol style="list-style-type: none">Client base with reference of the supply order on letter head of the bidder / importer for any of the three years in last 5 years along with satisfactory performance certificate of the minimum one installation (Copy of reference supply order and satisfactory performance needs to be attached).Copy of CDSCO Registration Certification.	

	PVT Ltd	7. Pediatric Rail End Cap-Male 8. Pediatric Rail Hinge-Female 9. Pediatric Arc small 90° 10. Pediatric Arc Medium 120° 11. Pediatric Arc large 180° 12. Pediatric Rail Hinge-Male 13. Rail Bolt M6 Cannulated 28mm 14. Rail Bolt M6 Cannulated 38mm 15. Pediatric Rail Arc Rotation End 16. Rail Bolt M6 27mm 17. Rail Bolt M6 30mm 18. Pediatric Rail 150mm 19. Pediatric Rail 200mm 20. Pediatric Rail 250mm 21. Pediatric Rail 300mm 22. Rail Set Screw M6X7.5mm 23. Locking Nut M6X6mm 24. Rail Bolt M6*7mm 25. Rail Bolt M6*5mm 26. Drill Guides 6*3.2*40mm 27. Drill Guides 6*3.5*80mm 28. Drill Guides 6*5*40mm 29. Drill Guides 6*5*80mm 30. Pediatric Rotation End Connector 31. Guide wire sleeve 5*1.8*40mm 32. Guide wire sleeve 5*1.8*180mm 33. Compression And Distraction Device 34. Allen Key 6mm 35. Half pin Introducer D Type 36. 90° screw Driver 37. Drill bit cannulated 3.2x1.8X200mm. 38. 3mm straight screw driver 39. Container for pediatric fixator instrument set 40. Drill Bit 3.2X 200mm
--	---------	--

The bidder(s) are requested to upload their infirmities within 07 days from the issuance of this notice on website www.jkmsclbusiness.com and in case, the file(s) is/are heavy and the bidder are not in position to get the same uploaded on JKMSCL website, the same may be sent to corporation on the email of tender inviting authority gmjkmscl.mil@gmail.com which shall be forwarded to technical evaluation sub-committee for scrutiny and recommendations in the matter.

In case the documents are not submitted in time, your bid shall be evaluated only on the basis of documents submitted earlier.

(Mohd. Ashraf Choudhary) JKAS
General Manager (Adm)
JKMSCL,

No.NIT/JKMSCL/GM (ADM)/2024/7439-40

Dated: 28/02/2024.

Copy to the:

1. Managing Director, JKMSCL.
2. I/C Website to upload the minor infirmities on the official website of JKMSCL.
3. Office Copy
4. e-mail id :- gmjkmscl.mil@gmail.com.