



JKMSCL

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Kashmir, Near Haj House, Bemina 190018, Srinagar

email: mdjkmscl2@gmail.com; website: www.jkmsclbusiness.com

C O R R I G E N D U M

In light of the representation(s) submitted by the prospective bidder(s) thereof for the finalization of Rate Contract for the procurement of "Machinery & Equipments" uploaded vide No. Mach/2025/667 dated 23.07.2025, the **amendments in the technical specifications of the following item** as recommended by the technical experts are mentioned below :

S.No.	Name of the item : Intra Aortic Ballon Pump (IABP)	
	Existing Specifications	Ammendments made thereof.
i.	Automatic in-vivo calibration function, when Fiberoptic balloon is in use	Automatic in-vivo/one time calibration function when Fiberoptic balloon is in use.
ii.	Automatic in vivo recalibration every 2 hours or sooner should patients or environmental conditions change when Fiberoptic balloon is in use	Automatic in vivo re-calibration every 2 hours (or sooner) or Advanced installed software should automatically re-calibrate the system should the patient's or environmental conditions change when Fibre optic balloon is in use

Please Note:

1. Those firms/bidders who have already uploaded their bids are required to re-upload their bids as per amendments and corrigendum issued thereof.
2. All the bidders are requested to keep themselves updated & submit their e-bids through e-portal as per specifications & BOQs. The amendments/modifications shall be available on e-Portal and www.jkmsclbusiness.com.

(Mohd. Ashraf Choudhary) JKAS
General Manager (Adm),
JKMSCL.

No.: JKMSCL/Corg/2025/6169-72

Copy for information to the: -

1. General Manager-(K), JKMSCL.
2. Dy. General Manager (Tendering), JKMSCL.
3. P.A to Managing Director.
4. Assistant Programmer (M), JKMSCL for uploading on web portal.
5. File

Dated: 20 .09.2025.