## JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu
Corporate Office: Kashmir, Near Haj House, Bemina 190018, Srinagar
email: <a href="mailto:ikmscliepm@gmail.com">ikmscliepm@gmail.com</a>; website: <a href="www.ikmsclbuisness.com">www.ikmsclbuisness.com</a>

## CORRIGENDUM

In light of the representation(s) submitted by the prospective bidder(s) thereof for the procurement of "Machinery & Equipments" uploaded vide No. JKMSCL/Mach/2023/603 dated 05.09.2023, the amendments in the technical specifications as recommended by the technical experts for "Machinery & Equipment" is annexed as Annexure I (01 Page).

Accordingly, the critical dates are as under: .

- 1. Last date and time for submission of online bids: 04.12.2023 upto 1600 hrs
- Date and time for online opening of technical bids: 06.12.2023 at 1100 hrs

## Please Note:

JKMSCL

- Those firms/bidders who have already uploaded their bids are required to reupload their bids as per amendments and corrigendum issued thereof.
- All the bidders are requested to keep themselves updated & submit their e-bids through e-portal as per specifications & BOQs. The amendments/modifications shall be available on e-Portal and <u>www.jkmsclbusiness.com</u>.
- Although the rates for each instrument are to be quoted seprately wherever sets
  have been asked, the L1 shall be ascertained only as per commulative rates of
  each set of instruments.

(Mohd. Ashraf Choudhary) JKAS

General Manager (Adm), JKMSCL

No.: JKMSCL/Corg/2023/3817-18 Copy for information to the:-

1. Managing Director, JKMSCL.

General Manager-(Adm), JKMSCL.

3. File

Dated: 23 .11.2023

No	JKMSCL Specifications for which the	SLICE CT	
	representations have been submitted by	Amendments Recommended	Remarks If any
.1	Minimum physical/Digital tilt of 20 degree on either side.	NII	Digital tilt a must to reduce exposure to staff and reduce scanning time
3.4	Output Capacity (actual and not effective) ≥7.2 KW or more	Output Capacity (actual and not effective) >72KW or more	Typing error
	Anode heat dissipation rate >=1MHU per minute	Anode heat dissipation rate 1.2MHU per minute or more	
4.2	Scannable range ≥1800 mm	NI	
6.3	Dynamic CTA Enabling 4D CT DSA ,time resolved perfusion with a minimum range of cm	Dynamic CTA Enabling 4D CT DSA ,time resolved perfusion with a minimum range of 8cm or more	
6.4	CT-guided intervention 3D fluroscopy	Same specification is written against point 10.4.6	Can be removed here to remove repetition and ambiguity
9.3	Software Should perform the functions like scanning image Reconstruction, film documentation, MPR,CT angiography,MIP,3D ,vrt,3d SSD, fly through, readymade perfusion for stroke imaging.	Software Should perform the functions like scanning image Reconstruction, film documentation, MPR,CT angiography,MIP,3D ,vrt,3d SSD, fly through on both console and workstation. Readymade perfusion for stroke imaging on console or workstation	For wider participatio n
9.6	Following Applications on all 2 concurrent users: a)Stroke Analysis & Quantification b)Complete Liver Volumetry &	Following Applications on all 2 concurrent users on workstation : a)Stroke Analysis & Quantification b)Complete Liver Volumetry & Segmentation	
10.4.6	time needle visualization with monitor of 18" or more,cart for carrying monitor and foot switch.Please quote the price	CT fluroscopy guided biopsy with near real time needle visualization with facility for depth and angle calculation with monitor of 18" or more, cart for carrying monitor and foot suitch please quote the price separetely	
13.4	separetely  Offered system should be BIS European CE with 4 digit notified body no/ USFDA certified	Offered system should be BIS/European CE with 4 digit notified body no/ USFDA /CDSCO certified	

0 Vi.V

Dr Aijaz Hakeem

(Bruc. Dy.)

Dr Waree Hand

by shat Am

or Majed