



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of Govt of Jammu & Kashmir)

Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar, Jammu

Corporate Office Kashmir: at Bemina Bye Pass (Near Haji House) Srinagar

Telephone: 0191-2580842, Fax: 0191-2581845 (Jammu); Telefax: 0194-24920662 (Srinagar)

MINUTES OF MEETING **CONSIDERATION OF TECHNICAL BIDS** **30-04-2025, 2:00 P.M**

Sub: Finalization of tenders invited for procurement of "Hospital Furniture" Uploaded vide No: JKMSCL/Hosp-Fur/2024/623 Dated: 10-06-2024.

1. e- Bid for the finalization of Rate Contract for the procurement of "Hospital Furniture" was uploaded vide No: JKMSCL/Hosp-Fur/2024/623 Dated: 10-06-2024.
2. As per tender notification Pre-Bid meeting was held on scheduled date i.e. 27-06-2024 at 11: A.M at Corporate Office Jammu and Srinagar. A corrigendum was issued vide No. JKMSCL/ GM (Adm)/2024/2627-30 Dated: 12-07-2024, wherein some specifications of the tendered items were amended. It came to the notice that due to typographic error under below entioned Tender IDs, faulty nomenclature of items is being shown:
 - a. **Tender ID: 2024_HME-249844_27,**
 - b. **Tender ID: 2024_HME_-249844_9**
3. Accordingly, a corrigendum was issued vide No: JKMSCL/GM/e-NIT-592/2023/9160-64 Dated: 26-08-2023 and all the bidders were informed that for Tender ID:2024_HME-249844_27, rates offered shall be treated for Gynae Examination Table and for Tender ID: 2024_HME_-249844_9, rates offered shall be treated for Stretcher Trolley.
Scheduled date for receipt and opening of tenders were extended upto 01-08-2024 upto 1600 hrs and 03-08-2024 at 1100 hrs.and 09-08-2024 upto 1600 hrs and 10-08-2024 at 1100 hrs respectively.
Accordingly, e-bids were opened on 10-08-2024 at 11.00 a.m., in the presence of Technical Evaluation Committee constituted by Boards of Directors during 1st Board Meeting and amended by Boards of Directors during 3rd Board Meeting, JKMSCL for the purpose, and as per the system generated report, bids received against are attached as **Annexure "A"**.
4. After opening of e-bids, compilation, preliminary scrutiny and evaluation of bid documents was carried out by the Technical Evaluation Sub Committee constituted for the purpose.
5. Documents thus found deficit by the subcommittee were notified for the information of the participating bidders vide Notice No: JKMSCL/GM(P&S) / 2024/ 5750-54 dated 05-10-2024 with directions to submit deficit documents latest by 14-10-2024, which was later extended till 22-10-2024. Documents thus submitted by the bidders were again evaluated and a final report was submitted by the subcommittee.
6. Technical Evaluation/ Advisory Committee after perusal of Report observed that most of the bidders are not recommended by the subcommittee for non-submission of Quality certificates/Statement of Past Supplies & Performance/Proof of Supply of minimum 50% of the quoted item.
7. After threadbare discussions, Technical Evaluation/ Advisory Committee in its meeting held on 16-01-2025 decided to give another chance to the participating bidders, to submit deficit documents within 7 days, so as to ensure healthy participation among bidders. However, for submission of Statement of Past Supplies & Performance (ANNEXURE-G) of NIT and Proof of Supply of minimum 50% of the projected quantity to any Govt. Organization/Corporations/ PSU etc. under mentioned proforma was devised for the ease of bidders.

Handwritten signatures and initials in blue ink.

Financial Year	Item Quoted in NIT 623		Tentative Qty Projected in NIT	Detail of institutes where supplied	Qty supplied	Ref. To supply order
	Item Code	Name of Item				

It was also decided that the bidders can mention same/similar items against the tendered item as Proof of Supplies

8. As decided by Technical Evaluation/Advisory Committee a final notice was issued vide No: JKMSCL/GM(P&S)/2024/12168-72 dated 18-01-2025 with directions to submit deficit documents latest by 28-01-2025.

9. Documents thus submitted by the bidders were again evaluated and a final report was submitted which is attached herewith as "**Annexure "B". Details of the bidders recommended / not recommended by the subcommittee for further processing is as under:**

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
1.	WARSI MEDICO SYSTEMS AND SERVICES (Bidder/Manufacturer/MSME) Total Quoted Items-34	Not Eligible to TEC as per documents submitted	To Submit: 1. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB) 2. CDSO/Application submitted to concerned department as per the notification dated:16-05-2024 for the following items codes: HFS-002; HFS-003; HFS-011; HFS-021; HFS-025; HFS-030 HFS-035; HFS-036; HFS-037	Not Submitted Not Submitted for: HFS-011 (Class C) HFS-021 HFS-025 (Class C) HFS-030 HFS-036 HFS-037	Not Eligible to TEC as per documents submitted
2.	LIFE LINE ENGINEERING Gujarat (Bidder/Manufacturer/MSME) Total Quoted Items-19 (As per Annexure E)	Not Eligible to TEC as per documents submitted	No Infirmities Submitted		No Change
3.	BANSAL TRADING COMPANY (Bidder) M/s Geeta Steel Furniture Gujarat (Manufacturer) Total Quoted Items-21	Not Eligible to TEC as per documents submitted	To Submit: 1. ISO-9001 Certification for the manufacturer (Submitted document is not accredited to NABCB) 2. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB) 3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB) 4. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB) 5. Test certificate from Govt. approved/NABL Laboratory for Test procedures like impact test, Bend test, Salt Spray Chamber test, Epoxy powder coating and Phosphate coating for the quoted items. (Submitted documents are not from Govt. approved lab/NABL) 6. Certificate issued through Chattered accountant that manufacturer have Seven (7)	Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted (Submitted document is not from Govt. Agency) Not Submitted	Not Eligible to TEC as per documents submitted

Yu *the* *a* *key* *to*

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
			<p>Tanks Hot process and epoxy powder coating of minimum 50 microns with phosphate layer underneath for corrosion resistance pre-treatment plant and Edge bending machine installed in the manufacturing premises. (Submitted document is not from CA)</p> <p>7. List of items for quality control measures including details of quality control laboratory (Submit document certified from CA)</p> <p>8. Certificate from Govt. agency/Chartered Engineer for production capacity assessment (Submit document from Govt. Agency)</p> <p>9. Statement of Past supplies & Performance (Submit Annexure G separately for each quoted item along with the quantity supplies highlighting the item code)</p> <p>10. Proof of supplies of at least 50% of the projected quantity to any Govt. Organization, Corporations/ PSUs (Copies of orders to be enclosed with proper highlighting the quantity supplies against each quoted item code)</p> <p>11. CDSCO/Application submitted to concerned department as per the notification dated:16-05-2024 for the following items codes:</p>	<p>(Submitted document is not from CA)</p> <p>Not Submitted</p> <p>Not Submitted</p> <p>Submitted (Not for 3 years)</p> <p>Not Submitted</p> <p>Not Submitted</p>	
4.	MDT INTERNATIONAL Ambala (Bidder/Manufacturer/MSME) Total Quoted Items-34	Not Eligible to TEC as per documents submitted	Not Submitted any Infirmities		No Change
5.	NARULA UDYOG (INDIA) PVT LTD Delhi (Bidder/Manufacturer/MSME) Total Quoted Items-17	Not Eligible to TEC as per documents submitted	Not Submitted any Infirmities		No Change
6.	S.S AGENCIES, Jammu (Bidder) M/s Vinod Medical System Pvt. Ltd. Gujarat (Manufacturer) Total Quoted Items-14	Not Eligible to TEC as per documents submitted	<p>To Submit:</p> <p>1. ISO-9001 Certification for the manufacturer (Submitted document is not accredited to NABCB)</p> <p>2. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB)</p> <p>3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB)</p> <p>4. ISO-13485:2016 Certification for quality assurance (Submit document accredited to NABCB)</p>	<p>Not Submitted</p> <p>Not Submitted</p> <p>Not Submitted</p> <p>Not Submitted</p>	Not Eligible to TEC as per documents submitted
7.	L M AGENCIES Srinagar (Bidder) M/s Midmark (India) Pvt ltd Mumbai (Manufacturer) Total Quoted Items-22	Not Eligible to TEC as per documents submitted			Eligible to TEC as per documents submitted for all quoted items except: HFS-017; HFS-018 ; HFS-023 HFS-024

Yu

B S

em

11

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
8.	M P SURGICAL Delhi (Bidder/Manufacturer/ MSME) Total Quoted Items-22	Eligible to TEC as per documents submitted			No Change
9.	S K SCIENTIFIC AND SURGICALS Ambala (Bidder/Manufacturer/ MSME) Total Quoted Items-1	Not Eligible to TEC as per documents submitted			Eligible to TEC as per documents submitted for all quoted items except HFS-021
10.	PMT HEALTHCARE PVT LTD Gujarat (Bidder/Manufacturer/ MSME) Total Quoted Items-15	Not Eligible to TEC as per documents submitted	To Submit: 1. ISO-14001 Certification for environment friendly features (Submit document accredited to NABCB) 2. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB and not latest also) 3. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB)	Not Submitted Not Submitted Not Submitted	No Change
4.	MEDSOR IMPEX New Delhi Bidder/Manufacturer/ MSME) Total Quoted Items-34	Not Eligible to TEC as per documents submitted.	Not Submitted any minor infirmities		No Change
5.	BHAVNA STEEL EQUIPMENTS Ahmedabad (Bidder/Manufacturer/ MSME) M/s Bhardwaj Agencies (Authorized Representative) Total Quoted Items-27	Not Eligible to TEC as EMD & Cost of tender processing fees not submitted	To Submit: 1. UDYAM Certificate showing the Latest classification of unit from the concerned department (Submitted document is not latest) (Resubmit the classification of unit at the time of submission of technical documents for the Nit-623 of JKMSCL.	Not Submitted latest certificate (Manufactu rer is medium enterprises when verified online)	No Change
13.	IMED FURNS PRIVATELIMITED Chennai (Bidder/Manufacturer/ MSME) Total Quoted Items-22	Not Eligible to TEC as per documents submitted	To Submit: 1. ISO-9001 Certification for the manufacturer (Submitted document is not accredited to NABCB) 2. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB) 3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB) 4. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB) 5. Stainless Steel grade certificate (SS 304) from Govt. approved testing laboratory duly mentioning about the chemical composition of Items manufactured by the firm. 6. Test certificate from Govt. approved/NABL Laboratory for Test procedures like impact test, Bend test, Salt Spray Chamber test, Epoxy powder coating and Phosphate coating	Not Submitted Not Submitted Not Submitted Not Submitted	No Change

Yu ser B st 12 my W

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
			for the quoted items. 7. Certificate issued through Chartered accountant that manufacturer have Seven (7) Tanks Hot process and epoxy powder coating of minimum 50 microns with phosphate layer underneath for corrosion resistance pre-treatment plant and Edge bending machine installed in the manufacturing premises.	Not Submitted	
14.	CAREVEL MEDICAL SYSTEMSPVT. LTD U.P (Bidder/Manufacturer /MSME) Total Quoted Items-15	Eligible to TEC as per documents submitted except item Code HFS-11 and 25			Eligible to TEC as per documents submitted
15.	M/s New India Enterprises Srinagar (Bidder) M/s United Surgical Industries,UP (Manufacturer) Total Quoted Items-22	Eligible to TEC as per documents submitted for all the quoted items except item code HFS-11, HFS-24, HFS-25 HFS-026			Eligible to TEC as per documents submitted for all the quoted items except HFS-026
16.	METHODEX SYSTEMS PVT LTD Indore (Bidder/Manufacturer) Mr. Idrees Farooq, J&K (Authorized representative) Total Quoted Items-30	Eligible to TEC as per documents submitted for the following item codes: HFS-006 HFS-014 HFS-016 HFS-013 HFS-017 HFS-022 HFS-023	To Submit: 1. List of quoted items with Name of the manufacturer as per the Annexure-U (Bidder has submitted Annexure-E for some items which are not quoted in Annexure-U) [Resubmit the list of quoted items] 2. Certificate from Govt. agency/Chartered Engineer for production capacity assessment [Submitted document is not from Govt. Agency] 3. CDSO /Application submitted to concerned department as per the notification dated:16-05-2024 for the following items codes: HFS-002; HFS-003; HFS-011 (Class C) HFS-021; HFS-025 (Class C) HFS-030; HFS-035; HFS-036 HFS-037 4. Proof of supplies of at least 50% of the projected quantity to any Govt. Organization, Corporations/ PSUs (Copies of orders to be enclosed with proper highlighting the quantity supplies against all the item codes except HFS-003; HFS-004; HFS-007; HFS-008; HFS-009; HFS-010 HFS-011; HFS-015; HFS-018; HFS-020; HFS-024; HFS-026 HFS-030; HFS-031; HFS-032; HFS-034; HFS-035; HFS-036 HFS-037; HFS-038; HFS-039; HFS-040	Not Submitted Submitted is not from Govt agency Submitted document is without any list of items Submitted atleast 50% of projected quantity for: HFS-002 HFS-006 HFS-014 HFS-016 HFS-013 HFS-017 HFS-021 HFS-022	NO Change Eligible to TEC as per documents submitted for the following item codes: HFS-006 HFS-014 HFS-016 HFS-013 HFS-017 HFS-022 HFS-023
17.	DEEPAK INDUSTRIES Jammu (Bidder/Manufacturer/ MSME) Total Quoted Items 33 (Based on affidavit submitted)	Not Eligible to TEC as per documents submitted	Not Submitted any infirmities		No Change

Handwritten signatures and initials:
 ✓
 ✓
 ✓
 ✓
 ✓

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
18.	MICROTEKNIK AMBALA (Bidder/Manufacturer/ MSME) Total Quoted Items-34	Not Eligible to TEC as per documents submitted	To Submit: 1. ISO-9001 Certification for the manufacturer (Submitted document is not accredited to NABCB) 2. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB) 3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB) 4. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB) 5. Stainless Steel grade certificate (SS 304) from Govt. approved testing laboratory duly mentioning about the chemical composition of Items manufactured by the firm. 6. Test certificate from Govt. approved/NABL Laboratory for Test procedures like impact test, Bend test, Salt Spray Chamber test, Epoxy powder coating and Phosphate coating for the quoted items. 7. Certificate issued through Chartered accountant that manufacturer have Seven (7) Tanks Hot process and epoxy powder coating of minimum 50 microns with phosphate layer underneath for corrosion resistance pre-treatment plant and Edge bending machine installed in the manufacturing premises. 8. List of items for quality control measures including details of quality control laboratory. 9. CDSKO /Application submitted to concerned department as per the notification dated:16-05-2024 for the following items codes: HFS-002; HFS-003; HFS-011 (Class C) HFS-021; HFS-025 (Class C) HFS-030; HFS-035 HFS-036; HFS-037	Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted	No Change
19.	LIFECARESYSTEM Haryana (Bidder) M/s Genuine Medica Pvt Ltd New Delhi (Manufacturer) Total Quoted Items-31 (not clarified)	Not eligible to TEC as per documents submitted	Not Submitted EMD & Cost of Tender Processing fees, therefore cannot be evaluated		No Change
20.	MEDIMEK INDUSTRIES Maharashtra (Bidder/Manufacturer/ MSME) Total Quoted Items-22	Not eligible to TEC as per documents submitted			Eligible to TEC as per documents submitted
21.	M/s NILKAMAL LIMITED Mumbai (Bidder/Manufacturer) Total Quoted Items-21	1. Not Eligible to TEC as per documents submitted	2. Statement of Installed Manufacturing Capacity (Annexure-E) 3. Statement of Past supplies & Performance (Submit Annexure G separately for each quoted item along with the quantity supplies highlighting the item code)	Not Submitted on stamp paper Not Submitted separately	No Change

✓

✓

✓

✓

✓

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
			<p>4. Proof of supplies of at least 50% of the projected quantity to any Govt. Organization, Corporations/ PSUs (Copies of orders to be enclosed with proper highlighting the quantity supplies against each quoted item code)</p> <p>5. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB)</p> <p>6. Certificate from Govt. agency/Chartered Engineer for production capacity assessment (Submitted document is not from Govt. Agency)</p> <p>7. CDSCO/Application submitted to concerned department as per the notification dated:16-05-2024 (Submit detailed document for the quoted)</p>	<p>Not Submitted as per format asked</p> <p>Not Submitted</p> <p>Not Submitted</p> <p>Not Submitted for all quoted items</p>	
22.	ALLIANCE IMPEX New Delhi (Bidder/Manufacturer/MSME) Total Quoted Items-11	Not Eligible to TEC as per documents submitted		Not submitted any infirmities	No Change
23.	UNIK SURGICAL PRIVATE LIMITED Patna (Bidder/Manufacturer/MSME) M/s Indo Kashmir, Jammu (Authorized representative) Total Quoted Items-32	Not Eligible to TEC as per documents submitted		Not submitted any infirmities	No Change
24.	GODREJ AND BOYCEE MFG CO. LTD Mohali (Bidder/Manufacturer) Total Quoted Items-18	Not eligible to TEC as per documents submitted	<p>To Submit:</p> <p>1. Self-attested catalogue of the following quoted items: HFS-006 HFS-010 HFS-021 (Submit for non-motorized) HFS-022; HFS-026</p> <p>2. Self-attested duly filled compliance sheet for the following quoted items: HFS-002 (without any deviation) HFS-003 (without any deviation) HFS-022; HFS-026; HFS-031</p> <p>3. Certificate issued through Chartered accountant that manufacturer have Seven (7) Tanks Hot process and epoxy powder coating of minimum 50 microns with phosphate layer underneath for corrosion resistance pre-treatment plant and Edge bending machine installed in the manufacturing premises</p>		Eligible to TEC as per major documents submitted to subjective to submission of documents at Serial number 1, 2 & 3 before physical demonstration
25.	RAJ STEEL INDUSTRIES Srinagar (Bidder/Manufacturer/MSME) Total Quoted Items-25	Eligible to TEC as per documents submitted except HFS-11, HFS-25, HFS-30			Eligible to TEC as per documents submitted
26.	HABIBA STEEL INDUSTRIES KASHMIR (Bidder/Manufacturer/MSME) Total Quoted Items-11	Not Eligible to TEC as per documents submitted	<p>1. ISO-13485:2016 Certification for quality assurance (Submit document is accredited to NABCB)</p> <p>2. EPF and ESIC registration No. of manufacturer issued by the concerned department.</p>	<p>Not Submitted</p> <p>Not Submitted</p>	No Change
27.	SANJIVANI HOSPITAL FURNITURE, MAHARASTRA	Not Eligible to TEC as per	<p>To Submit:</p> <p>1. ISO-9001 Certification for the manufacturer (Submitted)</p>	Not Submitted	No Change

Yes

Yes

Yes

Yes

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
	(Bidder/Manufacturer/MSME) Total Quoted Items-10	documents submitted.	<p>document is not accredited to NABCB)</p> <p>2. ISO-14001 Certification for environment friendly features (Submitted document is not accredited to NABCB)</p> <p>3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submitted document is not accredited to NABCB)</p> <p>4. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB)</p>	Not Submitted Not Submitted Not Submitted	
28.	L.K INDUSTRIES SRINAGAR (Bidder/Manufacturer/MSME) Total Quoted Items-06	Not Eligible to TEC as per documents submitted	<p>To Submit:</p> <p>1. ISO-9001 Certification for the manufacturer (Submitted document is not accredited to NABCB)</p> <p>2. ISO-14001 Certification for environment friendly features (Submit document accredited to NABCB)</p> <p>3. ISO-45001/ OHSAS 18001 Certification for occupational health safety management (Submit document accredited to NABCB)</p> <p>4. ISO-13485:2016 Certification for quality assurance (Submitted document is not accredited to NABCB)</p> <p>5. Stainless Steel grade certificate (SS 304) from Govt. approved testing laboratory duly mentioning about the chemical composition of items manufactured by the firm.</p> <p>6. Test certificate from Govt. approved/NABL Laboratory for Test procedures like impact test, Bend test, Salt Spray Chamber test, Epoxy powder coating and Phosphate coating for the quoted items.</p> <p>7. Certificate issued through Chartered accountant that manufacturer have Seven (7) Tanks Hot process and epoxy powder coating of minimum 50 microns with phosphate layer underneath for corrosion resistance pre-treatment plant and Edge bending machine installed in the manufacturing premises.</p> <p>8. EPF and ESIC registration No. of manufacturer issued by the concerned department.</p> <p>9. List of items for quality control measures including details of quality control laboratory.</p> <p>10. Certificate from Govt. agency/Chartered Engineer for production capacity assessment.</p>	Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted Not Submitted	No Change
29.	ROYAL SAFE COMPANY, HARYANA (Bidder/Manufacturer) Total Quoted Items-8	Eligible to TEC as per documents submitted except HFS-003			Eligible to TEC as per documents submitted
30.	MEDITECH HEALTHCARE HARYANA	1. Not Eligible to TEC as per	Not submitted any infirmities		No Change

Yu

on 24/11/20

S. No	Bidders	Remarks Report-I	Shortcomings	Status	Remarks Report-II
	(Bidder/Manufacturer/MSME) Total Quoted Items-01	documents submitted			
31.	M/S BALAJI IMPEX DELHI (Bidder)	Not Eligible to TEC as per documents submitted	Not submitted any infirmities		No Change

10. Samples of the qualified bidders were evaluated by the team of experts from Machinery Division, Jammu and Directorate of Health Services, Jammu at Sample Ware House, JKMSCL Gangyal Jammu w.e.f. 03-04-2025 in scheduled manner, in presence of representative of the firms. Sample Evaluation report thus submitted by the Technical experts is attached as **Annexure "C"**, however, Firm wise details are as under:

1. Godrej and Boyce Mfg Co ltd

S. No	Code	Name of Item	Compliance / Deviation
1	HFS-002	FOWLER BED	Rejected Not as per tender Specifications <ul style="list-style-type: none"> Size of Knee Rest: Over all - 470 mm L x 800mm W. Only Section : 307 mm L x 800 mm W. Size of Leg Rest: Over all - 560 mm L x 800mm W . Only Section : 610 mm L x 800mm W. Size of Fix Section: Over all - 235 mm L x 800mm W. Only Section : 610 mm L x 800mm W.
2	HFS-003	SEMI FOWLER BED	Rejected Not as per tender Specifications <ul style="list-style-type: none"> Buffer to Buffer: (2080 mm L (Range 2097 – 2182 mm) x 900 mm W) (+2% variation) X 560 mm H (Without Mattress). (Less Than Limits Bed Frame size:2070 mm L x 900 mm W. Back Rest Only Section: 710 mm L x 800 mm W. Body Rest Only Section 910 x 800 mm The bed has Laminated head and foot panels.
3.	HFS-004	PLAIN HOSPITAL BED	Rejected Not as per tender Specifications <ul style="list-style-type: none"> Overall size (Approx): 2030 (range 2048-2131 mm) mm (L) x 900mm (W) x 615 mm. (H)) (± 2%) (Without Mattress & From Ground to Top Frame). Bed Platform frame size : Approx 1900 mm L x 900 mm W. 4 Nos Flat Stiffeners 60 x 12 mm The head bow and leg bow is having inside bending radius 100 mm.2 head bow height 1060 mm and leg bow height 790 mm connected to bed with bracket size 106 mm x 25 mm x 1.2 mm (18 G). Nil Mattress Arrestor
4.	HFS-006	Bed Side Locker	Approved as per tender Specifications
5.	HFS-008	INSTRUMENT TROLLEY	Rejected Not as per tender Specifications <ul style="list-style-type: none"> The height of railing is 2" (50 mm). The distance between the two shelves is 580 mm and distance from ground to bottom shelf is 290 mm
6.	HFS-009	CRASH CART	Sample Not Provided
7.	HFS-010	DRESSING TROLLEY	Sample Not provided
8.	HFS-011	STRETCHER TROLLEY	Sample Not Provided
9.	HFS-013	EXAMINATION COUCH	Rejected Not as per tender Specifications <ul style="list-style-type: none"> NO Stainless Steel
10.	HFS-017	SALINE STAND WITH SS ROD	Sample Not Provided
11.	HFS-018	REVOLVING STOOL	Rejected Not as per tender Specifications <ul style="list-style-type: none"> Approx 480 mm L to 670 mm S.S 304 revolving top having diameter 300 mm x 0.7 mm (22 G) mounted on MS Tubular tripod base.

Vin

San

✓ X

le y

W

12.	HFS-021	BIRTHING BED (Non Motorized)	<ul style="list-style-type: none"> • Four Legs. Sample Not provided
13.	HFS-022	WHEEL CHAIR	Sample Not provided
14.	HFS-23	Three fold Bed side screen	Rejected Not as per tendered specifications <ul style="list-style-type: none"> • Middle Span 1070 mm Green Color
15.	HFS-026	HOSPITAL LOCKER PIGEON HOLE	Sample Not provided
16.	HFS-031	OVER BED TABLE	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Overall Size: 1080 mm L x 380 mm W x 950 mm H. • Top Size: 850 mm L x 380 mm W.
17.	HFS-038	WHEEL CHAIR FOLDING	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Overall dimensions: 1020 mm L x 660 mm W x 890 mm H. • Back support height above seat : 375 mm. • Seat width : 430 mm W x 400 mm L. • Hoop Of Steel (Chrome Plated) foamed seat
18.	HFS-040	INSTRUMENT CABINET	Sample Not provided

2. Royal Safe Company

S. No	Code	Name of Item	Compliance/Deviation
1.	HFS-002	FOWLER BED	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Fixed Height of 610 mm. • Size of Knee Rest: Over all - 730 mm L x 880mm W. • Size of Leg Rest: Over all - 600 mm L x 800mm W . • Bracket Size 40 mm • Knee Rest 360 mm L • Leg Rest 620 mm
2.	HFS-003	SEMI FOWLER BED	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Buffer to Buffer: (2100 mm L x mm W) (+2% variation) X 560 mm H (Without Mattress). • Bed Frame size:2010 mm L x 910 mm W. • Back Rest Only Section: 730 mm L x 800 mm W. • Body Rest Only Section 1080 x 800 mm • round brackets size 40 mm
3.	HFS-004	PLAIN HOSPITAL BED	Rejected Not as per tender Specification <ul style="list-style-type: none"> • sleeping level at a height of 290 mm • Nil Mattress Arrestor
4.	HFS-006	Bed Side Locker	Rejected Not as per tender Specification <ul style="list-style-type: none"> • One drawer 100mm H x 330mm W x 360mm D fitted is provided below the top • Paint Quality and Workmanship not up to the mark (welding marks and holes in sample)
5.	HFS-018	REVOLVING STOOL	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Approx 580 mm L to 810 mm • S.S 304 revolving top having diameter 300 mm
6.	HFS-026	HOSPITAL LOCKER PIGEON HOLE	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Knobs not provided
7.	HFS-031	OVER BED TABLE	Rejected Not as per tender Specification <ul style="list-style-type: none"> • Overall Size: 1080 mm L x 450 mm W x 1000 mm H. • Top Size: 800 mm L x 450 mm W.
8.	HFS-032	PERFORATED STEEL BENCH	Approved as per tender Specification

Methodex Systems

S.	Code	Name of Item	Compliance /Deviation
----	------	--------------	-----------------------

Vu

ANW

B

B

ky

hs

No			
1.	HFS-006	Bed Side Locker	Rejected Not as per tender Specification <ul style="list-style-type: none"> One drawer 95 mm H x 340mm W x 380mm D fitted is provided below the top
2.	HFS-013	EXAMINATION COUCH	Rejected Not as per tender Specifications <ul style="list-style-type: none"> 50 MM Provided Size of Foot Step is 360 mm L x 330 mm W x 205 mm H above floor.
3.	HFS-014	SOILD LINEN TROLLEY	Approved as per tender Specification
4.	HFS-016	KICK BUCKET	Approved as per tender Specification
5.	HFS-017	SALINE STAND WITH SS ROD	Approved As per tender Specifications
6.	HFS-022	WHEEL CHAIR	Sample Not Provided
7.	HFS-23	Three fold Bed side screen	Approved As per tender Specifications

M P Surgical

S. No	Code	Name of Item	Compliance / Deviation
1	HF S-002	FOWLER BED	Rejected As per tender Specifications <ul style="list-style-type: none"> Size of Knee Rest: Over all - 740 mm L x 800mm W. Only Section : 380 mm L x 800 mm W. Four corner Plastic buffers of 100 MM dia for with castor application are provided in place of rubber. Mattress Provided with four section 4" thick PU Foam of 40 density covered PVC Rexene cover. The Back Rest size is 760 mm L x 850mm W x 100 mm T, Fix section size is 180 mm L x 850mm W x 100 mm T, Knee Rest size is 460 mm L x 850 mm W x 100 mm T and Leg Rest size is 610 mm L x 850 mm W x 100 mm T.
2	HF S-003	SEMI FOWLER BED	Accepted As per tender Specifications <ul style="list-style-type: none"> Buffer to Buffer: (2120 mm L x 1020mm W) (+2% variation) X 600mm H (Without Mattress). There are four locations on the bed platform to hold stainless steel Telescopic Saline rod 12mm dia with 30mm dia x 1.2mm (18G) stainless steel outer covering tube with a knob to mount syringe pump.
3.	HFS-004	PLAIN HOSPITAL BED	Approved As per tender Specifications <ul style="list-style-type: none"> Overall size (Approx): 2080mm (L) x 890mm (W) x 580 mm. (H) (± 2%) (Without Mattress & From Ground to Top Frame).
4.	HFS-006	Bed Side Locker	Rejected As per tender Specifications <ul style="list-style-type: none"> 40 cms x 40 cms x 81 cms H. One drawer 100mm H x 370mm W x 330mm D fitted is provided below the top.
5.	HFS-007	Adjustable Bed Side Table	Rejected As per tender Specifications <ul style="list-style-type: none"> Size: Top 810mm L x 455mm W. Height adjustment from 800 mm to 1080mm
6.	HF S-008	INSTRUMENT TROLLEY	Approved As per tender Specifications <ul style="list-style-type: none"> The distance between the two shelves is 485 mm and distance from ground to bottom shelf is 340 mm
7.	HFS-009	CRASH CART	Rejected Not as per tender Specifications <ul style="list-style-type: none"> Two lockable box units made of high impact polystyrene with 3 drawers should have dimension 305 mm L x 360 mm D x 310 mm H. The size of middle and bottom shelf is 650 mm L x 380 mm W made from SS 304 0.9 mm (20 G) sheet. The shelf are provided with railing on three sides. This railing is made from SS 304 Rod of diameter 10 mm. The size of top shelf is 670 mm L x 235 mm W made from SS 304

Ku

or

2

6

6

4

W

			0.9 mm (20 G) sheet. Two side support size 237 mm W and 45 mm made from SS 304 2.0 mm (14 G) sheet provided to top shelf. <ul style="list-style-type: none"> Pullout cardiac massage board made of MDF of minimum size 710 mm x 300 mm x 12 mm laminated on top and bottom of laminate of 1 mm and 0.6 mm respectively. MDF shall have water resistance property and it should be made from eco-friendly material.
8.	HF S-010	DRESSING TROLLEY	Accepted As per tender Specifications Overall Size: 1010 mm L x 500 mm W x 920 mm H.
10.	HFS-013	EXAMINATION COUCH	Rejected Not As per tender Specifications <ul style="list-style-type: none"> Overall size: 1900 mm L x 910 mm W x 840 mm Stainless steel legs not provided Sliding foot step under the front side of lower middle cabinet made from 0.9 mm (20 G)MS CRCA Sheet. Size of Foot Step is 350 mm L x 330 mm W x 220 mm H above floor.
13.	HFS-016	KICK BUCKET	Approved As per tender Specifications
14.	HFS-017	SALINE STAND WITH SS ROD	Approved As per tender Specifications
15.	HFS-018	REVOLVING STOOL	Rejected Not As per tender Specifications Height adjustable from Approx 575 mm L to 800 mm with accurately machine screw mechanism..
16.	HFS-20	DOUBLE STEP STOOL	Approved As per tender Specifications
19	HFS-23	Three fold Bed side screen	Rejected Not As per tender Specifications <ul style="list-style-type: none"> curtains colour is green
20	HFS-024	CYLINDER TROLLEY	Rejected Not As per tender Specifications
22.	HFS-026	HOSPITAL LOCKER PIGEON HOLE	Approved As per tender Specifications
23.	HFS-030	STRETCHER FOLDING	Approved As per tender Specifications
24.	HFS-031	OVER BED TABLE	Approved As per tender Specifications
26.	HFS-034	SURGEON'S STOOL WITH BACK REST	Rejected Not As per tender Specifications <ul style="list-style-type: none"> Single Wheel Three leg
27.	HFS-035	GYNAE EXAMINATION TABLE (PLAIN)	Rejected Not As per tender Specifications <ul style="list-style-type: none"> No ankle Straps EPDM rubber tube Not available
31.	HFS-039	BACK REST	Rejected Not as per tender Specifications Back height adjustable from 170 mm to 640 mm Depth: 600mm

New India Enterprises

S. No	Code	Name of Item	Compliance /Deviation
1	HFS-002	FOWLER BED	Approved as per tendered specifications
2	HFS-003	SEMI FOWLER BED	Approved as per tendered specifications
4.	HFS-006	Bed Side Locker	Approved as per tendered specifications
5.	HFS-007	Adjustable Bed Side	Approved as per tendered specifications

Yu

an

2 2

6 4

11

6.	HFS-008	Table INSTRUMENT TROLLEY	Approved as per tendered specifications
7.	HFS-009	CRASH CART	Approved as per tendered specifications
8.	HFS-010	DRESSING TROLLEY	Approved as per tendered specifications
9.	HFS-011	STRETCHER TROLLEY	Approved as per tendered specifications
10.	HFS-013	EXAMINATION COUCH	Approved as per tendered specifications
14.	HFS-017	SALINE STAND WITH SS ROD	Approved as per tendered specifications
15.	HFS-018	REVOLVING STOOL	Approved as per tendered specifications
16.	HFS-20	DOUBLE STEP STOOL	Approved as per tendered specifications
17.	HFS-021	BIRTHING BED (Non Motorized)	Approved as per tendered specifications
18.	HFS-022	WHEEL CHAIR	Approved as per tendered specifications
19.	HFS-23	Three fold Bed side screen	Approved as per tendered specifications
20.	HFS-024	CYLINDER TROLLEY	Approved as per tendered specifications
21.	HFS-025	CRITICAL PATIENTS TROLLEY	Approved as per tendered specifications
27.	HFS-035	GYNAE EXAMINATION TABLE (PLAIN)	Approved as per tendered specifications
28.	HFS-036	SCOOP STRETCHER	Rejected - Not As per tendered specifications No Quick release buckle belts
29.	HFS-037	SPINE BOARD WITH STRAPS AND HEAD BLOCKS	Approved as per tendered specifications
32.	HFS-040	INSTRUMENT CABINET	Approved as per tendered specifications

Raj Steel industries

HFS-002	FOWLER BED	Approved as per tendered specifications
HFS-003	SEMI FOWLER BED	Approved as per tendered specifications
HFS-004	PLAIN HOSPITAL BED	Approved as per tendered specifications
HFS-006	Bed Side Locker	Approved as per tendered specifications
HFS-007	Adjustable Bed Side Table	Approved as per tendered specifications
HFS-008	INSTRUMENT TROLLEY	Approved as per tendered specifications
HFS-009	CRASH CART	Approved as per tendered specifications
HFS-010	DRESSING TROLLEY	Approved as per tendered specifications

Vu

clw

2/4

u y

v

HFS-011	STRETCHER TROLLEY	Approved as per tendered specifications
HFS-014	SOILD LINEN TROLLEY	Approved as per tendered specifications
HFS-015	WASH BASIN STAND TWO TIER	Sample not available
HFS-016	KICK BUCKET	Approved as per tendered specifications
HFS-017	SALINE STAND WITH SS ROD	Approved as per tendered specifications
HFS-018	REVOLVING STOOL	Approved as per tendered specifications
HFS-20	DOUBLE STEPSTOOL	Approved as per tendered specifications
HFS-021	BIRTHING BED (Non Motorized)	Approved as per tendered specifications
HFS-022	WHEEL CHAIR	Approved as per tendered specifications
HFS-23	Three fold Bed side screen	Approved as per tendered specifications
HFS-024	CYLINDER TROLLEY	Approved as per tendered specifications
HFS-025	CRITICAL PATIENTS TROLLEY	Approved as per tendered specifications
HFS-026	HOSPITAL LOCKER PIGEON HOLE	Approved as per tendered specifications
HFS-030	SRETCHER FOLDING	Sample not available
HFS-031	OVER BED TABLE	Approved as per tendered specifications
HFS-035	GYNAE EXAMINATION TABLE (PLAIN)	Approved as per tendered specifications
HFS-039	BACK REST	Not as per Specification Rejected
HFS-040	INSTRUMENT CABINET	Approved as per tendered specifications

Medimek Industries

1	HFS-002	FOWLER BED	Rejected Not as per Tender Specifications <ul style="list-style-type: none"> Size of Back Rest: Over all - 785 mm L x 800mm W. Only section : 705 mm L x 800 mm W. Size of Knee Rest: Over all - 470 mm L x 800mm W. Only Section : 335 mm L x 800 mm W. Size of Leg Rest: Over all - 580 mm L x 800mm W . Only Section : 610 mm L x 800mm W. Size of Fix Section: Over all - 210 mm L x 800mm W. Only Section : 610 mm L x 800mm W. Fix section size is 230 mm L x 850mm W x 100 mm T , Knee Rest size is 350 mm L x 850 mm W x 100 mm T and Leg Rest size is 620 mm L x 850 mm W x 100 mm T
2	HFS-003	SEMI FOWLER BED	Rejected Not as per Tender Specifications <ul style="list-style-type: none"> Buffer to Buffer: 1980 mm L x 920 mm W. Only Section: 740mm L x 790mm W. round brackets size 40 mm ODx38mm ID Telescopic Saline Tube
3.	HFS-004	PLAIN HOSPITAL BED	Rejected Not as per Tender Specifications SS Tube with two hooks
4.	HFS-006	Bed Side Locker	Approved As per Tendered Specifications
5.	HFS-007	Adjustable Bed Side	Approved As per Tendered Specifications

Yu

ch

2x

le

7

W

	HFS-008	Table INSTRUMENT TROLLEY	Approved As per Tendered Specifications
7.	HFS-009	CRASH CART	Rejected Not as per Tender Specifications • saline tube made of 12 mm dia
8.	HFS-010	DRESSING TROLLEY	Approved As per Tendered Specifications
9.	HFS-011	STRETCHER TROLLEY	Rejected Not as per Tender Specifications • MS flat size 20 mm x 5 mm • Stainless Steel
10.	HFS-013	EXAMINATION COUCH	Rejected Not as per Tender Specifications • Upper section of box approx size 1220 mm L x 460 mm W x 630 mm H with three sliding drawers of approx size 355mm L x 465 mm W x 75 mm H. • Size of Foot Step is 375 mm L x 330 mm W x 270 mm H above floor.
11.	HFS-014	SOILD LINEN TROLLEY	Approved As per Tendered Specifications
12.	HFS-017	SALINE STAND WITH SS ROD	Rejected Not as per Tender Specifications • Saline Stand is provided with S.S 304 Tube 12 mm diameter with two hooks , one star Knob at Top and PVC cap at bottom
13.	HFS-018	REVOLVING STOOL	Approved As per Tendered Specifications
14.	HFS-20	DOUBLE STEP STOOL	Approved As per Tendered Specifications • First step height 250 mm & second step size 460 m
15.	HFS-021	BIRTHING BED (Non Motorize d)	Rejected Not as per Tender Specifications • S.S. saline Tube with 12 mm dia S.S. rod shall telescope in SS socket tube approx 15.8 mm dia x 1.2 mm (18 G) welded on angular base bracket of 14G SS sheet. Nylon bracket provided to prevent colour damage It could be placed at two different locations • The backrest and middle section shall have 75 mm thickness P.U.foam mattress having 40 density with rexine cover and zip. The leg section shall have 110 mm thickness P.U.foam mattress having 40 density with rexine cover and zip for propose alignment with other adjacent sections. The PU foam shall have high resilience and radiolucent property.
16.	HFS-23	Three fold Bed side screen	Approved As per Tender Specifications
17..	HFS-031	OVER BED TABLE	Rejected Not as per Tender Specifications • Overall Size: 1050 mm L x 450 mm W x 970 mm H
18.	HFS-035	GYNAE EXAMINATION TABLE (PLAIN)	Approved As per Tender Specifications
19.	HFS-040	INSTRUMENT CABINET	Approved As per Tender Specifications

Ym

Ch

✓

✓

✓

✓

h

LM Agencies

HFS-002	FOWLER BED	<p>Rejected Not As per tender Specifications</p> <ul style="list-style-type: none"> • Overall Size 2140mm x 970 mm x 600mm • Bed Platform frame size: 2070 mm L x 910 mm W (+ 2%). • Mattress Platform size: 1960 mm L x 800 mm W (+ 2%). • Back rest Only section : 705 mm L x 790mm W. • Size of Knee Rest: Over all -280 mm L x 790 mm W. Only Section : 335 mm L x 800 mm W. • Size of Leg Rest: Over all - 600 mm L x 790 mm W . • The Back Rest size is 760 mm L x 850mm W x 100 mm T, Fix section size is 170 mm L x 850mm W x 100 mm T , Knee Rest size is 390 mm L x 850 mm W x 100 mm T and Leg Rest size is 610 mm L x 850 mm W x 100 mm T
2	HFS-003 SEMI FOWLER BED	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Buffer to Buffer: (2050 (range 2097-2182) mm L x 960 mm W • Bed Frame size:1990 mm L x 910 mm W. • Back Rest Only Section: 600 mm L x 790 mm W. • Size Body Rest: Only Section: 1110mm L x 790 mm W.
3.	HFS-004 PLAIN HOSPITAL BED	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Overall size (Approx): 2040 mm (L) x 920mm (W) x 580 mm. (H)) (+ 2%) (Without Mattress & From Ground to Top Frame). • No hooks • No Matress Arrestor
4.	HFS-006 Bed Side Locker	Approved as per Tender Specification
5.	HFS-007 Adjustable Bed Side Table	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Height adjustment from 820 mm to 1135mm
6.	HFS-008 INSTRUMENT TROLLEY	Approved As per Tender Specification
7.	HFS-009 CRASH CART	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • 12 Nos. hand out bins • 570 mm long and bent at top to have an arm of 450 mm approx. • NIL Round Rubber Buffer .
8.	HFS-010 DRESSING TROLLEY	Approved As per Tender Specification
9.	HFS-011 STRETCHER TROLLEY	Approved As per Tender Specification
10.	HFS-013 EXAMINATION COUCH	Approved as per Tender Specification
11.	HFS-014 SOILD LINEN TROLLEY	Approved as per Tender Specification
12.	HFS-015 WASH BASIN STAND TWO TIER	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Three legs plastic base mounted
13.	HFS-016 KICK BUCKET	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • SS tubular framework fitted with three 50 mm dia non-rusting castor wheels, all without brake
16.	HFS-20 DOUBLE STEP STOOL	Approved as per Tender Specification
18.	HFS-022 WHEEL CHAIR	Approved as per Tender Specification
24.	HFS-031 OVER BED TABLE	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Overall Size: 1120 mm L x 450 mm W x 1030 mm H.
27.	HFS-035 GYNAE EXAMINATION TABLE (PLAIN)	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Fixed rexine upholstered top 50 mm thick • Angle 40x30
30.	HFS-038 WHEEL CHAIR FOLDING	<p>Rejected Not as per Tender Specification</p> <ul style="list-style-type: none"> • Plastic • This hoop is made of aluminium • leatherite cushion material Not available

Vu

over 2 d k y

W

- Plastic strap for calf rest
- **Back is provided with pocket for keeping documents and tools - Not available**

Carevel Medical Systems

S. No	Code	Name of Item	Compliance /Deviation
1	HFS-002	FOWLER BED	Rejected Not as per Tender Specification <ul style="list-style-type: none"> • Bed Platform frame size: 2260 (Upper Limit 2182.8 mm) mm L x 1010 mm W ($\pm 2\%$). • Bed Platform frame size: 2200 mm L x 910 mm W ($\pm 2\%$). • Mattress Platform size: 2100 mm L x 800 mm W ($\pm 2\%$). • Mattress Provided with four section 4" thick PU Foam of 40 density covered PVC Rexene cover. The Back Rest size is 800 mm L x 895mm W x 100 mm T, Fix section size is 150 mm L x 850mm W x 100 mm T, Knee Rest size is 460 mm L x 850 mm W x 100 mm T and Leg Rest size is 470 mm L x 850 mm W x 100 mm T.
2	HFS-003	SEMI FOWLER BED	Rejected Not as per Tender Specification <ul style="list-style-type: none"> • Buffer to Buffer: (2060 (lower limit 2097.2mm) mm L x 1980mm W) (+2% variation) X 600mm H (Without Mattress).
3.	HFS-004	PLAIN HOSPITAL BED	Approved as per Tender Specification
5.	HFS-007	Adjustable Bed Side Table	Rejected Not as per Tender Specification Height adjustment from 790 mm to 1050mm
6.	HFS-008	INSTRUMENT TROLLEY	Approved as per Tender Specification
7.	HFS-009	CRASH CART	Rejected Not as per Tender Specification <ul style="list-style-type: none"> • Not available • The size of top shelf is 670 mm L x 330 mm W made from SS 304 0.9 mm (20 G) sheet.
8.	HFS-010	DRESSING TROLLEY	Approved as per Tender Specification <ul style="list-style-type: none"> • The size of bucket is 230 mm H x 242 mm Outer diameter and size of bowl is 340 mm OD x 85 mm H
9.	HFS-011	STRETCHER TROLLEY	Rejected Not as per Tender Specification <ul style="list-style-type: none"> • One additional flat supports made from MS flat size 30 mm x 5 mm should
10.	HFS-013	EXAMINATION COUCH	Accepted as per Tender Specification <ul style="list-style-type: none"> • Overall size: 1900 mm L x 930 mm W x 840 mm H
17.	HFS-021	BIRTHING BED (Non Motorized)	Rejected Not as per Tender Specification <ul style="list-style-type: none"> • 835 mm H (without mattress). • (2 hooks)
27.	HFS-035	GYNAE EXAMINATION TABLE (PLAIN)	Accepted as per Tender Specification

S K Scientific and Surgicals

S. No	Code	Name of Item	Compliance /Deviation
1	HFS-002	FOWLER BED	Rejected as per tendered specifications <ul style="list-style-type: none"> • Overall Size:- (2140 mm L x 910 mm W X 570 mm H) • Size of Back Rest: Over all-785 mm L x 800mm W. Only section : 450 mm L x 800 mm W. • Size of Knee Rest: Over all - 470 mm L x 800mm W. Only Section : 380 mm L x 800 mm W. • Size of Leg Rest: Over all- 590 mm L x 800mm W . Only Section : 610 mm L x 800mm W. • Size of Fix Section: Over all - 140 mm L x 800mm W. Only Section : 610 mm L x 800mm W. • The bed has easily detachable Polymer moulded head & foot panels. These Head and Foot Pannels are mounted in round bracket size 25 mm OD x 48 mm ID made from MS sheet having 2.0 mm (18 G) thick and welded with

Ku

San 22 2024

San

			bed frame and used along with PVC sleeve/or fitted with side heavy duty clamps. Four corner rubber buffers of 100 MM dia for with castor application
2	HFS-003	SEMI FOWLER BED	<ul style="list-style-type: none"> There are Two locations on the bed platform to hold stainless steel Telescopic Saline tube 12mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel outer covering tube with a knob to mount syringe pump. With three years warranty Mattress Arrestor Set Not Provided <p>Rejected as per tendered specifications</p> <ul style="list-style-type: none"> Buffer to Buffer (2190 mm L x 945 mm W) (+2% variation) X 575mm H (Without Mattress). Bed Frame size: 2140 mm L x 910 mm W. These head and foot panels are mounted in round brackets size 25mm OD x 48mm ID made from MS sheet having 2.0mm There are Two locations on the bed platform to hold stainless steel Telescopic Saline Tube 12mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel outer covering tube with a knob to mount syringe pump.
3.	HFS-004	PLAIN HOSPITAL BED	<p>Rejected as per tendered specifications</p> <ul style="list-style-type: none"> Specifications Overall size (Approx): 2090mm (L) x 910mm (W) x 525 mm. (H) (+ 2%) (Without Mattress & From Ground to Top Frame). Provision of Mosquito Net poles. for 600 mm height: Not provided SS 304 IV Pol made from 12 mm SS Tube with two hooks Telescopic Saline rod 12 mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel outer covering tube with a knob. Mattress Arrestor NOT AVAILABLE
4.	HFS-006	Bed Side Locker	<p>Rejected Not as per tendered specifications</p> <ul style="list-style-type: none"> One drawer 75mm H x 320mm W x 320mm D fitted is provided below the top. Under the drawer is an open storage space and below it is a closed-door cabinet
5.	HFS-008	INSTRUMENT TROLLEY	<p>Rejected as per tendered specifications</p> <ul style="list-style-type: none"> S. S tube 15.87 mm x 1.2 mm (18 G) on all four sides to support two stainless steel shelves made from 0.9 mm (20 G) SS 304 sheet: Supports not available The distance between the two shelves is 590 mm and distance from ground to bottom shelves 250 mm
6.	HFS-009	CRASH CART	<p>Rejected as per tendered specifications</p> <ul style="list-style-type: none"> The three drawers each to hold emergency medicines, ambu. bags, IV solutions, catheters, etc. separately and methodically in boxes of sizes approx - 85 mm W x 85 mm L x 50 mm H - 6 nos., - 85 mm W x 175 mm L x 50 mm H - 3 nos. and 85 mm W x 255 mm L x 50 mm H - 4 nos. : Not Available Stainless steel saline tube made of 12 mm dia. 304 grade S.S. approx. 710 mm long and bent at top to have an arm of 310 mm approx. at the end of which of 6 mm dia. S.S. hook shall be welded with TIG process. The size of top shelf is 670 mm L x 235 mm W made from SS 304 0.9 mm (20 G) sheet. Two side support size 200 mm W and 25 mm made from SS 304 2.0 mm (14 G) sheet provided to top shelf.
7.	HFS-010	DRESSING TROLLEY	<p>Rejected Not as per tendered specifications</p> <p>The size of bucket is 230 mm H x 205 mm</p>
8.	HFS-013	EXAMINATION COUCH	<p>Rejected as per tendered specifications</p> <ul style="list-style-type: none"> Couch fitted with stainless steel Legs made from 2.0 mm (14 G) SS 304 sheet with nylon bush at bottom of the leg & powder coated. : Not Available Upper section of box approx size 1220 mm L x 460 mm W x 630 mm H with three sliding drawers of approx size 335 mm L x 410 mm W x 55 mm H.
22.	HFS-026	HOSPITAL LOCKER PIGEON HOLE	<p>Approved as per tendered specifications</p>

Decision: After going through the report of Sub Committee and sample evaluation report submitted by the Technical Experts from Machinery Division, Jammu and Directorate of Health services, Jammu, technical Evaluation/Advisory Committee accepted the report of subcommittee and sample evaluation report submitted by the Technical Experts and It was unanimously recommended to open the Financial Bids of the bidders declared qualified by the Technical

Yr

or

SK

6 7

h

Experts as per tendered specifications. However, items where less than three bidders are qualified, be put to tenders afresh.

The meeting was attended by following Officers/Official:

S No.	Name of Officer/Official	Designation	Signature	Contact No.
1.	Mohd. Ashraf			
2.	VIKAS PANDITRA	A.E.G. Machind	Vikas Panditra	9419130648
3.	Chandee Kuan	ALH, JKMSCL	Chandee Kuan	9149703614
4.	Dr. Digi Jammu	DE of ETC JKMSCL	Dr. Digi	
5.	Shashinder Sharma	H.E.O., JKMSCL	Shashinder	9419134739
6.	Karan Singh	AAU		
7.	Dr. Ujjala Vagha	MO CC	Ujjala	
8.	A Komel Sharma	FM, DIL Jammu	Komel	9149641205
9.				
10.				


General Manager (Adm.)
JKMSCL.

No: JKMSCL/GM(Adm.)/623/2025/ 868-74

Dated 01/05/2025

Copy for information to the:-

1. Secretary to Govt. Health & Medical Education Deptt. J&K
2. Principal, Govt. Medical College, Jammu/ Srinagar.
3. Managing Director - J&K Medical Supplies Corporation Limited.
4. Controller, Drug & Food Control Organization, J&K.
5. Director, Health Services, Kashmir /Jammu.
6. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
7. General Manager (K), J&K Medical Supplies Corporation Limited.
8. Office file.