



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmssljepm@gmail.com; website: www.jkmssl.nic.in

M/s CHSPL Common Wealth System,
New Delhi.(Bidder)

No: JKMSCL/GM (Adm.)/2025/286-300

Dated: 12-04-2025.

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of Rate Contract for Washer Dis-infectant

Ref No.

1. NIT No: NIT/JKMSCL/Mach/2023/575 Dated 09.02.2023

2. Minutes of Purchase Committee vide No. JKMSCL/MD/2024-25/895-907 dated 05.02.2025.

3. LOI No. JKMSCL/GM/225/15352-57 DATED 10.03.2025

Sir,

As per the decision of the Purchase Committee, the Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour on the quoted rates by you in the NIT:

S. No.	Name of Item	Unit	Rates inclusive of Custom Duty, Custom Clearance and Taxes
1.	Washer Dis-infectant M/s Steris India Pvt. Ltd., Mumbai. (wholly on subsidiary of Steris corporation USA) (Importer) M/s Steris Corporation, USA.(Manufacturer) Model No.: AMSCO 2532 Washer Disinfectant.	US\$	US\$ 47,835.45 (US\$ Forty Seven thousand eight hundred thirty five and forty five only)

The approved rates for Comprehensive Maintenance Contract after the expiry of warranty period of 05 (Five) years are as follows:

S. No	Name of Item	CMC for five years
1	Washer Dis-infectant M/s Steris India Pvt. Ltd., Mumbai. (wholly on subsidiary of Steris corporation USA) (Importer) M/s Steris Corporation, USA.(Manufacturer) Model No.: AMSCO 2532 Washer Disinfectant.	6 th yr 248068.00 7 th yr 292721.00 8 th yr 307357.00 9 th yr 322725.00 10 th yr 338861.00
		Total: 1509732.00



Terms and Conditions:

- Delivery Site :** Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
- Validity :** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
- Taxes:** As mentioned above and as applicable.
- Warranty:** Five Years from the date of installation.
- Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
- Supplies** Within 90 days from the date of issuance of supply order.

7. **Penalty** In case In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%.
8. **Installation:** Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
9. **Availability of Spares:** Firm shall certify that the parts of the equipment shall remain available at least for a period of **five years** after the year of expiry of warranty period of five years.
10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
11. The supplier/manufacturer/importer shall ensure to supply the Brand new and unused item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.
13. **Fall Clause:** The prices under contract shall be subject to price fall clause. The prices charged for the goods supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically be reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
14. **Risk Purchase :** If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.
The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.
15. **Orders shall only be placed on the basis of actual requirement and funds as available under the relevant Head of Accounts.**
16. **Payment shall be released after successful installation and commissioning of the equipment duly verified by the intending department.**
17. **Approved technical specifications shall remain same as per NIT (Annexure I)**

All the terms and conditions of the NIT, Standard Procurement of Procedure (SPP) of JKMSCL & LOI shall be part of this Rate Contract.

Issued with the Decision and Authorization of Purchase Committee of JKMSCL vide its minutes No. JKMSCL/MD/2024-25/895-907 dated 05.02.2025.


Yours sincerely,
(Mohd. Ashraf Choudhary) JKAS
General Manager (Adm.), JKMSCL


Copy for information to the:

1. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
2. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
3. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
4. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
5. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
6. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
7. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
9. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
10. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
11. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
12. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
13. Asstt Programmer to upload on official website.
14. IT cell JKMSCL.
15. File

Washer Dis-infectant

Washer Disinfectant (Double Door) Capacity 264 Ltr.	Exterior Dimension 724 x 1981 x 724 mm	Qty 01
---	---	--------

- Washer Disinfectant with touch screen control is intended for use in the cleaning and intermediate-level disinfection of soiled reusable utensils, trays, glassware, rubber and plastic goods, simple hard-surfaced rigid surgical instruments (such as forceps and clamps) Imported and other similar items.
- The washer disinfectant shall be suitable for cleaning and disinfection of surgical Instruments/goods. The process shall include prewash, detergent wash and hot water disinfection, rinse, and drying cycles.
- It should be microprocessor Control system to ensure correct program sequence and display irregularities or deviations immediately.
- Washer Disinfectant should have four factory pre-programmed and validated cycles (Instruments. Utensils Auto Sink Instruments and Auto Sink Utensils), Nine additional washing cycles may be programmed to meet extra Customer requirements.
- The microprocessor control system features
 - I Equipped with audible warning system.
 - II Locking program cycle parameters with access code. Security lock-out feature that enables programs and temperatures to be locked and unchangeable without the proper access code.
 - III Cycle data is stored as a protection against power disruption.
- Washer disinfectant is equipped with a touch screen Customer interface. This 5.7" (115 mm) color touch screen is mounted at eye level
- Washer disinfectant should have Icon based touch screen control
- The unit should also have an interface standard for a batch printer. The unit should have also had USB port for cycle data.
- Chamber Capacity Chamber Volume should be approx. at least 250 Lts to hold at least 8 Nos of standard DIN trays.
- Washer Disinfectant should be clean 8 Din Tray in 30 minute and 16 Din tray 60 Min. It should also have fastest cleaning cycle of 15 Min without Disinfection
- Rack is equipped with rotary spray arms positioned to reach all load surfaces. Removable end-tips ensure easy spray arm cleaning and maintenance, and also have lumen flushing port facility to clean 6 lumen per cycle.
- The chamber should be constructed of argon-welded 20-gauge, #304 L. stainless steel (No. 4 finish) Chamber inhibits corrosive action of detergent. The chamber edges should not have pockets & folds, to avoid bacterial growth. The wash chamber should also be fitted with bright light for clear visibility of the washing process Chamber dimension should suit the capacity
- Washer Disinfectant door should be Manual drop-down doors are constructed of double pane tempered glass to allow operator to view chamber interior with doors closed Doors remain cool to touch while cycle is in progress. Each door is mounted on a compressed seal that reduces heat loss and increases heating capability
- Washer Disinfectant Pump should be Stainless-steel pump provides circulation using a 3 HP (2.2 kW) motor with up to 22 psig (151 kPa) at pump outlet.
- Washer Disinfectant Pump, spray system and all recirculating hard piping are of stainless-steel construction
- Washer Disinfectant should have three Chemical injection Peristaltic pumps
- Removable front service panels provide easy access to all piping, valves, electrical components, wiring Servicing from sides of unit is not required

- Washer Disinfector is equipped with safety lockout feature so program cannot start unless door is fully closed. Door is mechanically interlocked during cycle and cannot be opened until cycle completion
- The design should focus on saving the environment through reduced consumption of all utilities. The water consumption should not be more than 21 Gallon
- **Standards & Norms:**
- Should be US-FDA/European CE certified.
- Manufacturer should EN ISO15883 Compliant
- Medical Devices Directive 93/42/EEC as amended by 2007/47/EC