JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmscljepm@gmail.com **website** www.jkmscl.nic.in

JKMSCL

M/S S.R. Technomed,
Plot No. 22B, Bakshi Nagar,
Jammu-180001, e-mail: srtechnomed@gmail.com

No: JKMSCL/2019/GM|Adm|1553-66 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref No. 1. NIT/JKMSCL /Mach/2018/300 dated 23.10.2018

2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
E1304 High Power Photocoagulator Make : M/S Carl Zeiss India	INR	**4051278.00**	SGST: **243076.68** CGST: **243076.68**	1st yr: 81025.56 2nd yr: 81025.56 3rd yr: 121538.34 4th yr : 121538.34 5th yr :121538.34 **Total=526666.14**	Quoted Rate(1+2) **4537431.36** CMC (3): **=526666.14** Total (1+2+3) **=5064097.50**

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-

(a) Delay up to one- fourth period of the prescribed delivery period **2.5%**
(b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period **5%**
(c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period **7.5%**

Scanned by CamScanner

(d) Delay exceeding three- fourth of the prescribed period **-10%**

(e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.

12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.

Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

Priyanka Bhat (KAS)
General Manager-Adm.
JKMSCL.

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

HIGH POWERED PHOTOCOAGULATOR

Laser System	Diode-pumped solid-state	
Wavelength	532 nm	
Pulse Duration	Pulse Duration .01-3.0 seconds	
Pulse rate	.05-1.0 seconds and single pulse	
Energy Settings	50 mW to 2500 mW out of an endo-ocular probe	
Spot Size	50-1000 microns with Laser Link Z-1000, 50-500 microns with Converging Optics LaserLink	
Aiming Beam	Contrasting color, adjustable intensity 635 nm nominal, <1.0 mW	
Cooling System	Air-cooled	
Power Requirements	110 VAC,<6A,50/60 Hz, Single phase,220 VAC,<3A,50/60 Hz, Single phase	
Dimensions (Console)	Single Port 15.2cm (6in.) x 25.4cm (10in.) x 35.6cm (14in.) Dual Port 15.2cm (6in.) x 25.4cm (10in.) x 37.5cm (14.5in.)	
Weight	8 kg or 17.5 lbs	
Delivery Systems	LaserLink Z-1000 Slit Lamp Adapter (for Zeiss SL 130,30SL,20SL and 125SL); Converging Optics laser Link, Keeler & heine LIO Comprehensive selection of standard, straight, curved, aspirating, illuminating and stepped endophotocoagulation probes in 20,23 and 25 gauge sizes, among others	**Slit Lamp:** Magnification: -5x, 18x, 12x, 32x, with 10x eye pieces -6x, 10x, 16x, 25x, 40 x with 12.5x eye piece Field of view diameter = 46-6mm with 10x eye piece 31-5mm with 12.5x eye piece. Eye piece magnification = optionally 10x or 12.5x, super high point eye pieces. Compensation of ametropia = 8D. Width of slit image = continuous from 0-14mm, display 1 2 5 10mm. Length of slit Image= insteps 0.3 2.5 3.5 7 10 14; triple slit. Decentration of slit = ± 4° horizontally, click stop at 0° Swivel range of slit projection = 180°, scale for angular difference. Click stop at -10°, +10° Angle of incidence = 0 -20 tillable Filters: Blue, green, grey, swing-in, heat-reflecting filter, permanently integrated, differing screen, swing-in. Projection Illumation: 6v 20w halogen lamp. Rated voltage: 100-240v ± 10%, self-sensing, 50/60 Hz.
Available Accessories	Standard ,smart and Power Ease foot switches; Physician eye safety filters; Remote control; Comprehensive selection of standard, straight, curved , aspirating, illuminating and stepped endophotocoagulation probes in 20,23 and 25 gauge sizes, among others	

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmscljepm@gmail.com website www.jkmscl.nic.in

JKMSCL

M/S Hansraj Nayyar (Mediquip) Inc
Hans Towers, Hall Bazaar
Amritsar 143006 (India)
hansnayyar@yahoo.com

No: JKMSCL/2019/ Gm/Adm/1623-36 Dated:28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref.No. 1. NIT/JKMSCL/Mach/2018/300 dated 23.10.2018

2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
Vision Drum Make : HRN Model : HRN-129	INR	**4500.00**	IGST: 540.00	0.00	Quoted Rate(1+2) = **5040.00**
Medical Autoclaves Sterilizer Chamber Diameter 400 Chamber Depth 600 Model : 16HA/E/Manual Make : NAT STEEL	INR	**443620.00**	IGST 79852.00	1st yr: 20000.00 2nd yr:20000.00 3rd yr:20000.00 4th yr :20000.00 5th yr :20000.00 **Total=100000.00**	Quoted Rate(1+2) = **523472.00** CMC (3): **=100000.00** Total (1+2+3) **=623472.00**
Medical Autoclaves Sterilizer Chamber Diameter 500 Chamber Depth 900 Model : 20HA/E/Manual Make : NAT STEEL	INR	**612000.00**	IGST 110160.00	1st yr: 20000.00 2nd yr:20000.00 3rd yr:20000.00 4th yr :20000.00 5th yr :20000.00 **Total=100000.00**	Quoted Rate(1+2) = **722160.00** CMC (3): **=100000.00** Total (1+2+3) **=822160.00**
Medical Autoclaves Sterilizer Chamber Diameter 500 Chamber Depth 1200 Model : 20HB/E/Manual Make : NAT STEEL	INR	**683200.00**	IGST 122976.00	1st yr: 20000.00 2nd yr:20000.00 3rd yr:20000.00 4th yr :20000.00 5th yr :20000.00 **Total=100000.00**	Quoted Rate(1+2) = **806176.00** CMC (3): **=100000.00** Total (1+2+3) **=906176.00**

EB 020

EB 008 A

B

C

Scanned by CamScanner

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-
   (a) Delay up to one- fourth period of the prescribed delivery period **2.5%**
   (b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period - **5%**
   (c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period - **7.5%**
   (d) Delay exceeding three- fourth of the prescribed period **-10%**
   (e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.
10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.
14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.

Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely.

Priyanka Bhat (KAS)
General Manager-Adm,
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).

Scanned by CamScanner

MEDICAL AUTOCLAVES STERILIZER (HORIZONTAL)

Chamber Diameter (mm)	Chamber Depth (mm)	Chamber Volume (Ltr.)	Steam Generator Capacity (Ltr/Kw)	Floor Space (WidthxDepth)
400	600	81	45/18	990/1800
500	900	185	55/36	1200/2300
500	1200	248	55/36	1200/2700

Quality certification – US FDA/CE/BIS.

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmscljepm@gmail.com, website www.jkmscl.nic.in

JKMSCL

M/S L.M Agencies.
Auqaf Building
Ram Munish Baugh Srinagar,
Kashmir -190001,E-mail: kashirawaz@gmail.com, kashirawaz@yahoo.com

No: JKMSCL/2019/Gm|Adm|1567-80 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref No. 1. NIT/JKMSCL /Mach/2018/296 dated 14-07-2018
2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
EYE BANK SPECULAR MICROSCOPE Make : KONAM Model : KONAM CELL CHEK-D	INR	2464999.00	SGST: 147899.94 CGST: 147899.94	1st yr: 61769.00 2nd yr:72887.00 3rd yr:76532.00 4th yr :80358.00 5th yr :84376.00 Total :375922.00	Quoted Rate (1+2)= 2760798.88 CMC (3): 375922.00 Total (1+2+3)= 3136720.88

EBool

Terms and Conditions:

1. **Delivery Site :** Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity :** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes:** As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.

Scanned by CamScanner

7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-

(a) Delay up to one- fourth period of the prescribed delivery period **2.5%**
(b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period **5%**
(c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period **7.5%**
(d) Delay exceeding three- fourth of the prescribed period **-10%**
(e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.

12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.
Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

Priyanka Bhat (KAS)
General Manager-Admn
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

EYE BANK SPECULAR MICROSCOPE
Ultra wide viewing field areas of 1000 X 750 um, Multi- Sample Analytic Areas selectable upto 4 frames each of 400 X 300um, Camera of CMOS Sensor with dual capabilities for full graft imaging with 2X Zoom option, Illumination of Halogen Lamp, Reliable Cellular Analysis method, Centre Method , Flex Centre Method, Built in Pachymeter, Real Time Media Temperature Sensor, Moving range of Stage X : 16mm, y :16mm, Z : 16mm with a Tilt angle of 15 degree, with chamber Cap, Chamber Adaptors, Vial Adaptor Set (Vial Lid, Vial Adaptor, Spacer, Holder), Micrometer, User Friendly Software with an integrated Database, Quoted with suitable computer and printer, USFDA and European CE: approved. (Offered equipments accessories and consumables should be of same offered Original Equipment Manufacturer.)

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No 58 Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmscljepm@gmail.com website www.jkmscl.nic.in

JKMSCL

M/S L.M Agencies.
Auqaf Building
Ram Munish Baugh Srinagar,
Kashmir -190001,E-mail: kashirawaz@gmail.com, kashirawaz@yahoo.com

No: JKMSCL/2019/GM/Adm/1581-94 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref No. 1. NIT/JKMSCL/Mach/2017/288 dated 14.04.2018
2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
Phacoemulsification Machine **Model : ALCON System** **Make : Laureate System**	INR	**2369999.00**	**SGST: 142199.94** **CGST: 142199.94**	1st yr: 75000.00 2nd yr: 75000.00 3rd yr: 75000.00 4th yr: 75000.00 5th yr :75000.00 **Total : 375000.00**	**Quoted Rate (1+2)= 2654398.88** **CMC (3): 375000.00** **Total (1+2+3)= 3029398.88**

MC 1801

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated

Scanned by CamScanner

damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-

(a) Delay up to one- fourth period of the prescribed delivery period – **2.5%**
(b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period **5%**
(c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period – **7.5%**
(d) Delay exceeding three- fourth of the prescribed period **-10%**
(e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.

12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.
Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

28/2/18

Priyanka Bhat (KAS)
General Manager-Adm,
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager – K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

Phaco Emulsification Machine:	
1.	4 to 6 piezoelectric crystals U/S titanium hand piece with frequency of approx.40 kHz
2.	U/S delivery modes: continuous,pulse,brust and programmable emission mode (adjustable duty cycle)
3.	Continuous tip strok length control and adjustment for consistent U/S power delivery
4.	**Range of phaco tips like straight 10°,15°,30°, KELMAN**
5.	**M.I.C.S phaco tips for 1.8 mm/2.2mm/2.8mm incision co- axial phaco.**
6.	Inbuilt Bipolar Diathermy with 2MHz frequency, 7W power.
7.	**Peristaltic aspiration system with vacuum range of 5-500 mmhg and flow rate 2 to 50 cc/mn with adjustable aspiration rise time.**
8.	**Autoclavable 15 No. or Disposable 350 I/A tubing with facility of anterior chamber surge protection system.**
9.	**cutting rate ranging from 600 to 1500 cuts per minute with pneumatically powered guillotine cutter**
10.	Min 12 inches LCD with touch screen and user friendly control software.
11.	Programmable dual linear foot switch for linear/ panel control of all equipment functions.
12.	More than 1000 user programs saving facility.
13.	**European CE/US FDA approved.**
14	Phaco trolley with manually adjustable I/V stand
15.	Suitable UPS
	Accessories required:-
	U/S hand piece
	Co-Axial I/A hand piece
	Bimanual I/A hand piece
	Phaco tips
	Silicone Sleaves
	Test chambers
	I/A Tubing
	Diathermy Forceps
	Diathermy cable
	Tip Wrench
	Sterilization Tray
	Vitrectomy cutter

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmsclepm@gmail.com, **website** www.jkmscl.nic.in

JKMSCL

M/S Macson Traders
Karan Nagar, Srinagar
e.mail : macson_tr@rediffmail.com

No: JKMSCL/2019/Gen/Adm/1595-1608 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref No. 1. NIT/JKMSCL /Mach/2018/300 dated 23.10.2018

2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
Corneal Topography System Make : Optikon (Keratron) Model : REF 161104	INR	**1130000.00**	SGST: **67800.00** CGST: **67800.00**	1st yr: 53336.00 2nd yr:66086.00 3rd yr:81880.00 4th yr :101449.00 5th yr :125695.00 **Total =428446.00**	Quoted Rate(1+2) **1265600.00** CMC (3): **428446.00** Total (1+2+3) **1694046.00**
Vitrectomy Surgery Advances Technology Make : Optikon (R-Evolution) Model : GIUGIARO	INR	**8150000.00**	SGST: **489000.00** CGST: **489000.00**	1st yr: 384680.00 2nd yr:476619.00 3rd yr:590530.00 4th yr :731667.00 5th yr :906535.00 **Total=3090031.00**	Quoted Rate(1+2) **9128000.00** CMC (3): **3090031.00** Total (1+2+3) **12218031.00**
Optical Biometer Make : Topcon Model : Aladdin	INR	**2330000.00**	SGST: **139800.00** CGST: **139800.00**	1st yr: 105000.00 2nd yr:125000.00 3rd yr:125000.00 4th yr :137000.00 5th yr :150000.00 **Total 642000.00**	Quoted Rate(1+2) **2609600.00** CMC (3): **=642000.00** Total (1+2+3) **3251600.00**

EB 003

EB 012

EB 016

Scanned by CamScanner

EB027	**Keratoplasty sets** Make : ISCON Surgical Ltd Model : PRICON	INR	15000.00	SGST: 900.00 CGST: 900.00	00.00	Quoted Rate(1+2) = **16800.00**
EB019	**Cataract Set** Make : ISCON Surgical Ltd. Model : PRICON	INR	7000.00	SGST: 420.00 CGST: 420.00	00.00	Quoted Rate(1+2) = 7840.00

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-
   (a) Delay up to one- fourth period of the prescribed delivery period – **2.5%**
   (b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period – **5%**
   (c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period - **7.5%**
   (d) Delay exceeding three- fourth of the prescribed period **-10%**
   (e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

Scanned by CamScanner

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.
Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

28/3/19

Priyanka Bhat (KAS)
General Manager Adm,
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt, J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

Corneal Topography system

Patented Placido Disk Technology
Triangulation with cone-of-Focus, placid rings, and corneal surface

Placido Disk Technology with Cone –of- Focus Alignment System

Proven, Patented technology

- Patented Cone –of- Focus Alignment System and ARC step algorithm delivery sub-micron elevation accuracy.
- Using triangulation, the advanced Atlas Arc-Step also rithm provides reliable surface reconstruction
- 22-ring Placido disk optimized to avoid ring cross over, which means reliable results for a wide range of patients
- Long, comfortable 70mm working distance minimizes focusing error found in "small cone" systems

Smart Capture Image Analysis

Technology that helps your staff get the right image the first time

- Smart Capture analyzer 15 digit images per second during alignment and automatically selects the highest quality image
- Next-generation image processing provides more repeatable, reliable results, even in difficult cases
- Less dependence on operator technique means greater efficiency and fewer repeat exams
- Assess higher-order corneal aberrations and simulate visual acuity
- Optimize the selection of aspheric IOLs using Zernike corneal wave front analysis

Corneal Wave front Analysis

A valuable tool for patient education and treatment selection

Corneal Wave front Analysis takes corneal topography to a new dimension. Using ray tracing technology, the displays higher –order corneal aberrations, providing valuable insight for patient education and treatment planning.

- Educate patients about higher-order aberrations and simulate visual acuity
- Assess corneal refraction with image simulation and point spread function
- Measure corneal spherical aberration with Zernike analysis to optimize the selection of aspheric IOLs
- View the effects of individual higher order corneal aberrations
- View changes in contrast sensitivity through the modulation transfer function (MTF)
- Scotopic and photopic pupil images
- Automatic Horizontal Visible iris Diameter(HVID) or "white to white' measurement simplifies the selection of contact lenses
- Easily measure and display HVID for both eyes with OD/OS Comparison Display

Automatic Pupillometry and HVID Measurement

Tools that enhance contact lens fitting and refractive surgery planning

- Pupil size, measured at two levels of illumination (scotopic and photopic @ 700nm), provides insight into optical zone under varying light conditions.
- View centration of LASIK or orthokeratology treatment in relation to pupil center to assess treatment effectiveness
- Improve multifocal contact lenses selection by understanding patient's pupil size at two levels of illumination
- Streamline contact lens selection and fitting with automatic horizontal Visible Iris Diameter (HVID) measurement.

Ergonomic Design

Ideally suited for you and your patients

- Compact, integrated system with powerful computer and analysis software
- Unique chinrest positions patient for easy image capture and wide peripheral coverage and automatically detects OD/OS
- Non -visible placid ring illumination is comfortable for even the most light sensitive patients

Scanned by CamScanner

Pathfinder II Corneal Analysis Software

Take the art out of topography analysis

Pathfinder II Corneal Analysis Software for the Corneal Topography System is a comprehensive, easy to understand, and reliable anterior topographic screening tool to assist with refractive surgery screening and to help identify abnormal corneal conditions for monitoring and follow-up. Using next generation analysis technology, Path finder II instantly provides probabilities for 5 different corneal conditions by comparing topography exams to an extensive clinical database.

Show more

MasterFit Contact Lens Software

MasterFit Contact Lens Software helps streamline the fitting of gas permeable (GP) lenses.

Corneal conditions such as asymmetrical astigmatism, a decentered apex, contact lens wrapage, and irregular, peripheral and central astigmatism can dramatically affect the way a contact lens behaves on the eye. MasterFit II allows customized fitting of GP lenses according to your own fitting preferences.

- Simulate fluorescein patterns for custom and stock lenses, including spherical, toric, and aspheric designs
- Automatically design lenses to your preferences by customizing fitting options such as desired tear film clearance
- Improve trial lens fitting efficiency by adjusting lens parameters, such as peripheral curves, and simulating lens movement to compensate for lens-to-cornea relationship
- Email lens design and topography exam to your lab for efficient ordering and fulfilment

Review Software

Greater productivity through connectivity Review Software is a high-value tool which allows you to get more out of your Corneal Topography System. This new software builds upon the pervious tools that clinicians have come to trust, while providing a range of added functionality.

- Dynamic remote access to all your corneal topography exam data and patient education tools, such as corneal wavefront simulation
- Increase instrument availability for improved patient workflow
- Equivalent software functionality

<u>Vitrectomy surgery advanced technology</u>

The constellation vision system delivers an exceptional level of performance through its advanced technologies

- Featured Technologies
- Ultravit high speed vitrectomy probes
- Ultravit high speed vitrectomy probes deliver 7500 cpm dual pneumatic drive technology in 20,23,25+and 27 series. This technology has been designed to optimize your surgical experience and patient outcomes.
- **Duty cycle control**
- The constellation vision system allows the surgeon to modify duty cycle to control flow independent of vacuum and cut rate. The surgeon is given the ability to select 3 different duty cycle options at any given cut rate, port biased open 50/50 or port biased closed.
- **Integrated pressurized infusion**
- The Constellation vision system constantly monitors infusion pressure. The IOP compensation feature provides control of infusion pressure which results in more stable IOP
- **Illumination**
- The constellation vision system delivers state of the art illumination for visualizing tissues. The constellation illuminator incorporates ENGAUGE radio frequency identification device technology (RFID)to recognize light probe gauge size and automatically adjust light intensity.the constellation Xenon illuminator has been designed to provide long-lasting high brightness illumination.
- **Integrated purepoint laser**
- Features includes voice confirmation the multi function foot switch ENGAUGE RFID and dual laser attachment ports

Scanned by CamScanner

OPTICAL BIOMETER

- Should provide highly accurate laser optic measurements for every section of the eye-from the cornea to the retina. With its integrated Olsen formula and the optional Toric Planner featuring the Barrett Toric Calculator, should supplies the user with latest technology in IOL prediction for any patient.
- **Access to full eye data in a single click**
- Accurate IOL prediction is crucial for patient satisfaction in state of the art cataract surgery. Should provide all measurements needed to take full advantage of the latest generation IOL prediction methods for improved refractive outcomes with one click.
- **Thinking out of the box for a new level of performance**
- RBF stands for Radial Basis Function, which is like a neural Network, specialised in feature extraction and feature extraction and feature recognition and can handle multiple factors and non-linear relationships. Based on the input parameters axial length (AL), anterior chamber depth (ACD) and corneal curvature (K) it finds the right pattern that lead to an accurate IOL prediction.It is combined with a boundary model, this means that the RBF calculator only provides a result if the respective prediction is accurate with a very high probability. First clinical trials prove that the RBF performs excellent and may even outperform the current most advanced theoretical formulae.
- **Improved refractive outcomes**
- Learn more from Dr. Sheridan lam about how measured lens thickness allows for significant improvement of the IOL prediction accuracy leading to improved outcomes.
- **Excellent IOL prediction in all eyes**
- Latest generation IOL calculation methods such as Olsen and Barrett, implemented in EyeSuite IOL – the IOL planning suite of the lenstar – allow for excellent IOL prediction in short, average and long eyes with standard spherical or toric IOLs. Furthermore , Eyesuite
- **Improved outcomes with dual zone keratometry**
- Lenstar 's unique dual zone keratometry provides measurement of the axis and astigmatism, equivalent to the "Gold Standard " manual keratometry recommended for toric IOL by manufacturers.
- The closely spaced 32 measurement point pattern mpr·ves precision, both delivering more data and minimizing the need for software data interpolation.
- **Topography for torics- match the axis**
- With the optional T-Cone toric platform, the axis and astigmatism measurement of lenstar is extended with true 11-ring Placido topography. This additional data improves the efficacy and safety of toric IOL surgery, eliminating the risk of irregularities and allowing the user to double check the axis location.

KERATOPLASTY

1) Barraquer Wire Speculum, Solid Blades, Large

2) Hoffer Optic Zone Marker Size 3.0mm Dia

3) Osher-Neumann Radial Marker, 4 Blades

4) Osher-Neumann Radial Marker, 8 Blades

5) Grene Visual Axis Marker

6) Deitz Incision Depth Gauge, 0.50 mm

7) Thornton Globe Fixation Ring

8) Bores Incision Spreading Forceps

9) Bores Corneal Fixation Forceps, Straight

10) Kramer Corneal Fixation Forceps, Curved

11) Rycroft Air Injection Cannula, 27G

VY

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office Corporate Head Office Plot No 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House ,Bemina(190018),Srinagar

email: jkmscljepm@gmail.com, **website** www.jkmscl.nic.in

JKMSCL

M/S Winner Pharmaceuticals
Shop No. 272,Sector 11,
Nanak Nager, jammu
Email: winnerpharmaceutical@gmail.com

No: JKMSCL/2019/GM/Adm/1609-22 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" for the establishment of Eye Bank - issuance of rate contract thereof.

Ref No. 1. NIT/JKMSCL /Mach/2018/300 dated 23.10.2018
2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates with Taxes & conversion (1)	CMC Rates for 05 yrs (in Rs.) (2)	Total Amount (in Rs.) With taxes and CMC (1+2)
Ophthalmic chair unit Mfg : Warden Surgical Co. Pvt. Ltd	INR	**169495.00** **Accessories:** **7000.00**	1st yr: 5000.00 2nd yr: 5000.00 3rd yr:5000.00 4th yr 5000.00 5th yr :5000.00 **Total** 25000.00	Quoted Rate with tax = **176495.00** CMC (2) **=25000.00** Total (1+2) **201495.00**

EB 024

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder has failed to supply :-
   (a) Delay up to one- fourth period of the prescribed delivery period – **2.5%**

Scanned by CamScanner

(b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period - **5%**
(c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period - **7.5%**
(d) Delay exceeding three- fourth of the prescribed period **-10%**
(e) Fraction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.
Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

Priyanka Bhat (KAS)
General Manager-Adm.
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

Ophthalmic chair unit

Reclining (Full)
YES

Rotation

76°

Foldable armrests

YES

Horizontal Seat Displacement

YES

Seat: Max./ Min. Height in mm

490-660

Footrest

YES

Base Plate

YES

Base Plate with wheels

YES

Maximum Load

175 Kg.

Max. Power consumption

700W

Power

230 V AC/50-60Hz

Ophthalmic **Table top**

2 Instruments

Maximum Load on Table top

40 Kg.

Touch Screen Control Panel

Yes

Scanned by CamScanner

Power on Table top

0-12V/ 5A or 15V/5A

Power Near Vision Lamp

3W LED

Power of Room illumination

0-240V/ Max. 100W

Safety Top

Yes

Stroke elevated Table top

790-920 mm

Dimensions Trial Lens Drawer

560(w) x 483(d) x 590(h) mm

Inside dimension Trial Lens Drawer

460(w) x 350(d) x 70(h) mm

Dimension PC-Desk

1400(w) x 600(d) x 740(h) mm

Dimension PC-Desk

1800(w) x 845/600(d) x 740(h) mm

Total Power Supply

220-240V/ 50-60Hz

Maximum Power Consumption

1,000 W

Electronic Soft Start System

Yes

US FDA/European – CA Approved

Scanned by CamScanner

JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)
Corporate Head Office Corporate Head Office Plot No 58 Friends Colony Satyam Road Trikuta Nagar Jammu
Corporate Office: Corporate Office Kashmir Near Haj House ,Bemina(190018),Srinagar
email: jkmsclepm@gmail.com website www.jkmscl.nic.in

JKMSCL

M/S S.R. Technomed,
Plot No. 22B, Bakshi Nagar,
Jammu-180001, e-mail: srtechnomed@gmail.com

No: JKMSCL/2019/GM/Adm/1637-50 Dated: 28-02-2019

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of Rate Contract thereof.

Ref No. 1. NIT/JKMSCL NIT/JKMSCL/Mach/2018/298 dated 14-08-2019
2. Minutes of State Level Purchase Committee vide No. JKMSCL/PS/MD/2019/5655-68 Dated 11.02.2019.

Dear Sir,

The Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour, as per the rates mentioned against each. The rate contract shall remain valid for a period of two year (24 Months) from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder

Name of item under Group : Machinery & Equipment	Currency	Basic Rates without Tax with conversion (1)	Taxes (2)	CMC Rates for 05 yrs (in Rs.) (3)	Total Amount (in Rs.) With taxes and CMC (1+2+3)
300 mA X-ray Make : ADONIS Medical System Model : AE300 HF RX	INR	960000.00	SGST: 57600.00 CGST: 57600.00	1st yr: 5000.00 2nd yr:5000.00 3rd yr:5000.00 4th yr :5000.00 5th yr :5000.00 **Total =25000.00**	Quoted Rate(1+2) = **1075200.00** CMC (3): **=25000.00** Total (1+2+3) **=1100200.00**

MC 1789

Terms and Conditions:

1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
2. **Validity** : The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes**: As mentioned above.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at such rates, as given below, of value of stores which the bidder

Scanned by CamScanner

has failed to supply :-

(a) Delay up to one- fourth period of the prescribed delivery period – **2.5%**

(b) Delay exceeding one fourth but not exceeding half of the prescribed delivery period **5%**

(c) Delay exceeding half but not exceeding three- fourth of the prescribed delivery period – **7.5%**

(d) Delay exceeding three- fourth of the prescribed period **-10%**

(e) F raction of a day in reckoning the period of delay in supplies shall be eliminated if it is less than half a day. The maximum amount of agreed liquidated damage shall be 10%.

8. **Installation**: Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.

9. **Availability of Spares**: Firm shall certify that the parts of the equipment shall remain available atleast for a period **of five years** after the year of expiry of warranty period of five years.

10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.

11. The supplier/manufacturer/importer shall ensure to supply the item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.

12. JKMSCL is at liberty to get the feedback from the end-user department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.

13. **Comprehensive Maintenance Agreement**: You are further required to execute Comprehensive maintenance agreement before the release of payment.

14. **Orders shall only be placed on the basis of actual requirement and funds are available under the relevant Head of Accounts.**

All the terms and conditions of the NIT, terms mentioned in the LOI and Standard Procurement of Procedure (SPP) shall be part of the approved Rate Contract.

Issued with the approval of Managing Director, JKMSCL/Tender Inviting Authority.

Yours sincerely,

25/7/19

(Priyanka Bhat (KAS))
General Manager-Adm.
JKMSCL

Copy for information to the:

1. Principal Secretary to Govt., Health & Medical Education Deptt., J&K(Director, Board of Directors-JKMSCL).
2. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
3. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
4. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
5. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
6. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
7. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
9. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
10. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
11. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
12. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
13. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
14. Office file.

Scanned by CamScanner

Annexure I to Rate Contract No._____ dated _____

MC1789	**300 mA HF X-Ray Machine**
Clinical purpose	Radiography of the bones and fractures and other arthropathies. X-Ray Chest for the supportive diagnosis of the Pulmonary Tuberculosis X-Ray Pelvis (KUB) for renal disorders and stones. Sinusitis, Fractures of the Skull Cardiac diseases and cardiac enlargement Silicosis and other respiratory conditions, like Pleual effusion, hydrothorax, Pneumothorax Peritonitis by X-Ray abdomen.
Technical Characteristics	High Frequency X-Ray machine suitable for general Radiography. **X-Ray Generator** • High Frequency X-Ray generator having Frequency of 40 KHz more suitable for Radiography should be provided. • Power output of generator should be 25 KW or more. • Radiography KV range should be 40 to 110 KV or more. • mA range (Rad.) : 300mA or more • Exposure time (Rad.): 1 ms to 2 sec. with maximum numbers of steps. **Control:** • A very compact, Soft Touch Control Panel having following functions & indications should be provided. The panel can be supplied in floor or wall mount with Spill Proof design Following features should be available on the control panel. • Machine ON/OFF switch • Digital Display of KV& mAs. • K V & mAs increase and decrease switches. • Tube focal spot selection switch. • Ready and x-ray on switch with indicators. • Bucky Selection switch. • Self diagnostic Programme with Indicators for Earth fault error, KV error, filament error & Tube's Thermal Overload. **X-Ray Tube** • One No Dual focus Rotating Anode BEL/Toshiba/Imported X-ray tube thermally protected having focal spot: • 1mm or less small Focus, 2mm or less large Focus.
	• Anode heat storage capacity of tube should be more than 140 KHU. • One no manual collimator with aluminum filter & for adjustment of exposure area. **Column Stand:** • It should have floor to ceiling stand with vertical counter balanced travel. • It should have 360 deg. Rotation. • It should be provided one vertical bucky stand with machine. • Table. • Five position manual tilt table having buky grid ration of 8:1 with 85 lines per inches should be provided. • The bucky tray should accept cassette of 8"x10", 10"x12" and 14"x17" size.
User Interface	Manual
Noise (in dBA)	Noise-free system
Heat dissipation	Heat Dissipation: Should maintain nominal Temp and the heat should be disbursed through an cooling mechanism
Mobility, portability	Certified Room installation
Power Requirements	Power unit: Input voltage- 400V-440V AC, 50Hz ;3 -phase
Protection	Stabliser of appropriate capacity to be installed.
Power consumption	25 to 30 KW
Accessories (mandatory, standard, optional); Spare parts (main ones); Consumables/ reagents (open, closed system)	Machine should be supplied with following transducers: 2 No. BARC Approved whole body lead apporns with all attachments. One Pair of 8 meter H. V. Cable

Scanned by CamScanner

...osphere/Ambiance ...r conditioning, ...umidity, dust ...)	1) Operating condition: Capable of operating continuously in ambient temperature of 5 to 50 deg C and relative humidity of 15 to 80% in ideal circumstances. 2) Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.
User's care, Cleaning, Disinfection & Sterility issues	1) Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/disposable cover. 2) Sterilization not required
Certificates (pre-market, sanitary, ..); Performance and safety standards (specific to the device type); Local and/or international	1. Should be FDA/European CE/BIS approved product. 2. Manufacturer and Supplier should have ISO 13485 certification for quality standards. 3. Electrical safety conforms to the standards for electrical safety IEC 60601- General requirements(or equivalent BIS Standard) 37 Technical Specification RADIOLOGY 4. Shall meet internationally recognised for Electromagnetic Compatibility (EMI/EMC) for electromedical equipment: 61326-1. 5. Certified to be compliant with IEC 61010-1, IEC 61010-2-40 for safety. 6. AERB type approved
Local and/or international	Manufacturer/supplier should have ISO 13485 certificate for quality standard
Pre-installation requirements: nature, values, quality, tolerance	1) Availability of three phase uniform power supply. 2) Safety and operation check before handover. 3) To be installed in a separate room. 4) Facility for dark room should be available
Requirements for sign-off	Certificate of calibration and inspection of parts from the manufacturer
Training of staff (medical, paramedical, technicians)	1) Training of users on operation and basic maintenance; 2) Advanced maintenance tasks required shall be document
Operating manuals, service manuals, other manuals.	Should provide 2 sets(hardcopy and soft-copy) of: 1) User, technical and maintenance manuals to be supplied in english/hindi language along with machine diagrams; 2) List of equipment and procedures required for local calibration and routine maintenance; 3) Service and operation manuals (original and copy) to be provided; 4) Advanced maintenance tasks documentation; 5) Certificate of calibration and inspection. 6) Satisfactory certificate for any existing installation from government hospital

Scanned by CamScanner