



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina(190018),Srinagar

email: jkmssljepm@gmail.com; website: www.jkmssl.nic.in

JKMSCL

**M/S Heager India (P)
Ltd, New Delhi
(Bidder/Importer)**

No: JKMSCL/GM/2024/4509/25

Dated: 24/08-2024.

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of rate contract thereof.

Ref No.

1. NIT No. JKMSCL/Mach/2023/558 dated 26.11.2023

2. Minutes of Purchase Committee vide No. JKMSCL/MD/2024-25/228-40 Dated 01.07.2024

3. LOI No. JKMSCL/GM/2024/2602-05 dated 12.07.2024.

Dear Sir,

As per the decision of the Purchase Committee, the Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour on the quoted rates by you in the NIT :


S. No.	Name of item	currency	Basic Rate without Tax with Conversion (1)	Taxes (2)	Total Amount (in Rs) with taxes (1+2)
1.	Hysteroscopy Set M/s Heager GmbH, Germany (Manufacturer) Model: Heager Hysteroscopy Set	INR	4810750.00	577290.00	53,88,040.00 (Rupees fifty three lakhs eight eight thousand and forty only)


The approved rates for Comprehensive Maintenance Contract after the expiry of warranty period of 05 (Five) years are as follows:

S. No	Name of Item	CMC for five years
1	Hysteroscopy Set M/s Heager GmbH, Germany (Manufacturer) Model: Heager Hysteroscopy Set	1 st yr 227067.00 2 nd yr 283834.00 3 rd yr 283834.00 4 th yr 283834.00 5 th yr 283834.00
		Total: - 1362403.00

Terms and Conditions:

- Delivery Site :** Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
- Validity :** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
- Taxes:** As mentioned above and as applicable.
- Warranty:** Five Years from the date of installation.
- Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.

6. **Supplies** Within 90 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%.
8. **Installation:** Installation/commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
9. **Availability of Spares:** Firm shall certify that the parts of the equipment shall remain available at least for a period of five years after the year of expiry of warranty period of five years.
10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
11. The supplier/manufacturer/importer shall ensure to supply the Brand new and unused item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.
13. **Fall Clause:** The prices under contract shall be subject to price fall clause. The prices charged for the goods supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically be reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
14. **Risk Purchase :** If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.
The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.
15. **Orders shall only be placed on the basis of actual requirement and funds as available under the relevant Head of Accounts.**
16. **Payment shall be released after successful installation and commissioning of the equipment duly verified by the intending department.**
17. **Approved technical specifications shall remain same as per NIT (Annexure I)**
All the terms and conditions of the NIT, Standard Procurement of Procedure (SPP) of JKMSCL & LOI shall be part of this Rate Contract.
- 

Yours sincerely,

(Mohd. Ashraf Choudhary) JKMSCL
General Manager (Adm.),
JKMSCL

Copy for information to the:

1. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
2. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
3. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
4. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
5. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
6. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
7. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
9. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
10. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
11. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
12. General Manager -K(P&S, IT), J&K Medical Supplies Corporation Limited.
13. Asstt Programmer to upload on official website.
14. I.T cell JKMSCL.
15. File

SPECIFICATIONS FOR HYSTEROSCOPY SET

	All items must be from one manufacturer for smooth compatibility and working of the system.
1	Operative Mini Hysteroscopy Set
	Full HD Hysteroscopy 2.9mm 30-degree, Full HD premium quality compact construction of optics optimizes quality and brilliant images, length 301mm, with rod lenses optical system, Autoclavable, with fiber optic light transmission Incorporated.
	4.8 mm - 5.0 mm Oval Shape Continuous flow Operative hysteroscopy sheath (Inner Sheath & Outer Sheath) for 2.9 mm forward oblique telescope 30, with a working channel with stop cock for instruments of 5 FT. Auto lock system, Discrete inflow/outflow channel provides superior continuous irrigation & a crystal clear operating view. Separate/Common instrument channel allows operative intervention whilst maintaining distension pressure.
	3.8 mm - 4.4 mm Oval Shape Hysteroscopy Examination Sheath with Inflow & instrument channel of 5 fr.
	Semi rigid Scissors 5Fr. Double single action jaws- Qty-02
	Semi rigid alligator Grasping forceps double action jaw 5Fr- Qty-02
	Semi rigid spoon biopsy forceps double action jaw 5Fr- Qty-02
	Monopolar/bipolar Hook electrode 5Fr
	Bipolar Needle electrode 5 Fr
	Unipolar high frequency cable
	Bipolar high frequency cable
2	Operative Hysteroscopy
	Full HD Hysteroscope 4mm 30 degree wide angled, Length 30-33cm
	Autoclavable, Laser welded. Pure stainless steel, with incorporated fiber optic Light transmission, Scratch free Sapphire tip and three different adapters for fiber optic cable connection- 01 No.
	Resectoscopes working element passive with high frequency connector for single/double stem electrodes. Spring action thumb control.
	26/23/24 Fr. Continuous flow Inner outer sheath, rotatable, Dismantlable with auto click locking system and two stop cocks for in flow and out flow. Ceramic protection built in with the inflow sheath with obturator.
	Loop Electrode 02 no
	Ball Electrode 02 no
	Colin's Knife- 02 no
	Unipolar High frequency cable- 02 no
3	Bipolar Operative Hysteroscopy
	Full HD Hysteroscope 4mm 30 degree wide angled, Length 30-33cm
	Autoclavable, Laser welded, Pure stainless steel, with incorporated fiber optic Light transmission, Scratch free Sapphire tip and three different adapter for fiber optic cable connection- 01 No.
	Bipolar Resectoscope working element passive with high frequency connector for single/double stem electrodes. Spring action thumb control
	26/23/24 Fr. Continuous flow Inner outer sheath, rotatable, Dismantlable with auto click locking system and two stop cocks for in flow and out flow. Ceramic protection built in with the inflow sheath with obturator.
	Bipolar Loop Electrode 02 no
	Bipolar Ball Electrode 02 no
	Bipolar Colin's Knife- 02 no
	Bipolar High frequency cable- 02 no

4	Mini Bipolar Hysteroscopy Set
	All items must be from one manufacturer for smooth compatibility and working of the system.
	a) Full HD Hysteroscope 2.9mm 30 degree wide angled, Length 30-33cm Autoclavable. Laser welded. Pure stainless steel, with incorporated fiber optic Light transmission, Scratch free Sapphire tip and three different adapters for fiber optic cable connection. 01 No.
	b) Bipolar resectoscope working element passive with high frequency connector for single/double stem electrodes. Spring action thumb control.
	c) 18.5 -22 Fr. Continuous flow Inner outer sheath. rotatable. Dismantlable with auto click locking system and two stop cocks for in flow and out flow. Ceramic protection built in with the inflow sheath with obturator.
	Bipolar Loop Electrode- 02
	Bipolar Ball Electrode- 02
	Bipolar Colin's Knife- 02
	Bipolar High frequency cable- 02
5	Full HD Camera Unit- 01 No
	It should have 3CMOS/CCD full high definition digital camera with Full HD image production.
	It should have Resolution 1920 x 1080
	It should have fully compatible to the color systems PAL/NTSC.
	It should have LCD Screen with Touch panel
	It should have Image Sensor: 1/3 CMOS Progressive scan
	It should have image resolution Lines:1920x1080 pixels
	It should have Function Keys: 3 Individually Programmable and illuminated keys
	It should have Technical Data Controller
	It should have Operation Control panel :On-Off, menu, white balance, Photo Capture
	It should have Video Outputs: High Definition 2 X DVI (1080p), 2 composite. 2 S video
	It should have Standard compliance: IEC 60601
	It should have Directive compliance-CE-Mark/USFDA Mark
	It should have Native Full HD Camera with resolution 1920x1080 pixels. progressive Scan
	It should have Edge Enhancement and dynamic contrast function for best-in-class performance
	It should have Latest Sensor technology for Comparable image quality at no pixel errors due to automatic pixel select compensation
	It should have Multiple usage Connectivity for different application i.e. Laparoscopy, Urology Gynaecology
	It should have All rigid. Flexible or semi flexible endoscopes with standard eyepiece can be connected
	It should have Selective color enhancement (SCE) for a much-improved representation of the vessels and the tissue structure
	It should have Grid removal-reduces the honey comb structure of image bundle endoscopes
	It should have Clear view-Clear the view if liquid blurs the visibility
	It should have Full HD Image capturing and video recording via USB
	It should have Automatic white balance
	It should have Facility of attaching additions recorder functions
	It should have Camera head fully soak-able for disinfection.
6	Cold LED Light Source with Touch Screen color LCD / LED display- 01 No.
	It should have LED light source or equivalent to 300w Xenon light source with extra cooling facility
	Color temperature > 6900°K
	Recommended Lamp life 30000 or more hours.