



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)

Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu

Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina (190018), Srinagar

email: mdjkmssl2@gmail.com; website: www.jkmsslbusiness.com

M/s S.R Technomed,
Bakshi Nagar, Jammu. (Bidder)

No: JKMSCL/GM(Adm.)/2025/6589-6603

Dated: 03-10-2025.

Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of Rate Contract for the item "(Electro Surgical Unit)".

Ref. No.

1 NIT No. NIT/JKMSCL/Mach/2023/575 Dated 09-02-2023

2. Minutes of Purchase Committee vide No. JKMSCL/PS/MD/2025-26/395-407 dated 12.08.2025.

3. LOI No. JKMSCL/GM(Adm)/2025/4721-26 dated 18.08.2025.

Sir,

As per the decision of the Purchase Committee, the Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour on the quoted rates by you in the NIT:

S. No.	Name of item	Currency	Basic Rate without Tax with conversion (1)	Taxes (2)	Total Amount (in Rs) with taxes (1+2)
1.	Electro Surgical Unit M/s India Medtronic Pvt. Ltd., Mumbai. (Importer) M/s Medtronic Inc, USA Manufacturer) Model No. Valleylab Force -FX8	INR	1058726.78	127047.21	11,85,773.99 (Rupees eleven lakhs eighty five thousand seven hundred seventy three and ninety nine paise only)

The approved rates for Comprehensive Maintenance Contract after the expiry of warranty period of 05 (Five) years are as follows:

S. No	Name of Item	CMC for five years
1	Electro Surgical Unit	6 th yr 10620.00 7 th yr 10620.00 8 th yr 10620.00 9 th yr 10620.00 10 th yr 10620.00 Total: 53100.00

Terms and Conditions:

- Delivery Site :** Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.
- Validity :** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
- Taxes:** As mentioned above and as applicable.
- Warranty:** Five Years from the date of installation.
- Performance Security**
You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
- Supplies**
Within 90 days from the date of issuance of supply order.
- Penalty**
In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay

subject to maximum of 10%.

8. **Installation:** Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
9. **Availability of Spares:** Firm shall certify that the parts of the equipment shall remain available at least for a period of **five years** after the year of expiry of warranty period of five years.
10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
11. The supplier/manufacturer/importer shall ensure to supply the Brand new and unused item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.
13. **Fall Clause:** The prices under contract shall be subject to price fall clause. The prices charged for the goods supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically be reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
14. **Risk Purchase :** If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.
The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.
15. **Orders shall only be placed on the basis of actual requirement and funds as available under the relevant Head of Accounts.**
16. **Payment shall be released after successful installation and commissioning of the equipment duly verified by the intending department.**
17. **Approved technical specifications shall remain same as per NIT (Annexure I)**

All the terms and conditions of the NIT, Standard Procurement of Procedure (SPP) of JKMSCL & LOI shall be part of this Rate Contract.

Issued with the Decision and Authorization of Purchase Committee of JKMSCL vide its minutes No. JKMSCL/PS/MD/2025-26/395-407 dated 12.08.2025.


Yours sincerely,

3.10.25

(Mohd. Ashraf Choudhary) JKAS
General Manager (Adm.), JKMSCL

Copy for information to the:

1. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).

2. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
 3. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
 4. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
 5. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
 6. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
 7. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
 8. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
 9. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
 10. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
 11. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
 12. General Manager -K(P&S, IT), J&K Medical Supplies Corporation Limited.
 13. Asstt Programmer to upload on official website.
 14. IT cell JKMSCL.
 15. File
- 

ELECTRO SURGICAL UNIT

ELECTRO SURGICAL UNIT

All purpose digit electro surgical unit with Micro-processor based technology with capacity of 300 – 400 watts and can be used for all Open & laparoscopic surgeries including liver resection & liver transplantation with monopolar and bipolar modes available with capabilities of Mono and Bipolar (cutting and coagulation) with Monopolar under water cutting and coagulation facility.

Each unit should have:

1. Facility of automatic output control through computerized memory
2. Digital display for power setting with touch screen display.
3. Programmable power settings/ recall mode for last setting.
4. Tissue feedback system to adjust power to tissue changes.
5. Two monopolar and one bipolar output.
6. The machine should have minimum 6 bipolar modes of energy.
7. The device should have minimum 4 cutting and coagulation modes.
8. Bipolar coagulation should be available via footswitch and with auto-bipolar mode with bipolar current meter.
9. The system should have audio and visual patient plate error monitoring system.
10. Separate foot pedal for mono polar and bipolar.
11. Should be compatible for use with endoscopic coagulation.
12. The device should come with inbuilt memory of 8Gb for error log and storage.
13. The device should self-test every time it is switched on.
14. The device should be able to use two monopolar output simultaneously.
15. Facility to use monopolar and bipolar without switchover for ease during surgery.
16. The coagulation mode can be shared while using simultaneous monopolar modes.
17. Facility to enable and disable auto bipolar mode as per requirement by the surgeon.
18. The device should be future ready and any upgrade should be done free of cost.
19. The OEM should quote the latest available model present in global market.
20. Automatic limitation of maximum HF power.
21. Output power adjustment with touch screen.
22. The equipment should be compatible to be used along CUSA, Smoke Evacuator and Argon Plasma Generator.

Safety Standard:

1. IEC 60601-1:2012 for leakage currents & patient Auxiliary currents.
2. Safety class I as per IEC 601. Safety type.
3. Standards of voltage transients-Energy platform mains transfer (IEC 60601-1, IEC 60601-2-2, and ANSI/ AAMI HF18) so that system continues to operate normally with no errors or system failures when transfer is made between line AC and an emergency system voltage source. The system may momentarily shut down in safe mode.
4. CF Check for LF leakage with alarm as per IEC 601.1.
5. Check for HF leakage with alarm Maximum 150 mA Earth leakage current.
6. Maximum 0.5 mA Enclosure leakage current.
7. Maximum 0.1mA Monopolar leakage current.
8. Maximum 0.01mA Bi-polar output leakage current Maximum 0.1 mA.

Accessories.

The unit should be supplied with following accessories:

1. 3 Button switch pencil with electrode and with cable with cutting and coagulation modes, single use – 25 pcs.
2. Patient plate along with cable with split pads and polyhesive material, single

use - 50pcs.

3. Bipolar forcep Resuable - 1.
4. Bipolar Cord Reusable - 1.
5. Monopolar Cord for laproscopic surgeries Reusable - 1.
6. Disposable Laparoscopic Smoke Evacuator - 10.
7. Clean Coat Cautery L - Hook electrode Laproscopic - 10.
8. Clean Coat Cautery Spatula electrode Laproscopic - 10.

The device and accessories should be USFDA approved.

