



JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of Govt of Jammu & Kashmir)

Corporate Office Jammu: Plot No 58, Friends Colony, Satyam Road, Trikuta Nagar, Jammu

Corporate Office Kashmir: Opposite State Motor Garage, Near Hajj House, Bemina Srinagar

Telephone: 0191-3510489 (Jammu), Telefax: 0194-2432008 (Srinagar)

M I N U T E S O F M E E T I N G TECHNICAL ADVISORY/EVALUATION COMMITTEE 22-12-2025, 12:30 P.M

Sub: Acceptance of the report of the Physical Demonstration for tender invited for procurement of "Dental Machinery and Equipments" uploaded vide No: JKMSCL/Dental machinery & Equipments/2024/641 dated 27-11-2024.

e. Bid for the finalization of Rate Contract for the procurement of **Dental Machinery and Equipments** was uploaded vide JKMSCL/Dental Machinery & Equipments/2024/641 dated 27-11-2024.

1. As per tender notification Pre-Bid meeting was held on scheduled date i.e. 09-12-2024 at Corporate Office Jammu and Srinagar.
2. Corrigendums/Addendums were issued vide this office No: JKMSCL/Dental Material/Addm/641/2024/9589-92 dated 27-12-2024 and JKMSCL/Dental Machinery/Corg/641/2025/12265-68 dated 21-01-2025 & JKMSCL/GM(Adm)/2025/12793-97 dated 04-02-2025
3. Accordingly, e-bid was opened on 21-02-2025 at 11.00 a.m., in the presence of Technical Evaluation Committee constituted by Boards of Directors during 1st Board Meeting and amended by Boards of Directors during 3rd Board Meeting JKMSCL for the purpose, and as per the system generated report total of 95 items have been **bidded by the** participated firms in the said NIT.
4. After the opening of e-bids, the Technical Evaluation Committee decided that the compilation, preliminary scrutiny and evaluation of bid documents shall be carried out by the Technical Evaluation Sub Committees constituted for the purpose.
5. With reference to clause 2.1.9 Chapter II of the Standard Procurement Procedure approved during 2nd Board Meeting, it was decided by the technical advisory/evaluation committee that the Chairman, Technical Advisory / Evaluation Committee shall get the minor infirmities, if any, reported by the sub-committee, from the respective bidders, uploaded on website www.jkmsclbusiness.com by permitting a specific period to get the documents/infirmities uploaded on website and in case, the file(s) is/are heavy and the bidders are not in position to get the same uploaded on JKMSCL website, the same may be sent to corporation on the email address given by the JKMSCL Authorities which shall be forwarded to technical evaluation subcommittee for scrutiny and recommendations in the matter. After downloading the minor infirmities submitted by the bidders, the same shall be re-evaluated by the Sub Committee constituted for the purpose and the final report shall be placed before the Technical Advisory/ Evaluation Committee for further decision/ recommendation.

6. Accordingly, the minor infirmities reported by the sub-committee were asked from the bidders by the way of a Minor Infirmitiess Notice issued vide this office Notice No: JKMSCL/GM/Adm/MI/641/2025/2027-31, dated 31-05-2025, however it was observed that most of the bidders could not submit the infirmities and the sub-committee proposed to extend the date of submission of minor infirmities in the interest of healthy competition.



7. Thereafter the report of the subcommittee was placed before the Technical Advisory/Evaluation Committee in its meeting held at JKMSCL Corporate office at Jammu/Srinagar on dated 17-9-2025.

8. The Technical Advisory/Evaluation Committee decided to conduct the Physical Demonstration of all the items qualified in the technical bid as per the minutes of the meeting circulated vide No: JKMSCL/Dental Machinery/641/2025/6701-07 dated 06-10-2025.


9. Accordingly, letters were issued to the Principal, Govt. Dental Colleges Jammu/Srinagar and to the Director Health Services Jammu/Srinagar vide No: JKMSCL/GM/Adm/2025-26/7502-08 dated 11-11-2025 requesting for deputing technical experts for physical demonstration. Bidders were also informed vide Corrigendum No: JKMSCL/GM/Adm/Corg-641/2025-26/7496-7501 dated 11-11-2025 to conduct the Physical Demonstration of their items by the Technical Experts from end user department at Sample Warehouse of JKMSCL with effect from 25-11-2025 to 29-11-2025.

10. The technical experts from the end user department carried out the physical demonstration of the Dental Machinery Equipments made available by the concerned bidder and submitted its final report on 15-12-2025 wherein some items were approved and some were not approved. The detail of approved and not approved items is given as under:

No. of Bidders	Computerized Dental Occlusal Analysis System and biometric Equipments (Addendum Item) (DEN2-264)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	<p>M/s Villa India New Delhi (Bidder/Importer)</p> <p>M/s TEKSCAN + BIORESEARCH INC,USA (Manufacturer)</p> <p>Model: TSCAN NOVUS + BIOEMG +QUADRATENS + JT3D +JVA</p>	<p>As per Specifications</p> <p>Note: <u>Proprietary Article</u></p>	Approved






No. of Bidders	Lab Micromotor (DEN2-222)	Remarks of the Technical Experts	Approved/Not Approved
Bidder-1	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Mighty e-40	As per specifications	Approved
No. of Bidders	Surgical Motor with Contra angle and Straight Hand Piece (As per corrigendum) (Addendum item) (DEN2-209)	Remarks of the Technical Experts	Approved/Not Approved
Bidder-1	M/s Confident dental equipments limited Karnataka (Bidder/Importer) M/s Nakanishi Inc. Japan (Manufacturer) Model: Surgic AP2 (NSK)	As per specifications	Approved
No. of Bidders	Peizo surgical unit Den2-210 (J)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s VA Technologies Jammu (Bidder) M/s Unicorn Denmart Ltd Delhi (Importer) M/s NSK Japan (Manufacturer) Model: VarioSurge-3	As per specifications	Approved
No. of Bidders	Lead Apron (DEN2-179)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s Horizon Healthcare Solutions Delhi (Bidder/ Manufacturer/ MSME) Model: 3204LL-X Cordon Lead frontal apron	Not As per specifications	Not Approved

 3 


No. of Bidders	Conscious Sedation kit (DEN2-176)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s First Medical New Delhi (Bidder) M/s Consed International Kerala (Manufacturer) Model: CS/AT-2.0	As per specifications	Approved
No. of Bidders	Intra Oral Camera (DEN2-203)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medicam Maharashtra (Manufacturer) Model: DENTACAM-DLXPLUS	Demonstration shown for Model other than quoted. Not as per the specification.	Not Approved
No. of Bidders	Tri Headed Microscope (DEN2-121)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s Radical Scientific Equipments Pvt Ltd Ambala (Bidder/ Manufacturer) Model: RXLr-5NX	Not as per specifications Image Quality not up to mark. USFDA/CE certificates not attached.	Not Approved
Bidder-2	M/s Biomed Systems Srinagar (Bidder) M/s Nikon India Pvt Ltd (Importer/Indian Subsidiary) M/s Nikon Corporation Japan (Manufacturer) Model: Eclipse ci-L Plus with accessories	As per specifications	Approved
No. of Bidders	Water distiller table top (As per Corrigendum) (Addendum item) (DEN2-200)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Pvt Ltd Karnataka	As per specifications	Approved

	(Manufacturer) Model: Medidist		
No. of Bidders	Endo motor with Cordless hand piece (DEN2-205)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s Bashaw Enterprises Srinagar (Bidder) M/s Coltene Whaledent Pvt Ltd Maharashtra (Manufacturer) Clarify Model: Canal Pro CI2i	As per specifications	Approved
No. of Bidders	Digital OPG Machine without CEPH (As per Corrigendum)(Addendum Item) DEN2-244A	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s Villa India New Delhi (Bidder/Importer) M/s Cefla S.C Italy (Manufacturer) Model: Newton Go Version 2D OPG without Ceph	As per specifications	Approved
No. of Bidders	Mobile Suction Unit (DEN2-94)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Pvt Ltd Karnataka (Manufacturer) Model: MEDVAC	As Per the Specification	Approved
No. of Bidders	Light Cure Gun (DEN2-247)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder- 1	M/s Bashaw Enterprises Srinagar (Bidder) M/s Coltene Whaledent Pvt Ltd Maharashtra	Not as per specifications: Built in intensity meter non compliance	Not Approved



S




	(Manufacturer) Clarify Model: Coltolux Comfort		
Bidder- 2	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Bee Cool Plus Cordless Lightcure	Sample Not Shown	Not Approved
No. of Bidders	Digital Panoramic and Cephalometric Imagine system (OPG) DEN2-244	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder -1	M/s Villa India New Delhi (Bidder/Importer) M/s Cefla S.C Italy (Manufacturer) Model: Newton Go Version 2D OPG with Ceph	As per Specifications	Approved
Bidder-2	M/s VA Technologies Jammu (Bidder) M/s Unicorn Denmart Ltd Delhi (Importer) M/s Genoray Korea (Manufacturer) Model: Papaya Plus	As per specifications	Approved
No. of Bidders	Endodontic Motor Reciprocating motion(DEN2-231)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder- 1	M/s Bashaw Enterprises Srinagar (Bidder) M/s Coltene Whaledent Pvt Ltd Maharashtra (Manufacturer) Model: Canal Procl 2i	As per specifications	Approved
No. of	Phantom Head (DEN2-181)	Remarks of the	Approved/Not





Bidders		Technical Expert(s)	Approved
Bidder-1	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Phantom Head Table C-125A	As per specifications	Approved
No. of Bidders	Portable Light Cure Units (DEN2-118)	Remarks of the Technical Expert(s)	
Bidder-1	M/s Bashaw Enterprises Srinagar (Bidder) M/s Coltene Whaledent Pvt Ltd Maharashtra (Manufacturer) Model: S.P.E.C.3	Not as per specifications	Not Approved
Bidder-2	M/s Confident dental equipments limited Karnataka (Bidder/Manufacturer) Model: Bee Cool Plus Cordless Light Cure	Not as per specifications/Not Compliant	Not Approved
No. of Bidders	Soft Tissue Laser (8 Watt or More) (DEN2-08)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s First Medical New Delhi (Bidder) M/s Intense Medical and Dental System Pvt Ltd M.P (Manufacturer) Model: Bright Laser	As per specifications	Approved
No. of Bidders	Air Compressor (DEN2-04)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies (Bidder) Chisa Inc. Karnataka (Manufacturer) Model: EKD	As per specifications	Approved
Bidder-2	M/s Confident dental equipments limited	As per specifications	Approved

	Karnataka (Bidder/ Manufacturer) Model: Air Compressor 0.75 HP		
No. of Bidders	Pedo RVG Laptop (As per Corrigendum) (Addendum Item) DEN2-39(A)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Pvt Ltd Karnataka (Manufacturer) Model: Medx Digital Intra Oral Sensor	As per the Specification	Approved
No. of Bidders	Electrically Operated Multi- Program Movable Pedo Chair (DEN2-135)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s Besto Dent India Pvt Ltd Haryana (Bidder/ Manufacturer/MSME) Model: DELUXE	Demonstration Not Shown	Demonstration Not Shown
Bidder-2	M/s Suz-Dent India Pvt Ltd Gujarat (Bidder/ Manufacturer) Model: SUZAARYA	Demonstration Not Shown	Demonstration Not Shown
Bidder-3	M/s S.S Agencies Jammu (Bidder) M/s Chisa Inc. Karnataka (Manufacturer) Model: GEMS	As per Specifications	Approved
Bidder-4	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Wave Pedo Dental Chair	Demonstration Not Shown	Demonstration Not Shown
Bidder-5	M/s Bhardwaj Agencies Jammu (Bidder) M/s Clinix intelligent Medical systems Pvt Ltd Karnataka Model: MONALISA PEDO	Not As per Specifications: Pedo hand piece not available. Digital Display of Micromotor not available. Conventional type Control box attached.	Not Approved
Bidder-6	M/s S.R Technomed	As per Specifications	Approved

8 



	Jammu (Bidder) M/s CRB International Mohali, Punjab (Manufacturer) Model: EL-03		
No. of Bidders	Dental chair (Fully Loaded) without compressor (As per the corrigendum) (Addendum) (DEN2-19)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s BestoDent India Pvt Ltd Haryana (Bidder/ Manufacturer/ MSME) Model: Classic	As per Specifications	Approved
Bidder-2	M/s Suz Dent India Pvt Ltd Gujarat (Bidder/ Manufacturer) Model: SUZAARYA	As per Specifications	Approved
Bidder-3	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Private Limited Karnataka (Manufacturer) Model: Classic-M25	Not As per Specifications. Quality of the material of the Chair not up to Mark. Power of suction (High/low) not up to the mark.	Not Approved
Bidder-4	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Wave Dental Chair fully loaded	As per Specifications	Approved
Bidder-5	M/s Bhardwaj Agencies Jammu (Bidder) M/s Clinix intelligent Medical systems Pvt Ltd Karnataka Model: MONALISA ELECTRA	Not As per Specifications Quality of the material of the Chair not up to Mark. Digital display of the Micrometer not inbuilt.	Not Approved
Bidder-6	M/s S.R Technomed Jammu (Bidder)	As per Specifications	Approved

Handwritten signatures are present at the bottom right of the table, consisting of a signature of the Technical Expert(s) and a signature of the Approver.

	M/s CRB International Mohali, Punjab (Manufacturer) Model: EL-02		
No. of Bidders	Dental chair (Basic) with compressor 0.75 HP (As per the corrigendum) (Addendum) Basic Dental Chair Electronic DEN2-19(A)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s BestoDent India Pvt Ltd Haryana (Bidder/ Manufacturer/ MSME) Model: Classic	As per Specifications	Approved
Bidder-2	M/s. Suz Dent India Pvt Ltd Gujarat (Bidder/ Manufacturer) Model: SUZAARYA	As per Specifications	Approved
Bidder-3	M/s S.S Agencies Jammu (Bidder) M/s CHESA Inc. Karnataka (Manufacturer) Model: New Jwala	As per Specifications	Approved
Bidder-4	M/s Confident dental equipments limited Karnataka (Bidder/ Manufacturer) Model: Wave Dental Chair Basic	As per the Specifications	Approved
Bidder-5	M/s Bhardwaj Agencies Jammu (Bidder) M/s Clinix intelligent Medical systems Pvt Ltd Karnataka Model: MONALISA ELECTRA	Not as per the Specifications: Quality of chair material not up the mark. No inbuilt micrometer with digital display.	Not Approved
Bidder-6	M/s S.R Technomed Jammu (Bidder) M/s CRB International Mohali, Punjab (Manufacturer) Model: EL-02	As per specifications	Approved



No. of Bidders	RVG with Laptop (DEN2-39)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Pvt Ltd Karnataka (Manufacturer) Model: Medx + Digital Intraoral Sensor	As per Specifications	Approved
	B-Class Autoclave (DEN2-189)	Remarks of the Technical Expert(s)	Approved/Not Approved
Bidder-1	M/s S.S Agencies Jammu (Bidder) M/s Medtech Remedies Pvt Ltd Karnataka (Manufacturer) Model: Med claveB	As per Specifications	Approved

Decision:

"The committee members after threadbare discussion accepted the report of the Physical Demonstration of Technical Experts and recommended to open the financial bid of the approved items and for the items which resulted in the single bid after demonstration should proceeded further after ascertaining the Rate Reasonability.

The committee further recommended that for the items where single bid has been received. The physical demonstration of the said items should also be conducted at the earliest for further course of action.

The meeting was attended by following Officers/Official:

S.No	Name of Officer/Official	Designation	Signature	Contact No
1.	Dr. Shahid Hassan.	Associate Prof.		7006759992
2.	Dr. Mohsin Siddiq	Associate Prof		7006529237
3.	Mr. Moud Arif Dar	Assistant Prof		9419166990
4.	Dr. Ajaz Masroor	Associate Prof		7006983663
5.				

6.			
7.			
8.			
9.			
10.			
11.			
12.			
13.			
14.			
15.			
16.			
17.			

General Manager(Adm)
J&K Medical Supplies Corp. Ltd.

Dated: /01/2026

No: JKMSCL/Dental Machinery/641/2025/

Copy for information to the:-

1. Principal, Govt. Dental College, Jammu/ Srinagar.
2. Director Health Services Jammu/ Kashmir.
3. Managing Director, J&K Medical Supplies Corporation Limited.
4. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
5. General Manager (K), J&K Medical Supplies Corporation Limited.
6. Dy. General Manager (Tendering), J&K Medical Supplies Corporation Limited.
7. Medical Officer (Dental), J&K Medical Supplies Corporation Limited.
8. Office file.