



# JAMMU AND KASHMIR MEDICAL SUPPLIES CORPORATION LTD.

(Public Sector Undertaking of the Government of Jammu and Kashmir)  
Corporate Head Office: Corporate Head Office: Plot No. 58, Friends Colony Satyam Road Trikuta Nagar Jammu  
Corporate Office: Corporate Office Kashmir: Near Haj House, Bemina (190018), Srinagar  
email: [jkm scl@pm@gmail.com](mailto:jkm scl@pm@gmail.com); website: [www.jkm scl.nic.in](http://www.jkm scl.nic.in)

M/s Vikrant Life Sciences Private Limited.  
New Delhi (Bidder)

No: JKMSCL/2024/5842-56

Dated: 09/10-2024.

**Sub: E.bid for the finalization of Rate Contract for the procurement of "Machinery & Equipments" - issuance of Rate Contract thereof.**

Ref. No.

1. NIT No. JKMSCL/Mach/2023/596 dated 20-08-2023
2. Minutes of Purchase Committee vide No. JKMSCL / MD/2024-25/99-109 dated 01.05.2024 held on 29.04.24.
3. LOI issued vide No. JKMSCL/2024/1076-79 dated 13.05.2024.

Sir,

As per the decision of the Purchase Committee, the Jammu and Kashmir Medical Supplies Corporation J&K has approved the following item/item(s) under group "Machinery & Equipments" in your favour on the quoted rates by you in the NIT:

S. No.	Name of item	currency	Basic Rate without Tax with Conversion (1)	Taxes (2)	Total Amount (in Rs) with taxes (1+2)
1.	Deep Freezer (-86 °C) M/S Esteem Industries INC, Baddi (H.P.) (Manufacturer) Model No. ESTD-109.	INR	429000.00	77220.00	506220.00 (Rupees five lakhs six thousand two hundred twenty only)
2.	Water Bath M/S Esteem Industries INC, Baddi (H.P.) (Manufacturer) Model No. ESTW-411.	INR	9800.00	1764.00	11564.00 (Rupees eleven thousand five hundred sixty four only)

The approved rates for Comprehensive Maintenance Contract after the expiry of warranty period of 05 (Five) years are as follows:

S. No	Name of Item	CMC for five years
1	Deep Freezer (-86 °C)	6 <sup>th</sup> yr=29500.00 7 <sup>th</sup> yr=43660.00 8 <sup>th</sup> yr=59000.00 9 <sup>th</sup> yr=73160.00 10 <sup>th</sup> yr=88500.00 Total CMC amount : 293820.00
2.	Water Bath	6 <sup>th</sup> yr=678.00 7 <sup>th</sup> yr=1017.00 8 <sup>th</sup> yr=1356.00 9 <sup>th</sup> yr=1695.00 10 <sup>th</sup> yr=2035.00 Total CMC amount : 6781.00

### Terms and Conditions:


1. **Delivery Site** : Rate are FOR Drug warehouse of JKMSCL/any Govt. Institution of J&K State, wherever installation is required.

2. **Validity :** The rate contract shall remain valid for a period of two years from the date of issuance of the Contract which may be extended for a period of 90 days or till such time the new rate contract is finalized, whichever is earlier, with mutual consent of JKMSCL and successful bidder.
3. **Taxes:** As mentioned above and as applicable.
4. **Warranty:** Five Years from the date of installation.
5. **Performance Security** You are further required to deposit the performance security equivalent to 5% of the total invoice value in the shape of Bank Guarantee in favour of FA/CAO, JKMSCL within seven days from the date of issuance of Purchase Order(s) from time to time.
6. **Supplies** Within 60 days from the date of issuance of supply order.
7. **Penalty** In case of extension in the delivery period with liquidated damages, recovery of liquidated damages shall be made at the rate of 0.25% per day for every day of delay subject to maximum of 10%.
8. **Installation:** Installation / commissioning and necessary training for upkeep of the equipment shall be provided free of cost.
9. **Availability of Spares:** Firm shall certify that the parts of the equipment shall remain available at least for a period of five years after the year of expiry of warranty period of five years.
10. You have to certify that the rates mentioned above as quoted by your firm are the lowest rates applicable to all Govt./private Institution/PSU etc. Variation found if any, during the rate contract period by the Corporation/Intending department or any other agency shall be the sole responsibility of the Contractor/Supplier/manufacturer/importer.
11. The supplier/manufacturer/importer shall ensure to supply the Brand new and unused item(s) strictly as per specifications/sample(s) approved. Variations found, if any, in instant case/or in later stages by the Corporation/Intending department/any other agency which may lead to sub standard or Not of standard/approved quality, shall be the sole responsibility of the supplier/ manufacturer/ importer/contractor.
12. JKMSCL is at liberty to get the feedback from the end-user department/intending department regarding the functioning of the machinery & equipment supplied by your agency. Adverse report, if any, may lead to strict action against you under rules/terms & conditions of the NIT/guidelines issued by the Govt. from time to time.
13. **Fall Clause:** The prices under contract shall be subject to price fall clause. The prices charged for the goods supplies under the contract by successful bidder shall in no event exceed the lowest price at which the successful bidder sells the goods or offers to sell the goods of identical description to any other person(s)/Organization(s) including procurement agencies or any other department of Central Govt. or State Govt. as the case may be during the period till the performance of the contract. If at any time, during the period of the contract, the bidder reduces the sales price chargeable under the contract, he shall forth with notify such reduction to the JKMSCL, J&K and the price payable under the contract for the stores supplied after the date of coming into force of such reduction or sale shall stand reduced correspondingly. It imply that if the contract holder quotes/ reduces its price to render similar goods at a price lower than the contract price to anyone in the Central/State Govt. at any time during the currency of contract including extension period, the contract price shall be automatically be reduced with effect from the date of reducing or quoting lower price for all delivery of subject matter of procurement under contract and the contract shall be amended accordingly.
14. **Risk Purchase :** If the bidder is unable to complete the supply within the specified or extended period, the corporation shall be entitled to purchase the goods or any part thereof from elsewhere without notice to the bidder on his (i.e., bidders) account at his cost and risk, with the prior approval of Managing Director JKMSCL, J&K. The bidder shall be liable to pay any loss or damage which the purchasing officer may sustain by reasons of such failure on the part of the bidder.  
The bidder shall not be entitled to any gain on such purchases made against default. The recovery of such loss or damage shall be made from any sums accruing to the bidder under this or any other contract with the corporation/government. If recovery is not possible from the bill and the bidder fails to pay the loss or damage within one month of the demand, the recovery of such amount or sum due from the bidder shall be made from the bidder. In case supplier fails to deliver ordered goods, the risk purchases may be made at market rate from any other firm. It is mandatory for the approved supplier to acknowledge receipt of orders within seven days from the date of dispatch of order, failing which the procuring entity will be at liberty to initiate action to purchase the items on risk purchase provision at the expiry of the prescribed supply period.
15. **Orders shall only be placed on the basis of actual requirement and funds as available under the relevant Head of Accounts.**
16. **Payment shall be released after the execution of the agreement.**
17. **Approved technical specifications shall remain same as per NIT (Annexure I)**

All the terms and conditions of the NIT, Standard Procurement of Procedure (SPP) of JKMSCL & LOI shall be part of this Rate Contract.

*Issued with the Decision and Authorization of Purchase Committee of JKMSCL vide its minutes No. JKMSCL /MD/2024-25/99-109 Dated 01.05.2024.*

Yours sincerely,

  
(Mohd. Ashraf Choudhary) JKMS  
General Manager (Adm.), JKMSCL

Copy for information to the:

1. Principal, Govt. Medical College, Srinagar (Kashmir) (Director, Board of Directors- JKMSCL).
2. Principal, Government Medical College, Jammu (Director, Board of Directors JKMSCL).
3. Principal, Government Dental College, Srinagar (Director, Board of Directors- JKMSCL).
4. Principal, Government Dental College, Jammu (Director, Board of Directors- JKMSCL).
5. Director General, Indian System of Medicines, J&K (Director, Board of Directors- JKMSCL).
6. Director, Family Welfare, MCH & Immunization, J&K (Director, Board of Directors- JKMSCL).
7. Director, Health Services, Jammu (Director, Board of Directors- JKMSCL).
8. Director, Health Services, Kashmir (Director, Board of Directors- JKMSCL).
9. Mission Director, NHM, J&K (Director, Board of Directors- JKMSCL).
10. Controller, Drug & Food Control Organization, J&K (Director, Board of Directors JKMSCL).
11. F.A/Chief Accounts Officer, J&K Medical Supplies Corporation Limited.
12. General Manager –K(P&S, IT), J&K Medical Supplies Corporation Limited.
13. Asstt Programmer to upload on official website.
14. I.T cell JKMSCL.
15. File

## Technical Specifications

### Annexure I to Rate Contract No. \_\_\_\_\_

#### Deep Freezer (-86°C):

##### Technical Specifications:

**Upright Model:** CFC free high efficiency double refrigeration system for cooling and freezing filled in the bottom.

**Temperature:** 0-degree C-86-degree C + - 0.5-degree C (to work in -70 to -86-degree C range)

##### Temperature Control:

- Digital temperature controller (including display at suitable eye level) -

-Microprocessor Control/Microcontroller-for temperature setting

-Alarms for-Voltage, Over heat, Over cool as well as for under temperature, power fail, Door ajar conditions.

**Size:** 300 litres or more (up to 450 L) with minimum 4 Stainless Steel, rust free shelves. Fixed in casters for easy manoeuvrability, Polystyrene Insulated inner doors for the compartments. Pre-coated metal body to prevent environmental damage.

**Electricity:** 230 volts AC, 50Hz single phase. The line cord / Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

**Refrigeration system:** Heavy-duty maintenance free refrigeration system with hermetically sealed refrigeration compressors and reliable cascaded refrigeration to minimize noise and vibration. Air Cooled with security lock to prevent unintentional switch off. Short cooling time of 4 to 5 hours at 40°C ambient temperature. The equipment should be of continuous duty and frost-free. Convenient Air Filter Grill allows easy access for cleaning and changing. Access port for CO2 back up. It shall be fitted with 24x7 temperature recorders/ data loggers which allows for a minimum of 3 GB data storage and the data must be downloadable via a USB port. Deep freezer shall not have an automated defrosting system without a manual override. Audiovisual Electronic Alarm System independent of power supply and Remote alarm contact in case of equipment failure/ power failure Electrically heated doors for quick opening of frozen doors.

Accessories to include suitable boxes and racks for storage of specimen deposits/ DNA extracts/culture isolates in cryo-vials (16 in No. of suitable dimension for 2 ml cryo-vials). Cyro gloves (4 sets wrist length 12") to be provided as part of accessories.

##### Remarks:

The apparatus should confirm to Indian Standard Institution Guidelines with latest amendments in Indian Standard Specification for Laboratory Deep Freezers or equivalent National or international Standards covering Markings, tests and Safety requirements.

Voltage regulator of appropriate rating to be included to cope with 160-260 V

#### Water Bath:

##### Technical Specifications:

- Stainless Steel, insulated double walled,

- Inner wall of stainless steel,

- Thermostatic temp. control from ambient to 85- 90° C (A0.50) complete with immersion heater,

-Aluminium /SS cover,

-Brass drain cock,

-220-240 Volts AC, 50Hz. The line cord/ Power cord supplied with the equipment shall be of acceptable durability, length, and current carrying capacity complying with Indian Standards.

-Dimensions outside:- (approx.) 36x41x25 cms; inside:- (approx.)27x30x15 cms;

Capacity not given because the dimensions specified are required for the proper functioning of the water bath in the Laboratory for optimum utilization of the space.

- Power: (approx.) 480W;

- Digital microprocessor display to set temperature point preventing thermal runaway.

-Seamless reservoir with no welds to leak or rust, see- through cover is hinged and removable, and steeply gabled to accept taller samples.